

PAC HOSPITAL KAMRA (DISTT ATTOCK)
TENDER NOTICE FOR ELECTRO MEDICAL EQUIPMENT
FINANCIAL YEAR: 2025-26

1. Quotations are invited from the sales tax registered firms for purchase of following item according to the specifications mentioned against each for used in PAC Hospital Kamra:-

S No	Description	Qty
(a)	Defibrillator with AED Mode	01

(a) Specifications of Defibrillator with AED Mode:-

- | | | | |
|-------|------------------------------|---|---|
| I. | Impulse shape | : | Biphasic, current controller impulses, external (CCD) |
| II. | Energy in the automatic mode | : | 140-360J according to patient impedance |
| III. | Manual Energy Stages | : | 5J, 10J, 20J, 30J, 50J, 100J, 200J, 300J & 360J |
| IV. | Charge Time | : | < 6 sec for 200J |
| V. | Analysis Time | : | 7-12s |
| VI. | Asystole Threshold | : | < 200 μ V |
| VII. | Cardioversion | : | Manual Activation |
| VIII. | ECG Derivation | : | Einthoven I, II, III Goldberger aVL, aVR & aVF |
| IX. | Patient Resistance | : | 23 - 200 Ohm |
| X. | Heart Rate | : | 0-300 per min |
| XI. | Monitor | : | 3 TFT, LCD, blue mode |
| XII. | Dimensions | : | 6x18x23 mm to 12x24x30 mm |
| XIII. | Resolution | : | 320 X 240 pixels to 800 x 480 pixels |
| XIV. | Battery | : | Monitoring 5:30 hrs |
| XV. | Charge Time | : | Approx 3:20 hrs |
| XVI. | AED Mode | : | |

STANDARD / SPECIAL CONDITION

- (a) Payment as per PAC Board purchase procedure.
- (b) Item should be as per standard of ISO 13485.
- (c) Two years Warranty / Guarantee is required.
- (d) Store on receipt will be accepted after inspection by EME section of PAC Hospital
- (e) Store must be latest production series (factory new).
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (h) Free of cost training on operation, maintenance and calibration of machine at PAC premises must be provided during warranty period.
- (j) OEM Conformance and Quality Assurance Certificates are required at the time of delivery.
- (k) Authorized distributor / dealer certificate of the equipment / store is required.
- (l) Country of origin and manufacturer certification FDA or CE or MHLW is required.

- (m) Product will be acceptable only with OEM standardized packing and labeled / endorsed Part Number, Batch / LOT No, MFG / EXP Dates. No local / supplier sticker is acceptable.
- (n) If calibration is required, newly calibrated item is to be provided by the vendor.
- (p) The repair / sales Centre of the equipment / machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.
- (q) At least ten years after sale spare supportability is to be provided.
- (r) Delivery period 08 to 12 weeks after signing of contract.

2. Sealed Technical & Commercial quotations with the title "**Tender for Electro Medical Equipment with Description**" must be submitted to the undersigned on **03 February, 2026** up to **1030 Hrs** positively. **5% Bid Security** must be present in the sealed Commercial Quotation in the name of Officer Commanding PAC Hospital Kamra. The photocopy of **5% Bid Security** with black out value column should be present in the sealed Technical Quotation to ensure that original **5% Bid Security** is present in the sealed Commercial Quotation. Tenders would be opened at **1100 Hrs** on the same date.

3. **Only one offer** can be made in relevant currency for same item and quoted price must be in **PKR**. Multiple rates, if quoted, will be rejected. Multiple quotations against the tender will also be rejected. Validity of quotation must be minimum 90 days from opening date.

4. Previous performance of the suppliers will be considered while accepting / rejecting the tender. Demonstration of quoted unit may be taken before technical scrutiny finalization if deemed necessary by the concerned specialist.

Officer Commanding
PAC Hospital Kamra (Attock)
051-909902352

Syed Mohsin Hassan
Assistant Director (Web Design)
Thursday, 15 January, 2026, 2:34:39 PM