ANNUAL PROCUREMENT PLAN FOR PURCHASE OF MACHINERY IN SCHEME FOR STRENGTHENING OF SINDH GOVERNMENT QATAR HOSPITAL, ORANGI TOWN, KARACHI

FOR THE FINANCIAL YEAR 2025-26 & 2026-27

SINDH GOVERNMENT QATAR HOSPITAL, ORANGI TOWN, KARACHI

S.No	Description of Procurement	Required Quantity	Estimated Total Cost	Source of Funds ADP/ Non ADP SNE	Proposed procurement Method	Time of Procurement /Tentative				Remarks/ Method
						1st Quarter	2nd Quarter	3rd Quarter	4th Quarter	E-bids through IE-Pak Acquisition and Disposable System (EPADS)
1.	Procurement / Supply and Installation of Machinery & Equipment	as Per Required	PKR 97,080,000/	ADP Scheme 901/2025-26	Single Stage Two Envelope				✓	Yes

MEDICAL SUPERINTENDENT
SINDH GOVERNMENT QATAR HOSPITAL,
ORANGI TOWN, KARACHI.

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI


DAILY EXPRESS
روزنامہ
ایکسپریس
کراچی
پاکستان کے 11 شہروں سے بیک وقت شائع ہونے والا واحد اخبار
جلد 28 شمارہ 228 | جمعہ 6 ذیقعد 1447ھ 24 اپریل 2026ء صفحات 12 قیمت 40 روپے

سندھ گورنمنٹ قطر اسپتال اورنگی ٹاؤن کراچی

410 بستروں کا ٹیچنگ اسپتال کراچی

فون:76-3475-9933-92-21+، فیکس:3485-9933-92-21+

نمبر SGQH/1903 مورخہ 21.04.2026

ٹینڈر طلبی نوٹس

سندھ گورنمنٹ قطر اسپتال اورنگی ٹاؤن، کراچی، رول(2)46 سنگل اسٹیج۔ٹو اینویلپ طریقہ کار کے تحت برائے مالی سال 26-2025 اور 27-2026 اے ڈی پی اسکیم نمبر 901/2025/26 (سندھ گورنمنٹ قطر اسپتال، اورنگی ٹاؤن، کراچی کی پختگی) کے تحت میڈیکل ایکوپمنٹس، مشینری، انسٹرومینٹ، فرنیچر/فکسچر کی فراہمی کیلئے ایف بی آر/ایس آر بی کی فعال ٹیکس دہندگان فہرست پر موجود (جہاں قابل اطلاق ہو) سپلائرز/فرموں/ مینوفیکچررز/امپورٹرز/مجاز ڈسٹری بیوٹرز سے ای۔پاک ایکوئزیشن اینڈ ڈسپوزیبل سسٹم (ای پی اے ڈی ایس) کے ذریعے ای۔پیشکشیں طلب کرتا ہے۔مفصل قواعد وضوابط پر مشتمل بڈنگ دستاویزات https://portalsindh.eprocure.gov.pk/#/ پر دیکھی/ڈاؤن لوڈ کی جاسکتی ہیں۔

الیکٹرونک پیشکشیں صرف ای پی اے ڈی ایس کے ذریعے جمع کرائی جانی چاہئیں۔مینوئل پیشکشیں وصول نہیں کی جائیں گی، دلچسپی کے حامل بولی دہندگان کو الیکٹرونک پیشکشیں جمع کرانے کیلئے لنک https://sindh.eprocure.gov.pk/#/supplier/registration پر ای پی اے ڈی سسٹم پر اپنے آپ کو رجسٹر کرانا درکار ہے۔

بڈنگ دستاویزات میں دی گئی ہدایات کی مطابقت میں تیار کی گئی پیشکشیں 11.05.2026 بوقت دن 11:00 بجے ای پی اے ڈی ایس پر لازمی جمع کرادینی چاہئیں۔ٹینڈر فیس -/Rs.3000 کا اصل انسٹرومینٹ اور ہر آئٹم کے مقابل درج مقررہ بڈ سیکورٹی، ای۔ پیشکشیں جمع کرانے کی آخری تاریخ سے قبل پروکیورنگ ایجنسی کے پاس لازمی پہنچ جانے چاہئیں، جو اسی روز بوقت دوپہر 12:00 بجے فاطمہ انجینئر آڈیٹوریم، جناح سندھ میڈیکل یونیورسٹی (جے ایس ایم یو)، کراچی کے پتے پر کھولی جائیں گی۔

نوٹ:

- ای۔بڈنگ سے متعلق کسی بھی قسم کے استفسار کیلئے سندھ گورنمنٹ قطر اسپتال اورنگی ٹاؤن، کراچی بمقام دفتر میڈیکل سپرنٹنڈنٹ میں بذات خود رابطہ کیا جاسکتا ہے۔
- حکومت کی جانب سے عام تعطیل کے اعلان یا کسی ناگزیر وجوہات کی صورت میں، ٹینڈر/ پیشکشیں اسی وقت اور مقام کیساتھ آئندہ کام والے دن جمع کرائے اور کھولے جائیں گے۔
- خریدار، ایس پی پی رولز 2010 کی متعلقہ شقوں کے تحت کسی/تمام پیشکشوں کو مسترد کرنے کا حق محفوظ رکھتا ہے۔

کسی بھی قسم کی دشواری کی صورت میں، ممکنہ بولی دہندگان ایام کار/اوقات کے دوران ای پی اے ڈی ایس ہیلپ لائن 051-111-137-237 پر رابطہ کرسکتے ہیں۔

دستخط

(ڈاکٹر راشد سراج)

میڈیکل سپرنٹنڈنٹ

سندھ گورنمنٹ قطر اسپتال، اورنگی ٹاؤن

کراچی

INF/KRY/1558/2026

i WORK FOR SINDH
www.iwork4sindh.com
JOB PORTAL BY
INFORMATION DEPARTMENT

SINDH GOVERNMENT QATAR HOSPITAL ORANGI TOWN, KARACHI

NOTICE INVITING TENDER

Sindh Government Qatar Hospital, Orangi Town, Karachi invites E-Bids through E-Pak Acquisition & Disposable Systems (EPADS) from Suppliers / Firms / Manufacturers / Importers / Authorized Distributors on Active Taxpayer List of FBR / SRB (whichever is applicable for the supply of Medical Equipment's, Machinery, Instruments, Furniture / Fixture under ADP Scheme No. 901/2025/26 (Strengthening of Sindh Government Qatar Hospital, Orangi Town, Karachi) for the Financial Year 2025-26 & 2026-27 under Rule-46(2) Single Stage Two Envelope Procedure. Bidding documents containing detailed terms & conditions, can be viewed / downloaded from http://portalsindh.gov.pk/#/

Electronic bids should be submitted through EPADS ONLY. Manual bids shall not be received. Interested bidders are required to register themselves on EPADS System at the link http://sindh.eprocure.gov.pk/#/suppliers/registration for submission of electronic bids.

The bids prepared in accordance with the instructions in the bidding documents must be submitted on EPADS by 11-05-2026 at 11:00 A.M. The original instrument of tender fee Rs. 3000/= and fixed bid security mentioned against quoted items must reach the procuring agency before the deadline for submission of E-Bids, which will be opened on the same day at 12:00 Noon at the address at Fatima Engineer Auditorium, Jinnah Sindh Medical University (JSMU), Karachi.

N.B:-

- Any query for E-bidding may contact at Sindh Government Qatar Hospital, Orangi Town, Karachi at the Office of Medical Superintendent in person.
- In case Government announce any Public Holiday or any unfavorable circumstances, the tender/ bids will be submitted and opened on next working day, with same venue and time.
- The purchaser reserves the right to reject any/ all bids under the relevant provisions of SPP Rules 2010.

In case of any difficulty prospective bidders may contact EPADS helpline 051-111-137-237 during working days / hours.

Sd/-
(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

INF-KRY
1558/2026

WORK FOR SINDH
www.work4sindh.com
JOB PORTAL BY INFORMATION DEPARTMENT

GOVERNMENT OF SINDH
HEALTH DEPARTMENT

NOTIFICATION

No.SO(PM&I)2-1/2022-23(Main)/PC: In supersession of this Department's Notification of even number dated: 20.07.2022, 18.10.2023, 03.11.2023 & 04.01.2024 and with the approval of Hon'ble Chief Minister Sindh vide CMIT Summary, the Health Department's Procurement Committee (PC) for procurement of Drug/Medicines, Surgical/Disposable items, X-Ray Films / Chemicals, Machinery Equipments, Furniture / Fixture (etc.) is hereby re-notified as under:

1.	Professor Dr. Amjad Siraj Memon, Vice Chancellor, Jinnah Sindh Medical University (JSMU), Karachi.	**Chairman**
2.	Vice Chancellor, Liaquat University Medical & Health Sciences (LUMHS), Jamshoro.	Member
3.	Vice Chancellor, Peoples University of Medical & Health Sciences for Women, Nawabshah-Shaheed Benazirabad.	Member
4.	Vice Chancellor, Shaheed Mohtarma Benazir Bhutto Medical University Larkana.	Member
5.	Prof. Dr. Sajida Qureshi, Professor of Surgery, DUHS / Dr. Ruth K.M. Pfau CHK.	Member
6.	Chairman, Health Care Commission Sindh.	Member
7.	Special Secretary, Chief Minister's Inspection, Enquiries & Implementation Team Department, Govt. of Sindh.	Member
8.	Additional Secretary (GA), SGA&CD, Govt. of Sindh.	Member
9.	Additional Secretary (PM&I), Health Department, Govt. of Sindh.	Member / Secretary

TOR's:

a) Preparing and /or reviewing bidding documents;
b) Carrying out technical as well as financial evaluation of the bids;
c) Preparing evaluation report as provided in Rule 45;
d) Making recommendations for the award of contract to the Competent Authority.
e) Monitoring of the procured medicines with regard to their quality and genuiness.
f) The Committee shall be authorised to have vigilance on distribution of the supply to different Hospitals including all health facilities.
g) The Committee would be empowered to Co-opt. Expert(s) incase the committee desires any assistance in any specific/relevent fields.
h) The committee shall have all the powers to procure the medical/surgical equipments, machines and furniture or any other item to be installed and / or fixed within the Hospital in Sindh province at any level whether Tertiary, District & Taluka Hospitals.
i) The powers of the Procurement Committee would extend to all the projects under the Health Department for procurement of medicines, furniture, instruments & machines etc.
j) Perform any other function ancillary and incidental to the above;

SECRETARY HEALTH

No. SO(PM&I)2-1/2022-23(Main)/PC: Karachi, dated: 29th January 2024

C.C. to:

1. The Chairman and all Members of the Committee.
2. The Accountant General Sindh, Karachi.
3. The Managing Director, Sindh Public Procurement Regulatory Authority, Karachi.
4. The PS to Chairman, CMIE&ITD Karachi.
5. The PS to Chairman, Health Care Commission Sindh.
6. The PS to Special Secretary, CMIE&ITD, Karachi.
7. The Deputy Secretary (Staff) to Chief Minister Sindh / Chief Secretary Sindh.
8. The PS to Minister Health, Social Welfare and Public Health Engineering Dept. Sindh, Karachi.
9. The PS to Secretary Health, Govt. of Sindh, Karachi.
10. The PS to Addl. Secretary (PM&I), Health.

29/01/2024

(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH PM & I CELL HEALTH DEPARTMENT

GOVERNMENT OF SINDH
HEALTH DEPARTMENT
(PROCUREMENT MONITORING & INSPECTION CELL)

NOTIFICATION

No. SO(PM&I)2-1/2022-23/PC(TEC-III): In supersession to this department's notification of even number dated: 13.10.2023, on recommendation of Health Procurement Committee (HPC) and with the approval of worthy Secretary Health, the Experts Committee for Machinery / Equipment / Lab. Equipment / Furniture & Fixture etc. is hereby re-constituted with the following composition and ToRs:

i.	Dr. Shahid Pervez, Histopathologist,Aga Khan University Hospital Kyc.	Chairman
ii.	Dr. Benish Sultan, Associate Professor & Head of Pathology Lab, JSMU, Karachi.	Member
iii.	Prof. Dr. Tariq Mehmood, Representative of PAF, JPMC, Karachi.	Member
iv.	Dr. Syed Farjad Sultan, Associate Professor, Anaesthesia Department, DUHS / Civil Hospital Karachi.	Member
v.	Prof. Dr. Sajida Qureshi, Professor of Surgery, DUHS / Dr. Ruth K.M. Pfau Civil Hospital Karachi.	Member
vi.	Prof. Dr. Syed Mahmood Hassan, Professor of Pathology, JSMU, Karachi.	Member
vii.	Dr. Shoaib Malik, Associate Professor, Anesthesia Department, JPMC Karachi.	Member

TORs:

a) To revisit and approve proposed specifications of Machinery / Equipment etc. prepared by the Technical Committee.
b) To approve bidding documents prepared by Technical Committee along-with verification / certification of technical evaluation reports finalized by Technical Committee.
c) The committee will also check the samples as per given specification in the tender document (wherever applicable).
d) To submit final recommendation to the Procurement Committee within seven (07) days for approval.

SECRETARY HEALTH

No. SO(PM&I)2-1/2022-23/PC(TEC-III): Karachi, dated the 18th February 2025

A copy is forwarded for information & necessary action to:-

1. The Chairman, Health Procurement Committee / VC, JSMU, Karachi.
2. The Chairman and all members of the Experts Committee.
3. PS to Minister Health, Social Welfare and Public Health Engineering Dept. Sindh, Karachi.
4. PS to Secretary Health.

18/2/2025

(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH
PM & I CELL
HEALTH DEPARTMENT

GOVERNMENT OF SINDH
HEALTH DEPARTMENT
(PROCUREMENT MONITORING & INSPECTION CELL)

NOTIFICATION

No. SO(PM&I)2-1/2022-23/PC(TEC-II): In supersession to this Department's Notification of even number dated: 15.08.2022, on the recommendation of the Chairman, Health Department's Procurement Committee, Technical Experts Committee for Machinery / Equipment is hereby re-constituted comprising on following, to assist the Health Procurement Committee for preparation of bidding documents and evaluation of tender of companies / bidders for purchase of Machinery / Equipment / Furniture etc. for the year 2022-23:

a.	Mr.S.Zamir Hussain Shah,Bio Medical Engineer,Dr.Ruth K.M.Pfau Civil Hosp. Kar.
b.	Mr. Abdul Majid, Electrical Engineer, NICH, Karachi.
c.	Mr. Qazi Muhammad Amir, Bio-medical Technician, Elector-medical Workshop.
d.	Mr. Sunny Mohiuddin, Dy. Director (Procurement),Directorate of Pharmacy,Sindh, Kar.
e.	Mr. Rajeev, Pharmacist, Dr. Ruth K.M. Pfau Civil Hospital Karachi.

ToRs:

i. To prepare bidding documents, Specifications and Evaluation Criteria.
ii. To review existing formulary list and segregate into seven (07) Therapeutic Groups and submit to technical experts / leading professors of respective therapeutic groups for approval.
iii. To carry out scrutiny of submitted bidding documents.
iv. To prepare Bid Qualification Report (BQR), Bid Evaluation Report (BER) and submit to respective therapeutic groups experts for review and approval.
v. To prepare financial comparative statement and final Bid Evaluation Report (BER) and get endorsement from respective experts of therapeutic groups before final approval by Health Department Procurement Committee.
vi. To submit scrutinized working to Experts Committee constituted for Machinery / Equipment / Furniture / Fixture & Lab Equipment etc.

SECRETARY HEALTH

No. SO(PM&I)2-1/2022-23/PC(TEC-II): Karachi, dated the 01st February 2023

A copy is forwarded for information & necessary action to:-

1. The Chairman, Health Procurement Committee / VC, JSMU, Karachi.
2. The Executive Director, NICH, Karachi.
3. The Medical Superintendent, Dr. Ruth K.M.Pfau Civil Hospital Karachi.
4. All members of the Technical Experts Committee (Machinery / Equipment).
5. The P.S. to Secretary Health.

01/02/2023
(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH PM & I CELL HEALTH DEPARTMENT

GOVERNMENT OF SINDH
HEALTH DEPARTMENT

NOTIFICATION

No.SO(PM&I)2-1/2013/CRC): In supersession of this department's notification of even number dated: 09.01.2023 and in pursuance of Rule-31 of Sindh Public Procurement Rules 2010, Health Department, Govt. of Sindh is pleased to re-constitute Complaint Redressal Committee (CRC) comprising of following officers for scruitinizing the complaints of aggrieved bidders against tender invited by the Health Procurement Committee, Health Department under Frame Work Contract System and Health Institutions/Hospitals of Sindh Province :-

1)	Secretary Health, Govt. of Sindh	Chairman
2)	Representative of Accountant General Sindh, Karachi.	Member
3)	Independent professional from the relevant field.	Member

ToR's:

- To scrutinize the complaints received from the aggrieved bidders and decide the same strictly in accordance with SPP Rules 2010.

SECRETARY HEALTH

No.SO(PM&I)2-1/2013/CRC): Karachi, dated, the 10th July, 2024

Copy forwarded for information and necessary action to :

- The Chairman and all members of Health Procurement Committee.
- The Accountant General Sindh, Karachi.
- The Managing Director, Sindh Public Procurement Regulatory Authority Karachi.
- The Medical Superintendents (All).
- The District Health Officers (All).
- The Directors Health Institutions / Projects (All).
- The Chairman and all members of Complaint Redressal Committee.
- PS to Minister Health & Population Welfare Deptt., Govt. of Sindh, Karachi.
- PS to Secretary Health Govt. of Sindh, Karachi.
- PS to Additional Secretary (PM&I), Health Department, Govt. of Sindh, Karachi.

10/7/2024

(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH
PM & I CELL
HEALTH DEPARTMENT

OFFICE OF THE MEDICAL SUPERINTENDENT
SINDH GOVT QATAR HOSPITAL ORANGI TOWN KARCAHI
HEALTH DEPARTMENT, GOVERNMENT OF SINDH

STANDARD BIDDING DOCUMENT

PROCUREMENT OF MEDICAL EQUIPMENT, MACHINERY, DEVICES, EQUIPMENTS AND MISCELLANEOUS ITEMS FOR FOLLOWING ADP SCHEME

S#	ADP No#	NAME OF SCHEME
1	901	PROCUREMENT OF MEDICAL EQUIPMENT & INSTRUMENT RE-REVISE/MODIFIED STRENGTHENING OF SINDH GOVERNMENT QATAR HOSPITAL ORANGI TOWN KARACHI

FOR FINANCIAL YEAR 2025-26 & 2026-27

Due on: __-__-2026 at 11: 00 AM @ **Fatima Engineer Auditorium of Jinnah Sindh Medical University (JSMU), Karachi before the Chairman Central Procurement Committee.**

(Single Stage – Two Envelope Procedure)

("Most Advantageous Bid")

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

SECTIONS OF THE TENDER

TENDER TITLE
SECTION A SUPPLY, INSTALLATION OF MEDICAL MACHINERY/EQUIPMENTS **SECTION B** DEVICES, EQUIPMENTS AND MISCELLANEOUS ITEMS

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

TABLE OF CONTENT

S. No.	DESCRIPTION	PAGE
1.	Salient Feature	
2.	Terms & Conditions	
3	Documents Checklist	
4.	Evaluation Criteria & Bidder Details	
5.	Technical Specifications of Medical Equipment	
6	Human Resource Including Detail of Technical Team (Form A)	
7	Major Spare Parts (Form C)	
8	Compliance Sheet/Technical Evaluation Sheet (From D)	
9	Financial Evaluation Sheet (Form E)	
10	Manufacturer Authorization Form (Form F)	
11	List of Installation/ Reference Client (Form G)	
12	Performance Certificate (Form H)	
13	Bid Letter Form	
14	Contract Agreement Form	
15	Bank Guarantee Form	
16	Integrity Pact	
17	Bid Form & Price Schedule	
18	Schedule of Requirements	

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

SALIENT FEATURES OF THE TENDER

1.	Name of Work & Address	PURCHASE OF MEDICAL EQUIPMEN/MACHINERY FOR **Office of the Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.**
2.	Time & Date of issue of Tenders	From the date of publishing to one day before opening of Technical Proposal, during office hours. It should be (As mentioned in Notice inviting tender).
3.	Place of Issuance of Tender	**Office of the the Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.**
4.	Date of pre-bid meeting	In case requested by bidders or informed byProcuring Agency. Mentioned pre-bid meeting date if any applicationreceived considered for discuss same mentioned date
5.	Method of opening of Tender	Single Stage – Two Envelopes; as per SPPRA-Rules2010 (Amended up-to-date) on online EPAD system.
6.	Date & Time of Submission of Tenders	As published in newspapers It should be (As mentioned in Notice inviting tender).
7.	Date & Time of Opening of Tenders / Technical Proposals	As published in newspapers on EPAD system. It should be (As mentioned in Notice inviting tender).
8.	Date of Opening of Financial Proposals	As informed by the Chairman of ProcurementCommittee to all participants
9.	Venue of submission & opening ofTenders	Vice Chancellor, Jinnah Sindh Medical University(JSMU), Karachi.
10.	Validity of Tenders	90 days as per Sindh Public Procurement Rules, 2010 (Amended till date).
11.	Amount of Earnest Money/Bid Security	Fixed bid security on each quoted item amount in shape of Pay Order / Call Deposit from schedule Bankin favor of **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi..**to be submitted along with Financial Proposal and also attached photo copy without showing amount of that pay order else the offer will be rejected.
12.	Supply of Equipment	Within up to 16 weeks after the Purchase order in case of DDP. However, this period is extendable asper mutual agreement/demand of the situation.
13.	Installation Period	It will start after receiving of equipment at site. It will start after satisfactory installation of equipmentat site.
14.	Maintenance Period	It will start after the satisfactory installation of the equipment and issuance of Installation Certificate from the concerned official. This period will remain functional till three years with replacement of parts.
15.	Warranty Period with replacement ofParts and services	As mentioned in technical specification of the equipment, otherwise the warranty period of parts with Free of cost replacement will be for three years.
		Along with three years free service & maintenance from the date of Installation.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT GOVT. QATAR HOSPITAL ORANGI TOWN KARACHI

16.	(a) Contract Agreement Stamp Paper requirement for Agreement	The Contractor shall enter & execute a formal Agreement as per the "Form" annexed with such modification as may be necessary. - Rs. 0.35% of the Contract Value or as prescribed by Government Laws.
	Opening of L.C (Letter of Credit	Letter of Credit (L.C.) will be opened by Procuring Agency as consignee and as per the prevailing rule of same. Performa invoice of the quoted amount on C&Fbasis of the country of origin along with insurance cover note if applicable will also be submitted by the contractor(s). Bank charges in establishment of Letter of Credit (L.C.) fluctuation in currency if any and other miscellaneous charges on logistics, transportation clearing from sea port/ airport will also be paid/bear by the contractor(s).
17.	Release of Earnest Money/Bid Security	To un-successful bidders, after financial bid is openedor technical bid is rejected. Earnest Money will be released to successful Bidder after purchase order is released but after 5% Security Deposit submission.
18.	Security Deposit	5% Security Deposit in the shape Pay Order/Bank Guaranty will be submitted by Contractor with Procuring Agency which will be released as per item20.Bank Guarantee of equal amount will also be acceptable. Security Deposit is deposited as per item – 19.
19.	Release of Security Deposit	After completion of installation and maintenance ofDefect Liability period.
20.	Variation in Contract Price	No variation in price shall be allowed on any ground including currency fluctuation/variation or whatsoever.
21.	Discrepancy	If there is any discrepancy between Salient Featuresof Tender /Bidding Documents and respective contents mentioned elsewhere, Salient Features / Terms & Conditions and Additional Terms & Conditions of Tender will govern.
22.	Taxes	All taxes will be deducted as per prevalent laws ofCountry.
23.	Approved makes	As mentioned in items specifications.
24.	Special Note Regarding Equipment	Technical offers / Commercial offers failing to demonstrate below details would be rejected: a. Quoted system must be of advanced & latestversion.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT SINDH GOVT. QATAR HOSPITAL ORANGI TOWN KARACHI

		b. Tender must cover complete equipment. c. Tender must cover complete range of disposables/ kits d. Tender must provide all technical details up to the satisfaction of the end user. Items should be quality approved from the concerned international body of the respective industry.
25.	(a) Inspection of Imported equipment manufacturing site by the client.	After the award of Purchase Order, Supplier shall taketwo persons to the manufacturing site and bear all expenses of visit in case the cost any equipment of Supply Order is 50 million or above.
	(b) Training	Supplier will provide on-site successful training to allthe personnel working on/operating the said Equipment / Machine as long as the need prevail.
26.	Maintenance.	Maintenance cost for all items for Two years from thedate of successful Installation shall be undertaken by the Contractor (Maintenance includes all Parts & Labor, etc. with sufficient staff, during maintenance period).
27.	Default in Preventive Maintenance, Breakdown and Emergency Calls.	• In case of default by the Contractor with respect to maintenance, break down and emergency calls, the same will be carried out within 24 hours by the Procuring Agency and the cost so incurred will be paid from the Retention Money. • Moreover an additional 10% of the amount spent would be charged from the concerned contractor being defaulter.
28.	Cost to be quoted in BOQ.	• The Contractors shall quote price in **DDP only,** ofthe Equipment's including custom clearance etc. and delivering of Equipment at site with Installation & Commissioning cost. • It should also include cost during maintenance period including the parts and maintenance staff.
29.	Tax Exemption.	In case of purchase on C&F basis, the required certificate will be issued by this office that the importhas been made for this hospital, so as to avail the facility of exemption of duties/taxes, as per Government Rules/Policy according to the Sindh Public Procurement Rules, 2010 (Amended till date). Clearance from the custom and other charges will be the responsibility of bidder.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SIDH GOVT. QATAR HOSPITAL
LANGI TOWN KARACHI

30.	Criteria for bid evaluation.	Merit Point Evaluation **(Most Advantageous Bid)** The items ranked highest in merit points (obtained throughand based on technical and financial evaluation) shall be considered as most advantageous bid.
31.	Award of Contract	The Procuring agency will award the contract to the successful Bidder whose bid has been determined to be substantially responsive and has been determined to be the **Most Advantageous Bid,** provided further that the Bidder is determined to be qualified to perform the contract satisfactorily.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

TERMS & CONDITIONS OF TENDER

1. The Contractors responsible for **Supply, Installation, Testing and Commissioning of Medical Equipment, Machinery items , Instruments & Furniture & Fixture** while Installing Equipment at no extra-cost. So, the Contactors should take into account all such expenditures while quoting the bid.
2. **In Technical Bid**, the bidder must provide original data sheet, technical brochure, and all other relevant documents, along with copy of Pay Order / Bank Draft / Bank Guarantee without showing the amount figure otherwise, the bid will be ignored/rejected. The technical evaluation will be done on the basis of criteria given in tender document.
3. The vendor has to fill up **compliance sheet** with YES/NO or put value as desired in the column against provided specification. These specifications must be highlighted in the brochure attached; else the offer will be rejected.
4. **In Financial Bid**, the vendor should mention financial offer along with Scanned Copy of Pay Order / Bank Draft / Bank Guarantee Fixed bid security on each quoted item amount in favor of, **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi. Bid Security (in Original) must be submitted in sealed envelope to the procuring Agency before opening of tender.**
5. The Supply order shall be placed on DDP basis the bidder should supply the items within time as specified in Tender Document i.e. 16 weeks. This period is extendable as per mutual agreement/demand of situation.
6. The disclosure of firm's price at the time of opening of Technical bid will result in the rejection of the bid.
7. No tender will be entertained without earnest money. The earnest money will be forfeited to Government Treasury, in case of non-submission of security money within **seven (7) days** of receipt of the offer letter.
8. Only Manufacturer, its Sole Agent or Authorized Distributors can quote Device, Equipment / Machinery. No sub-distributor is allowed to quote.
9. **For Single Stage Two Envelope Sections of the tender**, first scrutiny of Technical Bids will be performed by the Procurement Committee. Financial Bids of only those firms will be opened who are qualified technically.
10. Certificate from the Sole Agent/Manufacturer has fully equipped Workshop Trained Personnel/ Engineers and Equipment should be provided in the technical bid, and it can be physically verified by the administration of **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.**, The warranty period will commence from the date of installation of the Equipment at **Office of the Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi..**
11. Firm must provide complete details of their financial standing, listing of similar items supplied in Pakistan, available local stock of items and details of office with number of personals.
12. The vendor has to quote only one rate for each item as per tender specifications. Hand written tenders or any over writing, cutting, should be signed.
13. Vendors should quote the items they intend to supply.
14. The bidder shall submit with the bid, an **Undertaking** on Non-Judicial stamp paper of Rs.50/- duly attested by Oath Commissioner, stating that the medical equipment will be supplied brand new, free from defects as per the approved specification, also provide two years warranty for service and maintenance and will bind himself for availability of spares for a period of **10 years** from the date of supply. In case of winding

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID RAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

up of his business, the supplier will be bound to make an alternate arrangement of spares and maintenance of medical equipment. In case of default, depreciated cost of medical equipment will be recoverable from vendor.

15. If it has been found that the information submitted by the bidder regarding his qualification and professional, technical, financial, legal or managerial competence as supplier / contractor is / was false and materially inaccurate or incomplete at any stage, then the bidder will be disqualified.
16. Suppliers / Contractors will be responsible for supply at consignee end, free of cost and also responsible for free installation work, if required.
17. Equipment will be handed over to the Store Keeper will be handed over to Store Keeper **Office of the Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.** (Price escalation will not be allowed.)
18. If vendor doesn't supply medical equipment in accordance with the supply order or fails to supply according to terms and conditions or in case of any other default, the Performance Security and Security Deposit would be forfeited and further action as per SPPRA rule would be taken against such vendor.
19. If the supplier fails to give supply and install within the stipulated period, Liquidated charges will be imposed.
20. The decision once taken will be final and will not be challenged in any Court of Law.
21. A copy of Income Tax and GST Registration Certificate should be attached along with Technical Bid.
22. The bidding firm must be registered with PNRA department for minimum past three years (for Radiology items).
23. The bidder shall furnish copy of ATL (Active Tax Payer List certificate).
24. Copy of valid Professional Tax (Excise & Taxation) Certificate, Income Tax Certificate and GST Registration Certificate should be attached with the technical bid.
25. The bidder shall furnish three years Income Tax Return Certificate.
26. Conditional Tender(s) against the Government rules and non-compliance of instructions mentioned in Tender Form & Hand written tender(s) will not be entertained.
27. The vendor should submit a **Data Sheet** according to the Tender Specifications Form as mentioned in the tender and highlight the tender specifications in their Technical Data Sheet. Additional Features canbe mentioned separately.
28. In case the information provided by bidder in Data Sheet as per **Clause No. 30** is incomplete or incorrect, the quoted item will be liable for rejection.
29. Supplier must have to provide Manual book with circuit diagram where applicable.
30. Supplier must have to arrange training on site for concerned technical staff for the supply of Equipment
31. Procurement Committee,, **Office of the Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.**
32. the right to reject or accept any / all tender(s) as per SPPR-2010 (Amended 2019).

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi** _____

MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

33. The system will have the minimum uptime 95% in case of down time beyond this period the following penalty will be applied:

    a) 95% to 100% No penalty.

    b) 90%to95%the warranty period will be extended by the number of days system in down.

    c) 85% to 90% warranty period will be extended 1.5 time the down time period

    d) 80% to 85% warranty period will be extended two time for the down time period and supplier is supposed to give justification of delay.

    e) Uptime less than 80% hospital will complain to the Principal to take action against the supplier and warrant\' will be the three time for the down time period.

    f) However, PPM (Periodic Preventive Maintenance) will not be considered in the downtime and PPM of the system will be done as per manufacturer recommendation.

34. Payment of DDP basis, the vendor shall submit Bill/Invoice for the purchase order value after delivery, installation and issuance of satisfactory report by the end user.

35. Bids shall remain valid for **90 days** after the date of bid opening and same may be extended in terms of Rule 38 (2) (3) (4) of SPPRA Rules.

36. Bid should be inclusive of all Government taxes (if applicable) and the same will be paid by the Contractor except withholding tax and 0.35% Stamp Duty which will be deducted at source in office of the Accountant General Sindh Karachi.

37. The bid security will be forfeited to the Government, if the bidder withdraws his bid after opening and before the expiry of the bid validity period or fails to sign the contract in stipulated time if the bid is accepted.

38. The Procuring Agency may reject all or any bid at any time prior to the acceptance of a bid or proposals, subject to the relevant provision of SPP Rules, 2010 (Amended 2019 up-to-date). Payment will be made within four weeks after receipt of bill / invoice duly fills in all respects.

39. Prices quoted shall remain valid up to 90 Days.

40. Registration from Sindh Revenue Board is not required in procurement of Goods.

I / We agree to above mentioned terms & conditions:
Name of Contractor_____Signature_____
(CNIC NO_____(Copy must be attached).)
Full Address _____
_____
Rubber Stamp _____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR RASHID SIRAJ) MEDICAL SUPERINTENDENT GOVT. QATAR HOSPITAL KORANGI TOWN KARACHI

SECTION A

Supply, Installation, Testing and Commissioning of Medical Equipment and Machinery Items

DOCUMENT CHECKLIST

Please review the following list of all documents to be enclosed with the Technical Proposal. These are the "MANDATORY DOCUMENTS" non-submission of any one of the mandatory documents will lead to disqualification and no further assessment of tender will not be done hence "TECHNICALLY REJECTED"

S.NO	DOCUMENT DESCRIPTION	YES / NO	PAGE #
1	Tender Purchase Receipt (Original) (Mandatory)		
2	Bid Security (Copy with value hidden in Technical Proposal; Original in Financial)		
3	Bidding Documents (Duly filled, Signed & Stamped by Bidder)		
4	Technical Proposal on Bidder's Letterhead		
5	Compliance sheet on letter head (Mandatory)		
6	Undertaking (as per sample form)		
7	Certificate (as per sample form)		
8	Income Tax & GST Registration Certificate		
9	Professional Tax Certificate (Sindh)		
10	SECP Incorporation Certificate (where applicable)		
11	PNRA Registration Certificate (where applicable)		
12	PEC Registration of Firm & Qualified Engineer's Certificate (whereapplicable) (Mandatory for Service works)		
13	DRAP Establishment license showing Importer/Manufacturer of medical devices.		
14	DRAP Product Establishment /Registration		
15	Company Profile		
16	Bank Account Maintenance Certificate		
17	Income Tax Returns (last 3 years)		
18	Audited Financial Statement (last 3year)		
19	Workshop for After Sales Services (Mandatory)		
20	Technical Team Details (Workshop details & Technical Staff Bio data with Training Certificates)		
21	Manufacturer's Authorization for the sole distribution in Pakistan (Attested by the Embassy)		
22	Catalogue / Brochures (original)		
23	Product Technical Data Sheet		
24	ISO 9001of bidder (where applicable) **(website link must be providedfor verification)**		
25	ISO 13485 of Manufacturer (where applicable)**(website link must beprovided for verification)**		
	CE/ FDA510(K) / JIS / MHLW (where applicable)**(website link must beprovided for verification)**		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

Note:

- Mandatory documents are mentioned.
- All pages of the bid, except for un-amended printed literature, shall be initialed by the Bidder.

BIDDER'S DETAILS FOR NOTICE PURPOSES

Bidder Name:	
Address:	
Tel No.	
Fax No.	
Contact Person Name:	
Mobile No.	
Email Address:	

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
GOVT. QATAR HOS
NGI TOWN KAR

SECTION A

Supply, Installation, Testing and Commissioning of Medical Equipment and Machinery Items

Note:

EVALUATION CRITERIA

Financial proposal must be submitted on company letter head duly signed and stamped. Bidder is required to type their offer in figure and as well as in words of the total amount; else the offer would be rejected.

EVALUATION/QUALIFICATION

1. The procuring agency reserves the right to evaluate and compare the bids on itemized basis OR on basis of a group of similar nature goods OR goods compatible with each other.
2. The following merit point system for weighing evaluation factors/criteria will be applied for technical proposals.
3. Bidders achieving minimum 70% marks will be considered as qualified beside compliant of all mandatory clauses. Documentary evidence must be attached in support of each parameter.
4. Any Bid not meeting the mandatory requirements of evaluation criteria will be disqualified /rejected straight away and will not be considered for further evaluation.
5. Provide website link of product certification for online verification.

A. PRODUCT EVALUATION

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
1	**Conformity to the Purchaser's Specifications (Mandatory)**	**30**
1.1	Fully compliant with the required specifications	30
1.2	Compliant with minor deviation (up to 5% subject to main function is not affected)	25
1.3	Compliant with minor deviation (up to 10% subject to main function is not affected)	20
1.4	Non-compliant to required specifications	Disqualify
2	**Trained Product Specialist**	**10**
2.1	Two marks for each Foreign Technical Graduate Engineer with PEC Registration in Sindhfor the quoted product (Factory/OEM level service training).	06
2.2	Two marks for each Foreign Trained Science Graduate in Sindh for the quoted product (Factory/OEM level application training).	04
3	**Product Certification**	**10**
	• USA Food & Drug Administration (USA-FDA) • European Community (CE) MDD • Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labor and Welfare Govt. of Japan	
3.1	All Three of the above	10
3.2	Any two	7
3.3	Only One	3
	TOTALMARKSPRODUCTEVALUATION(A)	**50**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SIDH GOVT. QATAR HOSPITAL
NGI TOWN KARACHI

B. BIDDER EVALUATION

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
4.	**Manufacturer's Authorization(Mandatory)**	**10**
4.1	Bidder is OEM direct representative or the Sole Distributor for last 10 years or more. (Authorization certificate attested by embassy must be provided)	10
4.2	Bidder is OEM direct representative or the Sole Distributor for more than 05 years but less than 10 years. (Authorization certificate attested by embassy must be provided)	8
4.3	Bidder is OEM direct representative or the Sole Distributor for 05 years or less (Authorization certificate attested by embassy must be provided)	5
4.4	Non submission of valid manufacturer's authorization	Disqualify
5.	**Human Resource**	**06**
5.1	Diploma of Associate Engineer (DAE) in electrical / electronic / biomedical / mechatronics / mechanical / industrial. DAE certificate must be submitted. (***0.5 mark for each certificate***)	2
5.2	Graduate Engineer with PEC Registration in biomedical, electronics, and mechatronics, mechanical, industrial. PEC registration card of the engineer must be submitted. (***1mark for each Engineer***)	4
6.	**Networking**	**04**
6.1	Networking setup across Pakistan (1 mark for each setup) (verifiable) (proof of registered office and workshop must be provided)	04
7.	**Past Experience/ Performance**	**10**
7.1	Bidder's prior experience for supplying the **same nature of product.** Performance certificate of at least 05 public hospitals having minimum 5 years' Experience along with Installation Report and Supply order/ Purchase Order of the firm in last 5 years, onletter head, signed and stamped by the Head of institution of public sector. ***One mark for each after sale satisfactory performance certificate.***	05
7.2	Bidder's prior experience for supplying the **same nature of product.** Performance certificate of at least 05 reputable private hospitals having minimum 5 years' Experience along with Installation Report and Supply order/ Purchase Order of thefirm in last 5 years, on letter head, signed and stamped by the Head of institution of private sector. ***One mark for each after sale satisfactory performance certificate.***	05
8.	**Turnover total during last three (03) financial years(Audited Statements of Accounts and Income Tax Return Forms must be attached as supporting documents)**	**10**
8.5	Turnover last 3 years equal or above PKR 1200 million	10
8.4	Turnover last 3 years equal or above PKR 1000 million	08
8.3	Turnover last 3 years equal or above PKR 800million	06
8.2	Turnover last 3 years equal or above PKR 600 million	04
8.1	Turnover last 3 years below PKR 600 million	00
9.	**Bonus points**	**10**
9.1	Free of Cost Extended Warranty (in accordance with the standard warranty period) atleast for two year including parts & services & labor. (Bidder offering greater extended warranty period will get the full marks)	10
9.2	Free of Cost Extended Warranty (in accordance with the standard warranty period) at least for one year including parts & services & labor.	05
	TOTAL MARKS BIDDER EVALUATION (B)	**50**
	GRANDTOTAL (A+ B)	**100**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

BIDDER EVALUATION CRITERIA

Bid found to be responsive with maximum accumulative points (Technical + Financial) shall be accepted as the **mostadvantageous bid** as per SPPR, 2010 (Amended 2021).

Bids will be evaluated on the basis of following evaluation criteria.

A. TECHNICAL EVALUATION

i) Technical evaluation has 70% weightage

ii) Bids will be evaluated on a Bidder's capacity and capability to undertake supplies and on quoteditem quality, performance and conformity to the required specifications.

iii) Technical capacity of a bidder includes availability of workshop, tools, technical staff, number of installation and experience in the province of Sindh for prompt services to Karachi Division. Past performance of the bidder at **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.** will also be reviewed.

iv) A bidder must achieve a minimum of 70 points out of total 100 points in technical evaluation, tobe considered for financial evaluation.

B. FINANCIALEVALUATION

i) Financial evaluation has 30% weightage.

ii) Only financial proposal of bidders with responsive technical proposal will be opened. Financial proposal of bidders who do not meet required mandatory documents and minimum 70 points out oftotal 100 points in technical evaluation will be returned unopened.

iii) Bidder who has offered the lowest value for the item willget30points.

iv) All remaining bidders after the lowest shall get points on their financial proposal, calculated as per the following formula.

$$\frac{\text{Value of lowest bidder} \times 30}{\text{Value of higher bidder [2nd, 3rd, 4th, etc.]}}$$

C. FINAL BID EVALUATION

i) Points obtained by bidders in technical and financial evaluation shall be prorated and finalizedas follows.

Evaluation	Weightage
Technical	70%
Financial	30%
Total	**100%**

ii) Bid of a bidder who is found technically responsive with highest accumulated points will be considered as the most advantageous bid.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SIND GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

SECTION A

Supply, Installation, Testing and Commissioning of Medical Equipment and Machinery Items

S. No	Name of items	Qty.	FIXED BID SECURITY
1.	**Autoclave Digital** **Specification:** 1. The outer surfaces and the chamber of the device must be completely made of stainless steel 2. Sterilization basket made of stainless-steel material inside the device 3. The volume used in the device should be at least 50-60liters 5. The device should be able to perform fully automatic, pre-vacuum, steam filling, sterilization, steam discharge and final vacuum operations automatically. 6. Cover seal on the device must be provided with silicone-based gasket 7. Should have a lockable cover system that does not open when there is pressure inside 8. Double-walled device, steam generator should be placed between walls 9. The water required for reinforcing the steam of the steam generator is located in tank and the water level in this tank should be observable 10. The vacuum pump used in the device must be of liquid ring type 11. Fresh air intake into the cell during the drying process must be provided via the HEPA filter The pressure gauges of the sterilization chamber separately and on the panel (preferable) 13. Device should have PLC control system, sterilization temperature between 110 ° C and 140 ° C at 1 ° C intervals, sterilization time at intervals of 1 minute up to 99 minutes and final vacuum repeat number between 2 -9 14. On the control panel of the device control 'must be warning system for open, overpressure, heater failure and low water level acid' 15. In case the electricity is cut off again during the sterilization phase: If the temperature of the cabin is within ± 1 ° C limits, it should continue its operation and if the alarm exceeds the limits, it should warn the power failure visually and audibly. 16. The device must have both electrical (heater-stopping), mechanical(overpressure) safety system and safety valve against pressure **Country of Origin: IMPORTED/ EXPORT QUALITY.** **Should be: ISO-13485, CE/ ISO-14001 (Anyone).** **Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end.**	02	108000
2.	**DENTAL UNIT (COMPLETE WITH CHAIR, LIGHT, HAND PIECE UNIT HAND PIECES, SUCTION AND COMPRESSOR 4 HAND-PIECE** **Specification:** **Patient's Chair** Manually adjustable flat headrest. armrest Swiveling for entry and exit of patient Multiple profiles and positions 3 favorite buttons for individualized instrument settings Programmable Starting and Rinsing positions Extendable foot rest for ergonomic treatment of long patients Fully hydraulically self-adjusting Chair with lifting capacity of 200 KGS. Automatic Emergency stop of chair once it meets an obstacle. Manual Emergency stop also to be available	02	108000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

(DR. RASHID SIRAJ
MEDICAL SUPERINTEND
GOVT. QATAR HO
NGI TOWN KA

	Synchronized movement of backrest and seat cushion by electromechanical motors prevents construction/over extension of the patient. The position of the patient's head on the headrest remains constant when the chair is lowered. Easy adjustment of chair & backrest positions. Standard Pneumatic Foot switch for controls of hand pieces. All the outlet & inlet for the services to the chair should be Concealed in the box to be at the foot area of the chair, as an infection control Measure. **Dentist's Element** Dentist Element with hanging hoses and adjustable arm with pneumatic brake system Control buttons for the individual chair functions, Programs, Light on/off and Intensity as well as for spittoon bowl and cup filler. Hand piece holders without lockouts Number of Hand piece positions including 3 way syringe: 05 Hand piece priority system (if one Hand piece in use others will not operate while lifting from holder) 3x Hanging hoses, fiber optic type 1 x 3-way syringe with autoclave able tip 1x High Speed air turbine push button type with 400,000 rpm 1x Slow Speed Air motor with internal water spray. Speed range from 2000-40000 rpm. Forward and reverse modes. Complete with straight and contra angle attachments 1x Built-in Ultrasonic Fiber Optic Scalar frequency 28-36 kHz with sterilize-able hand piece and tips of same manufacturer. Water regulation on the scalar hand piece. Scaling and Endo mode Removable silicone tray pad Removable hand piece holder and tubing for cleaning and disinfection. **Assistant's Element** Number of hand piece positions including 3way syringe: 04 1 x 3-way syringe with autoclave able tip 1 x High Volume Evacuation hose, autoclave able suction hand pieces with adjustment suction pressure 1 x saliva ejector hose, autoclave able suction hand piece with adjustment suction pressure **Water Unit** Automatic Bowl Rinse function Fixed on chair base Automatic Glass Fill function with adjustment Clean water system with the facility to change from clean (water bottle) to tap water with toggle switch Standard Wet suction through micro switches (read to connect to central / wet suction) **Patient's Light LED Light** For optimum illumination of the treatment site Flexible positioning via three flexible joints Anti-glare for the patient through clearly confined luminous spot Brightness is 8,000 – 26,000 Lux, 5,000- 6,900 Kelvin **02x Working Stools** Anatomically contoured seat, mounted on 5-castors which can automatically braked when no load is applied to the stool. Installation and all accessories must be included Operator, service manual and spare parts catalogue must be provided in both hard and digital copy All consumables/spare required replacement within 10 years of normal operation must be quoted optionally. Must have all tool, instruments and calibration equipment in Karachi End-user application training and technical training of engineer to be provided locally by a trained specialist along with training material as per manufactures guideline **OR EQUIVALENT**		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSP
ORANGI TOWN KARAC

	Country of Origin: UK, USA, Europe, Japan or equivalent, (Any One Certificates of CE, FDA, MHLW) **Should be:** ISO-13485 (Mandatory) **Warranty:** Warranty period shall be of three years starting from date of satisfactory installation of equipment		
3.	**Dental X-Ray Machine** **Specification:** Digital Dental X-Ray Machine DC Type (Mobile) Compatible with Digital Sensors, PSP Image Plate Scanners and Traditional Films 100% imported Scissor arm from the original manufacturer Microprocessor controlled Tube Current 2-4mA or better Tube voltage 60-70 kV or better Exposure time automatic adjustable Patient types: Child and Adult Automatic Selection of dose by Tooth selection Timer with integrated circuit Tube head with 360° rotation and beam limiter Power on light signal. x-ray emission light signal Focal point 0.2mm or better Filtration system Power supply 220V/50Hz **OR EQUIVALENT** **Country of Origin:** UK, USA, Europe, Japan or equivalent, (Any Two Certificates of CE, FDA, MHLW) **Should be:** ISO-13485 (Mandatory) **Warranty:** Warranty period shall be of three years starting from date of satisfactory installation of equipment	**1**	**75000**
4.	**Multiparameter Patient Monitor** **Technical Specification** **1.** Monitor for Adult / Paeds / Neonates **2.** Operating Features and Characteristics: **3.** Screen Size: At least 15" or more diagonal (Touch Operated) or better with non-fade LCD / TFT LCD / LED color display **4.** Electro-Surgical Interference Suppression / Protection **5.** Defibrillator Protection **6.** Freeze and Cascade Facility **7.** Waveform: Six wave forms or more **8.** Waveform Traces Speed: 25 & 50 mm / sec Parameters: ECG Input: 3&5-lead ECG cable and standard AAMI line for connection Lead Choice: I, II, III, aVR, aVF, aVL, V, V1-V6, **2.** Numeric: Heart Rate **3.** Real time, and freeze ECG Trace Non-Invasive Blood Pressure (NIBP): **1.** Method: Oscillometric Principle **2.** Numeric: Systolic, Diastolic, and Mean Pressure **3.** Selectable auto inflate interval settings	**5**	**72000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTEND
SINDH GOVT. QATAR HO
KORANGI TOWN KA

	4. Rising Cuff / Continuous Pressure Display Invasive Blood Pressure (IBP): **1.** IBP Two or more Channels Pulse Oximetry: **1.** Numeric: 0 – 100% oxygen saturation measuring range **2.** Waveform-Plethysmograph Pulse with Pulse Strength Indication **3.** Probe preferable Masimo/Nellcor Arrhythmia Analysis: **1.** Arrhythmia Analysis and ST Analysis Respiration: **1.** Breath Rate display and settable apnea alarms **2.** Sweep Speed: 6.25, 12.5 mm / sec or better **1.** Trend Data: Graphical / Tabular with at-least 48 hours' or better **2.** Should have feature for the connectivity with the Central Monitoring station Alarms: 1. High & Low (settable) on all parameters 2. Visual and audible indication of alarms Communication: 1. Capability to interface with LAN / WLAN for data transfer Accessories: **1.** The system must be complete with all reusable (sensors, probes, cables, or any other accessories) required for measuring all the above selected parameters for Adult, Paeds and Neonates **2.** Reusable Cuffs (Latex Free) of all available sizes of all body contours Power Requirement: Line voltages: 100-240VAC Line Frequency: 50/60 Hz **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end.**		
5	**Drill Machine for Orthopedic Surgery.** **Specification:** 1 DRILL AND REAMER Speed max. Depending of attach. Cannulation 4.0mm Weight (incl. battery &lid)1000 - 1500g Attachments 2 REAMER ATTACHMENT JACOBS CHUCK REAMER ATTACHMENT with Jacobs chuck0.5-7.5 mm chuck 0-250 rpm or better;4.0 mm; 3 REAMER ATTACHMENT AO-LARGE REAMER ATTACHMENT	**1**	**282000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SHAH
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
GANGI TOWN KARACHI

	0-250 rpm or better;4.0 mm; 4 REAMER ATTACHM.HUDSON/ZIMMER REAMER ATTACHMENT 0-250 rpm or better;4.0mm; 5 Researchable Battery with 4hrs or more battery backup 6 LID 7 DRILL ATTACHMENT AO-SMALL DRILL ATTACHMENT 0-1,200 rpm or better;2.8mm; 8 DRILL ATTACHMENT HUDSON/ZIMMER DRILL ATTACHMENT 0-1,000 or better rpm;4.0mm; 9 UNIVERSAL K-WIRE ATTACHMENT K-WIRE ATTACHMENT For K-wires from 0.6-4.0 mm 0-1,200 rpm or better 10 CHARGING UNIT 11 Mains voltage ranges 220V~to240V~ **Country of Origin:** UK, USA, Europe, Japan or equivalent. **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Any One Certificate) (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificates. Warranty: Warranty period shall be of three years starting from date of satisfactory installation of equipment		
6.	**Surgical Diathermy Machine.** **Specification:** • Microprocessor based electrosurgical unit for normal and under water cutting usages. • Automatic power adjustments • Operation in radio frequency range. • Controls for cutting, coagulation, spray and blends. • Monopolar cutting power of 300 watts with change of 05 watts or with less increments. • Mono polar coagulation power of 200 Watts or more. • Bipolar Cutting power of 100 watts or more • Bipolar coagulation power of 120 Watts or more. • Spray coagulation power of 120 watts or more • Different gradations of blending of cutting and coagulation power. • Digital display of all controls and set values of cutting and coagulation power. • Audio and visual alarms. • Power Supply: 220V, 50 Hz. **Accessories:** • Monopolar handle with cord. • Bipolar forcep with cord. • Attachment for monopolar coagulation. • Knife electrode. • Surgical electrode, ball-shaped. • Wire loop electrode. • Needle electrode. • Ball electrode.	**5**	**216000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

	• Bipolar coagulation forceps. • Reusable patient plate. • Double paddle foot switch, explosion proof. • Trolley with lockable antistatic castors may be provided locally. **Country of Origin:** UK, USA, Europe, Japan or equivalent. **Should be:** ISO-13485 (Mandatory) & CE, FDA, MHLW (Any One Certificate) (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificates. Warranty: Warranty period shall be of three years starting from date of satisfactory installation of equipment		
7.	**Echo Machine** **Specification:** Color Doppler with Fully Digital Beam former, having 2D / M-Mode and Doppler Facilities, (PW, HPRF, & Color Flow Imaging) with High Resolution Imaging Doppler Signal Quality; Receiving / having DICOM compatibility **Imaging & Technology** • **Imaging Modes:** 2D, 3D, 4D Imaging, Color Doppler, Pulsed Wave Doppler, Strain Elastography • **Image Optimization:** iBeam, HD-Scope, Glazing Flow, NTE iClear, UWN+ Contrast Imaging • **Automatic Adjustments:** iTouch+ for real-time dynamic image optimization • **Specialized Applications:** o Smart OB: Automatic fetal biometrics measurement o Smart Face: Automatic removal of occlusions in fetal face imaging o Smart HRI: Hepatic steatosis evaluation o Smart Bladder: Automatic bladder volume measurement o Auto EF: Automatic ejection fraction measurement based on diastole/systole frames o RIMT: RF-data-based real-time IMT measurement **Ergonomic Design & Hardware** • **Display & Controls:** o 21" High-resolution LED monitor Resolution: 1920x1080 or above, o Viewing angle>=176 degrees, Tilt: 105°degrees or better. o Down/up with lifting arm range: 300±20 mm o Rotate angle range: 180 degrees o 13 inch high sensitivity anti-glare color touch screen Resolution: 1280 x 1080 or above. o Touch screen panel angle adjustable for easy visualization: 50 degrees in rotation o Fully adjustable floating control panel o Waterproof control panel for easy disinfection o Rotatable control panel for ergonomic positioning o Noise-reduced design for quiet operation o Ergonomic footrest for user comfort **Imaging Mode:**	**1**	**360000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSP
ORANGI TOWN KARA

	○ B, THI, PSH, M, Anatomic M, Color mode, Free Xros M, Free Xros CM, Color Doppler, Power Doppler, PW, CW, Trapezoid B mode **B-Imaging parameter:** ○ Gain: 0 ~ 100 or better ○ Focusing: Zone Vue: dynamic pixel focusing in a zone ○ Max. depth: max 400mm, probe dependented or better ○ Dynamic range: 30-260 or better. ○ B frame rate: Max:600 f/s or better. ○ Zooming: Zoom function with picture in picture indication on both frozen and live images with Pan zoom and Spot zoom (up to 10 times) **M-Imaging parameter:** ○ Display formats: V2:3, V3:2, V3:1, H2:3, Full (V: vertical, H: horizontal) with M sweep speeds 6 steps or better. ○ Raw Data Post processing: Gain , Speed , Dynamic Range , Gray Map , Tint Map , Display format **Spectral Doppler imaging parameter:** ○ Raw Data Post processing: Gain , Speed , Dynamic Range , Gray Map , Tint Map Display Display formats: V2:3, V3:2, V3:1, H2:3, Full, Duplex/Triplex (PW only) (V: vertical, H: horizontal) ○ Gain: 0 ~ 100, PRF: PW: 0.7-18.9kHz & CW: 0.2-100.0kHz or better. ○ Scale: PW: max. 729.2cm/s; min. 21.5cm/s & CW: max. 3850.0cm/s; min. 9.6cm/s or better. **Color/Power Doppler imaging parameter:** ○ 0-100, 2/step: 0 ~ 100, Color frame rate Max 222f/s or better **TDI and Quantitative Analysis:** ○ Imaging mode: TVI, TEI, TVD, TVM with TDI frame rate Max frame rate: 3175 f/s or better. **Storage:** ○ Hard disk (1TB HDD, could be upgraded to 512GB SSD with SSD price) ○ **Cine memory:** Maximum cine memory up to 32577 frames (B storage server) or 163.84s (M storage server) (depends on the mode) or better. ○ Retrospective and prospective storage cine length: 480s • **Portability & Power:** ○ Compact, mobile design with ease of transport ○ Built-in battery with up to 60 minutes of scanning time		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

	○ Magnetic protective cover for cable management • **Diagnostic Ultrasound System· Height adjustable panel** · PC Keyboard · Main Unit · EU Power Cord · Built-in Wireless Adapter · DICOM Basic · DICOM Worklist · iScape View · iNeedle · IMT . Stress Echo · iWorks · DVR Module · ClamAV · ECG with Leads · Free Xros M · Free Xros CM · Tissue Doppler Imaging · TDI QA · LVO · Tissue Tracking QA · RIMT · Auto EF · R-VQS · V-Mapping · Built-in Battery · CW · · Linear Probe 3-13MHz **Smart Workflow & Automation** • **iScan Helper:** Guided scanning tool for learning and practice • **iWorks:** Standardized scanning protocol to avoid missing planes • **Smart Vue:** Automatic recognition and restoration of standard planes • **Auto Measurement Tools:** ○ Auto-trace and measurement for improved efficiency ○ Auto Doppler trace for velocity assessment ○ Smart HR Flow for optimized vascular imaging **Connectivity & Data Management**		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

	• **DICOM Compatibility:** DICOM 3.0 for seamless integration with hospital systems • **Storage & Export Options:** o Internal storage for patient data and imaging o USB, Ethernet, and Wi-Fi options for data transfer o PACS compatibility for hospital data management • **Software Updates:** Living Technology ensures ongoing software improvements **with Standard Probes** **- Convex Probe** 1 – 5 MHz Multi-Frequency Convex Probe for B/M/CDI/PW. **Cardiac Probe** 2 - 4 MHz Multi-Frequency Phased Array Adult Probe for Cardiac studies. **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end..		
8.	**ETT Machine** **Specification:** 1. Dedicated Computer based ETT System 2. Analysis of ST levels, ST slopes and ST –Index 3. Report: 12 leads, rhythm and full disclosure arrhythmia, exercise summary and trend 4. Holter and Stress Test Review 5. Full High-Definition Color Monitor with screen size of 19" or more 6. 12 Channels Concurrent Acquisition and Display 7. Multiple user selectable display formats, i.e., 6 Channels, 12 Channels, ST trends etc. 8. Display QRS Detection for selected ECG Leads 9. Can display Heart Rate, Speed, Workload, Exercise Time or more 0. Patient Worklist Capability and Report Customization 1. Electronic Report Formats: PDF / XML /DICOM Treadmill: 2. Fully Programmable, Medical Graded Unit with belts 13. Walking Surface: Minimum Width = 20 inches or more and Minimum Length of 60 inches or more 4. Can be controllable from main unit 5. Speed Adjustable: 0 – 18 km / h or better 6. Emergency Stop Button 7. Weight Bearing Capacity: Minimum 200 Kg or more 8. Automatic Blood Pressure Measurement Device 9. Automatic Baseline Drift Control Filter 0. Complete integrated full functional workstation 1. Dedicated imported cart supplied by from same manufacturer Other Parameters: 2. Operating Requirement: AC 220 V & 50Hz 23. Online pure sinewave UPS (Recommended by manufacturer) for at least 30mins operational backup time for complete unit	**1**	**216000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

	Accessories: Complete with all standard accessories Type and Quantity of Reusable Cuffs (Latex Free) of all available sizes for all body contours (To be specified by the Procuring Agency and End-user) Desktop PC minimum i5 or better for data analysis with permanent licensed Operating System Windows Laser Jet Printer or better System capability to connect with Picture Archiving & Communication System (PACS) / Hospital Information System (HIS) (The Procuring Agency and End-user to specify) **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end..		
9.	**O T Light Double doom with Camera** **Specification:** • Light Source: LED technology designed to minimize shadowing and optimize visibility in the operating theater. • Electric Field Adjustment: Capability to adjust the light intensity to match surgical needs. • Laminar Flow Properties: Designed to support sterile environments by ensuring proper airflow. • Main Light Intensity: Minimum 160,000 lux at 1 meter distance. • Satellite Light Intensity: Minimum 160,000 lux at 1 meter distance. • Light Diameter 140~300mm or better. • Depth of illumination 20% = 1400mm and 60% have 800mm • Unit have Patented, proven reflector technology Unit should have No color deterioration. • Unit have Optimum hygiene • Unit have Safety glass for best possible clarity. Should have Laminar Flow Technology. • Color temperature adjustment Standard 4350 and optional 3500- 5100 Kelvin or better • Electric field adjustment • Temperature increase at head height <1°C • Color rendering index to 99 Ra and R9 = 97 • Unit Have Life time of LED *Camera/Video Integration* **Video Options:** Support **integrated or carrier/arm-mounted FHD/4K cameras** in the light head or separately, depending. **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.	2	420000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SINDH GOVT QATAR HO.

	Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end..		
10.	**OT Table Electro Hydraulic with Remote Control** **Specification:** • Stainless steel structure with slow rebound latex foam mattress • Five section body layouts for multiple surgical position request • Reliable 6-movement and silent lifting • Table top sliding for X-ray and C-Arm • Individual kidney bridge elevation system • Built-in back up battery • 495mm lowest table table position for endoscopic surgery • Bluetooth wireless remote control **TECHNICAL FEAUTRES** • Designed specially to meet the different demands of general surgical operations, latex foam mattress, • Concealed rubber lift locker, single remote controller • Full Length: 2050mm or above • Width: 520mm or above • Height adjustment: 720-1070mm or above • Trendelenburg: 36° • Reversed Trendelenburg: 26° • Tilt left/right: +21/21° • Longitudinal shift sliding: 320mm (Optional) • Loading capacity: 400 kg • Leg plate up/down: +36°/-90° • Head rest up/down: +60°/-90° • Back plate up/down: +80°/-50° • Reflex position: 110° • Flex position: 225° • Split leg place: 90° (each) • Kidney bridge elevation: 125mm • Backup battery: included, for 80 movements • Power supply: 100-240V/ 50-60Hz **Standard accessories:** • mattress pad (1 piece), • cable controller (1 pc),	2	360000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

	• arm support (1 pair), • shoulder support (1 pair), • body support (1 pair), • leg support (1 pair), • anesthesia frame(1 set) **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.		
11.	**Pheco Machine** **Specification:** • A smart synthesis of advanced technologies, functionality and reliability. • Compact platform for cataract and retinal surgery • The device boasts an easy-to-use touch screen and a user-friendly GUI, • Super simple set-up, for everyday cost-effectiveness. • Fluidics is based on two high efficiency pumps, • Equipped with high flow rate and vacuum values • Managed using easy assembly disposable and reusable I/A cassettes. • The flow-based peristaltic pump by a latest generation stepper motor • The integrated programmable IV pole gravity fed and the air controlled infusion. • Simultaneous linear control of two different system functions with smooth up-and-down and simple left to right movements. • Measuring the actual phaco tip stroke during surgery • Technology minimizes the energy delivered to the eye for a safer and more efficient surgery and faster patient rehabilitation. • Ensure maximum chamber stability and patient safety. • Dynamic IOP Control can maintain constant intraocular pressure. • Fluid delivery gravity fed (determined by the height of irrigation source) • Aspiration pump type peristaltic • Actuating medium (Venturi pump) pressurized air coming from an external source • Available vacuum range: 5 - 650 mmHg or better. • Available flow rate range: 1 - 65 cc/min (peristaltic only) or better. • Anterior Hand-piece type: pneumatically powered guillotine cutter • Available cutting rate: up to 2000 cuts/min or better. **Complete with the following** • Pheco Hand-piece six crystal 01 • 20g Pheco tip 01 • Silicone Sleeve for 20g tips 01 box • Pheco Tips Wrench 01 • Drainage Bag 02 • Test Chamber 02 • Anterior Vitrectomy Cutter 01 • Diathermy Bipolar Cable 01 • Diathermy Forceps 01 • Easysys Reusable I/A Cassette 01 • Wrench for easysys reusable I/A cassette 01 **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.	**1**	**360000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

	Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end.		
12.	**Binocular Microscope** **Specification:** Standard microscope with infinity color corrected optics Binocular head, adjustable inter papillary distance Tube head with an inclination of 30 Degree or more Rotating quadruple/quintuple nose piece 2-layer with mechanical sliding stage, XY moving range Condenser carrier, vertically/horizontal adjustable Bright field condenser, N.A. of 1.25 Aperture iris diaphragm LED illumination 220V 50 Hz, AC User Adjustable Settings: Coaxial coarse and fine focusing system, with focusing stop mechanism Variable intensity control system Eyepiece lenses: - 10x wide field, focusable, Objective lenses (Achromat): - 4xScanning lens 10X 20X 40x 100x oil immersion **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.	**2**	**23400**
13.	**Slide Dryer Hot Plate** **Specification:** • **Digital Display:** Real-time temperature monitoring and LCD display. • **Memory Function:** Retains the last set temperature upon restarting. • **Safety Features:** Over-temperature alarm protection, often with 1-5°C thresholds. • **Construction:** Durable, corrosion-resistant, and antimicrobial powder-coated steel. • **Temperature Range:** Typically ambient to 95°C or 99°C. • **Temperature Precision:** High precision 1°C or 0.5°C. • **Capacity:** Ranges from roughly 28 (slimline) to 60+ slides. • **Surface Design:** Often flat or with a 65° louvered incline for optimal drainage. • **Accessories:** Optional stainless steel lids for faster heating. • **Power:** Usually 110V-240V AC 50/60Hz **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.	**1**	**12000**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

14.	**Binocular Microscope with Teaching Aid.** **Specification:** microscope with infinity color corrected optics Binocular head, Siedentopf binocular, 360° rotating, 30° inclined. adjustable inter papillary distance Tube head with an inclination of 30 Degree or more Rotating quadruple/quintuple nose piece 2-layer with mechanical sliding stage, XY moving range Condenser carrier, vertically/horizontal adjustable Bright field condenser, N.A. of 1.25 Aperture iris diaphragm LED illumination 220V 50 Hz, AC User Adjustable Settings: Coaxial coarse and fine focusing system, with focusing stop mechanism Variable intensity control system Eyepiece lenses: - 10x wide field, focusable, Objective lenses (Achromat): - 4xScanning lens 10X 20X 40x 100x oil immersion **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years startingfrom date of installation of equipment to the consignee end.	1	30000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

SECTION B

Supply & Installation of Devices, Equipment and Miscellaneous Items

Technical Evaluation Criteria for Section "B" Maximum Allocable Marks Score 70% to Qualify

Document Check List (Mandatory Documents)

If any bidder failed to submit any of the following documents, the same shall be disqualified.

Sr. No.	DOCUMENT DESCRIPTION	YES / NO	PAGE #
1	Tender Purchase Receipt (Original)/Pay order		
2	Bid Security Copy with value hidden in Technical Proposal; Originalin Financial)		
3	Technical Proposal on Bidder's Letterhead		
4	Compliance sheet on letter head		
5	Undertaking (as per sample form)		
6	Income Tax & GST Registration Certificate		
7	Valid Professional Tax Certificate		
8	FBR Income Tax Returns (last 3 years)		
9	Audited Financial Statement (last 3year)		
10	Factory / Workshop Details (duly equipped)		
11	Valid Authorization Letter from Manufacturer in reference to this particular tender.		
12	Catalogue / Brochures / Drawing		
13	ISO 9001 of Bidders / Manufacturer (where applicable) **(website link must be provided for verification)**		

Note:

- All pages of the bid, except for un-amended printed literature, shall be initialed by the Bidder.
- Documents must be submitted in binding, with filled Table of Content, Page Numbering and tagging wherever required.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar._____

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI._____

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop._____

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar._____

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi._____

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT SINDH GOVT. QATAR HOSPITAL ORANGI TOWN KARACHI

EVALUATION / QUALIFICATION

1.1 The procuring agency reserves the right to evaluate and compare the bids on itemized basis OR on the basis of a group of similar nature goods OR goods compatible with each other.

1.2 The following merit point system for weighing evaluation factors/criteria will be applied for technical proposals.

1.3 Bidders achieving minimum total **70% marks** will be considered to qualify. Documentary evidence must be attached in support of each parameter.

1.4 Any Bid not meeting the mandatory requirements of evaluation criteria will be disqualified / rejected straight away and will not be considered for further evaluation.

A. PRODUCT EVALUATION

S#	PARAMETERS / SUB-PARAMETERS	Total Marks
1	**Conformity to the Purchaser's Specifications (Mandatory)**	**40**
1.1	Fully compliant with the required specifications	40
1.2	Compliant with minor deviation (up to 10% subject to main function is not affected)	30
1.3	Non-compliant to required specifications	Disqualify
2	**Product Certification**	**15**
2.1	USA Food & Drug Administration (USA-FDA) and European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labor and Welfare Govt. of Japan	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labor and Welfare Govt. of Japan **OR** USA - Food and Drug Administration (USA-FDA) and European Community (CE)	12
2.3	USA - Food and Drug Administration (USA-FDA)	10
2.4	European Community (CE) MDD	08
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labor and Welfare Govt. of Japan	08
	TOTAL MARKS PRODUCT EVALUATION (A)	**55**

B. BIDDER EVALUATION

S#	PARAMETERS / SUB-PARAMETERS	Total Marks
1	**Manufacturer's Authorization**	**10**
2	**Delivery Schedule at consignee's end**	**10**
2.1	Up to 40 days DDP	10
2.2	More than 40 days	00
3	**Past Experience / Performance of Last 3 years**	**10**
3.1	Last three Years' Experience of five Public Sector Tertiary Care Hospitals (1 Marks of each)	05
3.2	Last three Years' Experience of five Private Sector Tertiary Care Hospitals (1 Marks of each)	05

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

S#	PARAMETERS / SUB-PARAMETERS	Total Marks
4	**Average Annual Turnover during last three (03) financial years (Audited Statements of Accounts and Income Tax Return Forms must be attached as supporting documents)**	**10**
4.1	Turn over below PKR 100 million	00
4.2	Turn over above PKR 100 million	04
4.3	Turn over above PKR 200 million	06
4.4	Turn over above PKR 300 million	08
4.5	Turn over above PKR 500 million	10
5	**Replacement Warranty**	**05**
5.1	Replacement Warranty for 1 year	05
	TOTAL MARKS BIDDER EVALUATION (B)	**45**
	GRAND TOTAL (A + B)	**100**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar._____

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI._____

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop._____

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar._____

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi._____

(DR. [illegible] SIRAJ)
MEDICAL SUPERINTEND
GOVT QATAR HO
[illegible]

BIDDER EVALUATION CRITERIA

Bid found to be responsive with maximum accumulative points (Technical + Financial) shall be accepted as the **mostadvantageous bid** as per SPPR, 2010 (Amended 2021).

Bids will be evaluated on the basis of following evaluation criteria.

A. TECHNICAL EVALUATION

i) Technical evaluation has 70% weightage

ii) Bids will be evaluated on a Bidder's capacity and capability to undertake supplies and on quoted item quality, performance and conformity to the required specifications.

Technical capacity of a bidder includes availability of workshop, tools, technical staff, number ofinstallation and experience in the province of Sindh for prompt services to Karachi Division. Past performance of the bidder at **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.**

iii) will also be reviewed.

iv) A bidder must achieve a minimum of 70 points out of total 100 points in technical evaluation, to be considered for financial evaluation.

B. FINANCIALEVALUATION

i) Financial evaluation has 30% weightage.

ii) Only financial proposal of bidders with responsive technical proposal will be opened. Financial proposal of bidders who do not meet required mandatory documents and minimum 70 points out oftotal 100 points in technical evaluation will be returned unopened.

iii) Bidder who has offered the lowest value for the item willget30points.

iv) All remaining bidders after the lowest shall get points on their financial proposal, calculated as per the following formula.

$$\frac{\text{Value of lowest bidder}}{\text{Value of higher bidder [2nd, 3rd, 4th, etc.]}} \times 30$$

C. FINAL BID EVALUATION

i) Points obtained by bidders in technical and financial evaluation shall be prorated and finalized as follows.

Evaluation	Weightage
Technical	70%
Financial	30%
Total	**100%**

ii) Bid of a bidder who is found technically responsive with highest accumulated points will be considered as the most advantageous bid.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

SECTION B

Supply & Installation of Devices, Equipment and Miscellaneous Items

S. No	Name of items	QTY	Fixed Bid Security
1.	**Split Air Conditioner 1.5 Ton** **Specification:** BTU Capacity: 18000 BTU Compressor W: 5300 Low Voltage Startup: 150V Eco Friendly Refrigerant: R410A Energy Efficient Class A+: 60% Electricity s Auto Voltage Adaptation: 150V-260V **COOLING** BTU/H Capacity (H/S/L*): 19100/17750/5118 Output Capacity (W) (H/S/L*): 5600/5200/1500 Power Input (W)(H/S/L*): 2100/1529/390 Energy Efficiency Ratio (EER): 3.6 **HEATING** BTU/H Capacity (H/S/L*): 22550/19600/4260 Output Capacity (W) (H/S/L*): 6600/5750/1450 Power Input (W)(H/S/L*): 2200/1580/390 Energy Efficiency Ratio (EER): 3.6 With compatible imported/export quality stabilizer. Complete installation and commissioning are the responsibility of the vendor Warranty three years including all parts Compressor warranty ten years Import/Export Quality ISO/CE certified Warranty 03 years from the date of installation	30	270000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT GOVT. QATAR HOSPITAL SINDH RANGI TOWN KARACHI

HUMAN RESOURCE INCLUDING DETAIL OF TECHNICAL TEAM

(FORM A)

S. #	Name	Designation	Posted Location	Date of Joining	Education	Trainings

Note: Supporting Documents must be provided for each employee.

CEO/Proprietor/Managing Director

Name:_____

Email Address:_____

Cell Phone:_____

Director/General Manager Services

Name:_____

Email Address:_____

Cell Phone:_____

Director/General Manager Sales

Name:_____

Email Address:_____

Cell Phone:_____

Technical Focal Person for this Project

Name:_____

Email Address:_____

Cell Phone:_____

Address:_____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

SPECIALIZED TESTING AND CALIBRATION TOOLS(FORM B)

Supporting Documents must be provided for each specialized tool

S. #	Tool Description	Make	Model	Date of Purchase	Date of Last Calibration

MAJOR SPARE PARTS(FORM C)

- Purchase documents must be provided,
- It may include Sensor, PCB, LCD display, Pumps, motors, detectors, Laser and other items recommended by manufacturer etc., list of recommended spare parts must also be provided.
- Parts such as cables, connectors will not be considered major spare parts.
- Bidder using an inventory management software may provide a software generated list instead of the below format.

S. #	Part Number	Part Description	Make	Model	Date of Purchase	Qty

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

(DR. PRASHAD SIRA...)
MEDICAL SUPERINTENDENT
SINDH GOVT. QAT
ORANGI TO...

COMPLIANCE SHEET/ TECHNICAL EVALUATION SHEET

(FORM D)

Name the Vendor:

Make/Manufacturer:

Model: Year of introduction:

Country of Origin:

Country of Manufacturing: Country of

Major parts sourcing:

Use full life as per manufacturer (in Years): CE

(MDD) link:

FDA(US) Link:

JIS/MHLW Link: ISO

13485 Link:

Pre-installation requirement.

Electrical Connection:

Grounding:

Backup Supply:

Water (RO/DI with flow):

Drain:

Any other Requirement:

S. #	Tender Specification	Offered Specification	Page # of the supporting Document

Note: Supporting documents must be provided

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT GOVT. QATAR HOSPITAL NGI TOWN KARACHI

FINANCIAL EVALUATION SHEET(FORM E)

	Annual Turnover for the last 3 years Year (PKR)
Year 2024	
Year 2023	
Year 2022	

Financial Information in PKR	Year 2024	Year 2023	Year 2022
Total Assets (TA)			
Total Liabilities (TL)			
Current Assets (CA)			
Current Liabilities (CL)			
Total / Gross Revenue (TR)			
Profits Before Taxes (PBT)			
Net Profit			
Current Ratio			

Supporting Document		Yes / No
Bank Statement	Year 2024	
	Year 2023	
	Year 2022	
Audited Accounts	Year 2024	
	Year 2023	
	Year 2022	
Bank Account Maintenance Certificate		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

MANUFACTURER'S AUTHORIZATION

(FORM F)

[SEE CLAUSE 11.1 (A) OF THE INSTRUCTION TO BIDDERS]

To:

Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.

WHEREAS_____**[name of the Manufacturer]**_____who are established and reputable Manufacturers of _____**[name and /or description of the goods]**_____having factories at____ _____**[address of factory]**_____do hereby authorize _**[name and address of Bidder / Agent]** _____to submit a bid, and subsequently follow-up / negotiate and sign the Contract with you against Invitation for Bids (IFB) / Tender Notice for the goods manufactured, by us, under the patent name of

_____for performance of the contract.

We hereby commit and assure our full guarantee and warranty / guarantee as per Clause 12 of the GeneralConditions of Contract for the goods offered for supply by the above mentioned firm against this Invitation for Bids.

[Signature for and on behalf of Manufacturer]Note:

This letter of authority should be on the letterhead of the Manufacturer and should be signed by a person competent and having the power of attorney to bind the Manufacturer. It should be included by the Bidderin its bid.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar._____
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI._____
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop._____
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar._____
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi._____

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT GOVT. QATAR HOSPITAL ORANGI TOWN KARACHI

LIST OF INSTALLATION / REFERENCE CLIENT(FORM G)

Equipment:

Make/Manufacturer:

Model:

S. #	Department	Name of Institute	City	Date of Installation	Date of Satisfactory Certificate

Note: Attach the following supporting documents.

- Performance certificate (issued in last one year) on letter head of bidder/Hospital
- Installation Report signed by Head of department and Biomedical Engineer
- Supply order / purchase order

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
GOVT. QATAR HOSPITAL
KORANGI TOWN KARACHI

PERFORMANCE CERTIFICATE(FORM H)

[on bidder's letter head]

This is to certify that the institute is satisfied by the performance of the firm and equipment as detailed below.

Purchase Order No: ,dated
Make & Model no.:
Serial Number:
Quantity:
Name of the consignee:

The Firm has fulfilled his contractual obligation regarding the following services:
Satisfactory Installation, performance, and commissioning of equipment.
Furnishing detailed operation and maintenance manual for each equipment.
Training of the operators/users in operating the equipment to the satisfaction of the Institute.
Maintain up time of %
Perform Preventative maintenance at_Interval during warranty/contract period
Provide Spare parts timely

Biomedical Engineering Department.

Name _____

Designation _____

Date _____

Signature with Stamp _____

Head of the Institute/Medical Director/Medical Superintendent

Name _____

Designation _____

Date _____

Signature with Stamp _____

Representative of the firm

Name _____

Designation _____

Date _____

Signature with Stamp _____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDE
GOVT. QATAR HO
NGI TOWN KA

BID LETTER FORM

From:
(Registered name and address of the bidder)To:

Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.

Dear Sir / Madam,

Having examined the bidding document and amendment thereon we the undersigned, offer to provide services to the works including in conformity with the terms and conditions of the bidding document and amendments there on, for the following project in response to your tender call dated_____

Tender Title: .....

We undertake to provide services/execute the above project or it part assigned to us in conformity with the said bidding documents for an estimated sum of Rs._____(Rupees - _____) (total bid amount in words and figures) which may vary in accordance with the schedule of prices attached herewith and coverage options made by Drug Testing Laboratory Sindh or its user organization.

If our bid is accepted, we undertake to;

1) Provide services/execute the work according to the time schedule specified in the bid document,
2) Obtain the performance guarantee of bank in accordance with bid requirements for the due performance of the contract, and
3) Agree to abide by the bid conditions, including pre-bid meeting minutes if any, which remain binding upon us during the entire bid validity period and bid may be accepted any time before the expiration of that period.
4) We understand that you are not bound to accept the lowest or any bid you may receive, nor to give any reason for the rejection of any bid and that you will not defray any expenses incurred by us in bidding.

Place:
Date:

Bidder's signature
and seal.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

CONTRACT FORM

THIS AGREEMENT made the ..... Day of..... (year) Between the Procuring Agency (hereinafter "the_____") of one part and .....(Name of Vendor) of .....(City and country of Vendor) (Hereinafter "the Supplier") of the other part:
WHEREAS the_____is desirous that certain Supplies, as described in the bid documentand briefly outlined below, should be provided by the Vendor.

Date of tender call:
Title of the project:
Brief outline of the work:

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS;

In this agreement words and expression shall have the same meanings as are respectively assigned to themin the bid document referred to.
The following document shall be deemed to form and be read and construed as part of this Contract, viz..

1) Bid document(s)
2) Pre-bid conference minutes if any,
3) Clarification on bid document issued if any,
4) Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.of award notification.

In case of conflict among documents mentioned above, the documents mentioned above in reverse order will prevail over other documents. In consideration of the payments to be made by the **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.** to the Supplier as hereinafter mentioned, the Supplier hereby covenants with the **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.** (Tender Title. . . . . .
. . . . . .) and to remedy defects therein conformity, in all respects, with the provisions of the contract.

The **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.** hereby covenants to pay the Supplier in consideration of the provision of the goods and services and the remedying of defects therein, the contract price or such other sum as may become payable under the provisions of the contract at the timesand in the manner prescribed by the contract.

Brief particulars of the services which shall be supplied/provided by the Supplier are as under:

Solutions, service or material	Quantity	Unit price	Amount	Remarks

IN WITNESS whereof the parties hereto have caused this Agreement executed the day and year abovewritten.
Signed, sealed, delivered by_____the (for the Procuring agency)Signed, sealed, delivered by_____the (for the Supplier)

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi**_____

DR. RASHID SIRAJ
MEDICAL SUPERINTENDENT
GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

FORM OF PERFORMANCE SECURITY
(Bank Guarantee)

Guarantee No.: _____
Executed on: _____
Expiry date: _____

[Letter by the Guarantor to the Employer]

Name of Guarantor (Bank) with complete address (Scheduled Bank in Pakistan):

_____

Name of Principal (Contractor, Manufacturer, Supplier or any bidder) with complete address:

_____

Penal Sum of Security (express in words and figures):

_____

_____

Letter of Acceptance No._____Dated: _____

KNOW ALL MEN BY THESE PRESENTS, that in pursuance of the terms of the Bidding Documents and above said Letter of Acceptance (hereinafter called the Documents) and at the are e of the said Principal we, the Guarantor above named, are held and firmly bound unto the **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.** (hereinafter called the Employer) in the penal sum of the amount stated above for the payment of which sum well and truly to be made to the said Employer, we bind ourselves, our heirs, executors, administrators and successors, jointly and severally, firmly by these presents.

THE CONDITION OF THIS OBLIGATION IS SUCH that whereas the Principal has accepted the Employer's
above said Letter of Acceptance for _____ (Name of Contract) for the _____(Name of Project).

NOW THEREFORE, if the Principal (Contractor) shall well and truly perform and fulfill all the undertakings, covenants, terms and conditions of the said Documents during the original terms of the said Documents and any extensions thereof that may be granted by the Employer, with or without notice to the Guarantor, which notice is, hereby, waived and shall also well and truly perform and fulfill all the undertakings, covenants terms and conditions of the Contract and of any and all modifications of said Documents that may hereafter be made, notice of which modifications to the Guarantor being hereby waived, then, this obligation to be void; otherwise to remain in full force and virtue till all requirements of Condition of Contract are fulfilled. Our total liability under this Guarantee is limited to the sum stated above and it is a condition of any liability attaching to us under this Guarantee that the claim for payment in writing shall be received by us

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

within the validity period of this Guarantee, failing which we shall be discharged of our liability, if any, under this Guarantee.

We,_____(the Guarantor), waiving all objections and defenses under the Contract, do hereby irrevocably and independently guarantee to pay to the Employer without delay upon the Employer's first written demand without cavil or arguments and without requiring the Employer to prove or to show grounds or reasons for such demand any sum or sums up to the amount stated above, against the Employer's written declarationthat the Principal has refused or failed to perform the obligations under the Contract which payment will be effected by the Guarantor to Employer's designated Bank & Account Number.

PROVIDED ALSO THAT the Employer shall be the sole and final judge for deciding whether the Principal (Contractor) has duly performed his obligations under the Contractor has defaulted in fulfilling said obligations and the Guarantor shall pay without objection any sum or sums up to the amount stated above upon first written demand from the Employer forthwith and without any reference to the Principal or any other person.

IN WITNESS WHEREOF, the above-bounden Guarantor has executed this Instrument under its seal on the date indicated above, the name and corporate seal of the Guarantor being hereto affixed and these presents duly signed by its undersigned representative, pursuant to authority of its governing body.

Witness:

_____
Guarantor (Bank)

1._____
(Name, Title, Signature & Seal)

Signature:_____
2. _____ Name:_____

(Name, Title, Signature & Seal) Title: _____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSP
ORANGI TOWN KARA

INTEGRITY PACT

DECLARATION OF FEES, COMMISSION AND BROKERAGE ETC PAYABLE BY THE SUPPLIERS/CONTRACTORS/CONSULTANTS

Contract Number: **NO.**_____ **Dated:** _____
Contract Value: **Rs.** _____
Contract Title: _____

M/s._____hereby declares that it has not obtained or induced the procurement of any contract, right, interest, privilege or other obligation or benefit from Government of Sindh (GoS) or any administrative subdivision or agency thereof or any other entity owned or controlled by it (GoS) through any corrupt business practice.

Without limiting the generality of the foregoing, **M/s.**_____represents and warrants that ithas fully declared the brokerage, commission, fees etc. paid or payable to anyone and not given or agreed to give and shall not give or agree to give to anyone within or outside Pakistan either directly or indirectly through any natural or juridical person, including its affiliate, agent, associate, broker, consultant, director, promoter, shareholder, sponsor or subsidiary, any commission, gratification, bribe, finder's fee or kickback,whether described as consultation fee or otherwise, with the object of obtaining or inducing the procurement of a contract, right, interest, privilege or other obligation or benefit, in whatsoever form, **Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.,** except that which has been expressly declared pursuant hereto.

M/s._____certifies that it has made and will make full disclosure of all agreements and arrangements with all persons in respect of or related to the transaction with PA and has not taken any action or will not take any action to circumvent the above declaration, representation or warranty.

M/s._____accepts full responsibility and strict liability for making any false declaration, not making full disclosure, misrepresenting facts or taking any action likely to defeat the purpose of this declaration, representation and warranty. It agrees that any contract, right, interest , privilege or other obligation or benefit obtained or procured as aforesaid shall, without prejudice to any other right and remedies available to PA under any law, contract or other instrument, be voidable at the option of PA. Notwithstanding any rights and remedies exercised by PA in this regard, **M/s.** _____agrees to indemnify PA for any loss or damage incurred by it on account of its corrupt business practices and further pay compensation to PA in an amount equivalent to ten time the sum of any commission, gratification, bribe, finder's fee or kickback given by **M/s.** _____as aforesaid for the purpose of obtaining or inducing the procurement of any contract, right, interest, privilege or other obligation or benefit, in whatsoever form, from PA.

_____ M/s. _____

_____
Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

(DR. RASHID SIRAJ) MEDICAL SUPERINTENDENT (SINDH GOVT) QATAR HOSPITAL ORANGI TOWN KARAGHI

Bid Form & Price Schedule

BID FORM

To:

Medical Superintendent Sindh Government Qatar Hospital Orangi Town Karachi. Dear Sir,

Having examined the Bidding Documents, the receipt of which is hereby duly acknowledged, we, the undersigned, offer to supply and deliver the goods specified in the said Bidding Documents for the sum of

[Total Bid Amount Rs. _____], [Bid Amount in words _____ only]or such other sums as may be ascertained in accordance with the Schedule of Prices attachedherewith and made part of this bid.

1. The free of cost / donation / discounts offered and the methodology for their application are:

_____

_____

2. We undertake, if our bid is accepted, to deliver the goods in accordance with the delivery schedulespecified in the Schedule of Requirements.
3. If our bid is accepted, we shall obtain an unconditional guarantee of a bank in the sum of 5% of theContract Price for the due performance of the Contract, in the form prescribed by the Procuring Agency.
4. We agree to the validity of this bid for 90 days from the date fixed for financial bid opening and itshall remain binding upon us and may be accepted at any time before the expiration of that period.
5. Until a formal Contract is prepared and executed, this bid, together with the written acceptancethereof and notification of award, by the Procuring Agency, shall constitute a binding Contract between us.
6. We understand that you are not bound to accept the lowest or any bid you may receive.

Name _____

In the capacity of _____

Signed _____

Duly authorized to sign the Bid for and on behalf of _____

Date _____

(A) PRICE SCHEDULE IN PAK RUPEES delivered duty paid (DDP BASIS)

FOR GOODS OFFERED WITHIN THE PROCURING AGENCY'S COUNTRY

S#	Detailed Specification of Goods	Model/ Cat No.	Nameof Manufacturer	Country of Origin	Portof Shipment	Quantity of Stores	Unit	Rate Per Unit	Total Price
1	2	2	4	5	6	7	8	9	10
Total Amount in Pak Rs.									

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

Name_____

In the capacity of _____

Signed_____

Duly authorized to sign the Bid for and on behalf of _____

Date_____

Port of Shipment and Country of origin of "MAJOR PART(S) OF THE EQUIPMENT" must be clearly reflected separately in the Technical and Financial bids. The "Origin" means the place where the "goods" are mined, grown, or

NOTE:

(B) PRICE SCHEDULE IN FOREIGN CURRENCY
(CFR / CNF/ C&F / CPT BASIS)
FOR GOODS OFFERED FROM OUTSIDE THE PROCURING AGENCY'S COUNTRY

S #	Detailed Specification of Goods	Model / Cat No	Name of Manufacturer	Country of Origin	Port of Shipment	Quantity of Stores	Unit	Curr-Ency	Rate Per Unit	Total Price
1	2	2	4	5	6	7	8	9	10	11
Total Amount in Foreign Currency										

Name _____

In the capacity of _____

Signed _____

Duly authorized to sign the Bid for and on behalf of _____

Date _____

NOTE:

Port of Shipment and Country of origin of "MAJOR PART(S) OF THE EQUIPMENT" must be clearly reflected separately in the Technical and Financial bids. The "Origin" means the place where the "goods" are mined, grown, or

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar._____
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI._____
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop._____
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar._____
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi._____

(DR. RASHID SIRAJ)
MEDICAL SUPERINTENDENT
SINDH GOVT. QATAR HOSPITAL
ORANGI TOWN KARACHI

Schedule of Requirements

SCHEDULE OF REQUIREMENTS

The entire quantity of the ordered supplies shall be delivered in 16 weeks or earlier from the dateof issuance of supply order / contract award without any penalty.
Delay in the delivery shall result in penalties to be paid by the Supplier without any argument or question according to the prevailing SPPRA Rules.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-medical Technician, Elector-medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

Checked & approved by Expert Committee for Machinery Equipment:

i.	Dr. Shahid Pervez, Histopathologist, Aga Khan University Hospital Kyc.	
ii.	Dr. Binish Arif Sultan, Associate Professor & Head of Pathology Lab, JSMU, Karachi.	
iii.	Prof. Dr. Tariq Mehmood, Professor Emeritus, Radiation Oncology, Member of PAF, JPMC, Karachi.	
iv.	Dr. Syed Farjad Sultan, Associate Professor, Anaesthesia Department, DUHS / Civil Hospital Karachi.	
v.	Prof. Dr. Sajida Qureshi, Professor of Surgery, DUHS / Dr. Ruth K.M. Pfau Civil Hospital Karachi.	
vi.	Prof. Dr. Syed Mahmood Hassan, Professor of Pathology, JSMU, Karachi.	
vii.	Dr. Shoaib Malik, Associate Professor, Anesthesia Department, JPMC Karachi.	

CONCURRED BY:

Professor Dr. Gulshan Ali Memon
Vice Chancellor, Peoples University of Medical & Health Sciences for Women, Nawabshah-Shaheed Benazirabad/
Member

Professor Dr. Ikram Din Ujjan
Vice Chancellor, Liaquat University Medical & Health Sciences, Jamshoro./
Member

Professor Dr. Nusrat Shah
Vice Chancellor, Shaheed Mohtarma
Benazir Bhutto Medical University
Larkana / **Member**

Dr. Ahson Qavi Siddiqi
CEO, Health Care Commission
Sindh/ Member

Prof. Dr. Sajida Qureshi
Professor of Surgery,
DUHS / Dr. Ruth K.M. Pfau
CHK./ Member

Special Secretary,
Chief Minister's Inspection, Enquiries &
Implementation Team Department,
Govt. of Sindh/ Member

Mr. Mohammed Ibrahim Memon
Additional Secretary (PM&I), Health
Department, Govt. of Sindh,/
Member / Secretary

Additional Secretary (GA),
SGA&CD, Govt. of Sindh/
Member

Professor Dr. Amjad Siraj Memon
Vice Chancellor,
Jinnah Sindh Medical University/
Chairman