



TENDER DOCUMENTS

PURCHAS OF

MEDICAL LABORATORY EQUIPMENT

FOR THE YEAR 2026-27



PEOPLES MEDICAL COLLEGE HOSPITAL

NAWABSHAH SHAHEED BENAZIRABAD


Date of issuance of Tender from the date of publication up to
Date of Submission of Tender
Date of Opening of Tender

21-09-2026 up to 03.00 P.M
22-09-2026 up to 01.00 P.M
22-09-2026 @ 02.00 P.M



TERMS & CONDITION OF CONTRACT GENERAL.

1. I/We _____ Muslim Adult, by
cast _____ R/O _____ bind myself /
ourselves to supply the said item to the Peoples Medical College Hospital Nawabshah (Shaheed Benazir Abad) during
the **financial year 2026-27** to until further orders. The articles are shown in tender form and the rates are specified
against each. The articles shall be of the best quality. The decision of the Medical Superintendent / officer In-charge
authorized by him to act on his behalf (therein called the said officer) in regard to the quality and kind of the articles
shall be final and binding upon me / us.
2. The supplies shall be delivered whenever required in the presence of the said officer or his representative
and myself/ ourselves or my /our authorized agent my / our agent / agents shall be responsible
person/persons and his / their name shall be known to the said officer in case it is discovered otherwise this
contract /tender will be cancelled and security money forfeited.
3. Should any delay occur on my / our part or should I / We / Our agent fail to supply the articles at he given time and
place the said officer may purchase them from the Market out of my / our security deposit and the differences
between the contract and Market rate shall be paid well by me / us.
1. The security money deposited by me /us shall be returned to me / us after successful completion of the contract and on
my/our furnishing the usual No demand certificate.
2. Tender from shall be accompanied by Earnest money as per tender form in shape of Call Deposit / Pay Order.
3. I will supply all the items as offered by me in Tender during the financial year at same rate and cost without
delay/fail.
4. The articles / stores will be delivered upon proper receipt signed by the authorized officer on the delivery challan.
The original challan will be attached with the bills.
5. The Medical Superintendent, Peoples Medical College Hospital Nawabshah (Shaheed Benazir Abad) reserves the
Right to cancel the part or whole Tenders without any reasons.
6. I / We perfectly understand all the above conditions and general directions to the contractor. I /we bind myself/
ourselves to abide by them and I /we also understand my/our contract, is liable to terminate in case of breach of any of
the terms of contract. In that case my/ our security deposit will be forfeited by the Medical Superintendent Peoples
Medical College Hospital Nawabshah (Shaheed Benazir Abad).
7. The Medical superintendent Peoples Medical College Hospital Nawabshah (Shaheed Benazir Abad) reserves the right
to increase or decrease the quantity of any item of schedules as and when it is deemed necessary without assigning
any reason.
8. Delivery orders will be supplied within the stipulated date and time as given in the order, failing which security
money will be forfeited?
9. In case of late delivery penalty of 2% of the total cost of the ordered items will be imposed after the expiry of the
given period.
10. Medical Superintendent Peoples Medical College Hospital Nawabshah (Shaheed Benazir Abad) reserve the right to
impose the following penalties for any breach of the contract by bidders.
- (a) Forfeiture of the Security money
 - (b) Forfeiture of payment
 - (c) Black listing of the firm
11. I / We shall abide by the General Sales Tax rules as applicable.
12. The approved bidder(s) has/have to deliver the material on F.O.R basis at Peoples Medical College
Hospital Nawabshah (Shaheed Benazir Abad) (Hospital Premises).
13. Income Tax, GST and SST will be deducted as per rules, on the purchase / Services.
14. The Bid be valid for nineteen (90) days from the date of opening of the tender.
- 15. Tender valid up to 30th June 2025.**



OTHER DIRECTIONS (EVALATION CRITERIA)
TECHNICAL PROPOSAL should have the following documents.

1. Original Tender Purchase Receipt.
2. Copy of the Bid offer with quoted items with terms & conditions (without rates) with signature & Stamp.
3. Photostat copy of Pay Order / Bank Draft of Earnest Money showing without rates.
4. Photo Copy of Authority letter should be provided by the bidder of coated items companies.
5. Photo Copy of Valid National Tax No Certificate upgrade
6. Photo Copy of Valid General Sale Tax Certificate updated.
7. Photo Copy of Valid Professional Tax Certificate (From Excise and Taxation Department).
8. An affidavit of Rs.100/- (Non-Judicial Stamp paper) that the firm is not black listed in any Government Department.
9. Sample of quoted item should be provided at the time of opening of tender otherwise tender will be rejected.
10. Three years' Experience and past performance in relevant field.
11. Last Three-Year Income Tax Returns Challan.
12. Three years Annual Turnover statement with Bank Certificate as per required in tender documents (must be year-wise). Rs.01 Million.

FINANCIAL PROPOSAL should have the following documents.

1. Original Pay Order / Bank Draft of Earnest Money.
2. Original Copy of Bid offer with quoted price.
3. Over writing not allowed
4. Amount written must be Figure and words
5. I/We solemnly declare that the information furnished by me/us is correct to the best of my/ our knowledge and if found incorrect our contract will be liable to be terminated.

Signature of the Contractor

Name:

Name of Firm

Full Address


C. N.I.C No

Income tax No.

G.S.T No

Cell #

Land Line #



TECHNICAL PROPOSAL

Tel: No. (0244) – 9370264.



Fax No. (0244) – 9370266

PEOPLES MEDICAL COLLEGE HOSPITAL NAWABSHAH SHAHEED BENAZIR ABAD

BIDDING DOCUMENT PURCHASE OF MEDICAL LABORATORY EQUIPMENT FOR THE YEAR 2026-27

Date of issuance of Tender from the date of publication up to

22-09-2026 up to 03.00 P.M


Date of Submission of Tender

22-09-2026 up to 01.00 P.M


Date of Opening of Tender

22-09-2026 @ 02.00 P.M

S.#	Category	Name of the Articles	Req Qty	Packing	Manufacturer
1.	Misc:	Acid Solution for Selectra 1000 ml	10 bottles		
2.	Misc:	Blood-Lancet	300 Nos		
3.	Misc:	Blue Tips	600 pkts		
4.	Misc:	Yalow Tips	200 pkts		
5.	Misc:	Calibrator Chemistry for Selectra	20 Nos		
6.	Misc:	Capillary Tube No Heparin	50 Box		
7.	Misc:	Cover Slip 18 x 18	10000 Nos		
8.	Misc:	Cover Slip 22 x 22	10000 Nos		
9.	Misc:	Dengue kit	500 Nos		
10.	Misc:	Distil water 30 lit can	100 Nos		
11.	Misc:	E.S.R PIPPET (Disposable)	4000 Nos		
12.	Misc:	E.S.R stand	100 Nos		
13.	Misc:	EDTA Vacuum Tube	5000 Nos		
14.	Misc:	Field stain A+B	20 Nos		
15.	Misc:	Gel Tube Vacuum	5000 Nos		
16.	Misc:	Glass Slide	1000 Nos		
17.	Misc:	HBsAg (Device)	5000 Nos		
18.	Misc:	HBV DNA PCR 96 test	30 Nos		
19.	Misc:	HBV DNA Extraction Kit 250 Test	15 Nos		
20.	Misc:	HCl N/ For Selectra	30 Nos		
21.	Misc:	HCV Device	5000 Nos		
22.	Misc:	HCV DNA PCR 96 test	30 Nos		
23.	Misc:	HCV RNA DNA Extraction Kit 250 Test	15 Nos		
24.	Misc:	HIV Device	3000 Nos		
25.	Misc:	ICT Malaria Device	1000 Nos		
26.	Misc:	Juster 100-1000 Micro	10 Nos		
27.	Misc:	Juster 10-100 Micro	10 Nos		
28.	Misc:	Juster 5-50 Micro LIT	10 Nos		
29.	Misc:	Juster 10-100 Micro LIT	10 Nos		
30.	Misc:	Micro Scope(student best scope) with LED Camera (UK, USA, Germany or equivalent)	02 Nos		



31.	Misc:	Micro Pipette EDTA	30 pkts		
S.#	Category	Name of the Articles	Req Qty	Packing	Manufacturer
32.	Misc:	Multi Sticks 10 parameter	100 bottles		
33.	Misc:	PCR reaction Tube 0.1 ml	200 pkts		
34.	Misc:	Plain Vacuum Tube Red Cape	5000 Nos		
35.	Misc:	Pregnancy Kit 1x 50	20 Boxes		
36.	Misc:	PT Bottles	4000 Nos		
37.	Misc:	Quality Control High	20 Nos		
38.	Misc:	Quality Control Normal	20 Nos		
39.	Misc:	Reaction Rotor for Selectra	30 Nos		
40.	Misc:	RNA DNA Free Gloves	100 Nos		
41.	Misc:	RNA DNA Free sample cup 1.5 ml	200 Nos		
42.	Misc:	RNA DNA Free sample cup 2 ml	200 Nos		
43.	Misc:	RNA DNA Free tips 01 ul to 10 ul	100 Nos		
44.	Misc:	RNA DNA Free tips 10 ul to 100 ul	100 Nos		
45.	Misc:	RNA DNA Free tips 100 ul to 1000 ul	100 Nos		
46.	Misc:	Simple Cup for Selectra 2 ml	1000 Nos		
47.	Misc:	Sodium Citrate Bottle Vacuum	4000 Nos		
48.	Misc:	Sunny Plaster	100 pkts		
49.	Misc:	Surgical Gloves	1000 Nos		
50.	Misc:	System Solution 1000 ml for Selectra	30 bottles		
51.	Misc:	T.B Rapid Device	500 Nos		
52.	Misc:	Test-tube Disposable	8000 Nos		
53.	Misc:	Test-tube Rack	50 Nos		
54.	Misc:	Thermo Shaker Incubate for PCR 24 Whole	01 No		
55.	Misc:	Typhoid Device	4000 Nos		
56.	Misc:	Trope T Device (UK, USA, Germany or equivalent)	2000 Nos		
57.	Misc:	Urine Continuer	10000 Nos		
58.	Misc:	Urine Multi Strip	50 Box		
59.	Misc:	Vacuum Tube for CP	20000 Nos		
60.	Misc:	VDRL	5,000 Nos		
61.	Misc:	ALBUMIN	1000		
62.	Misc:	ALKALINE PHOSPHATASE	8000		
63.	Misc:	ALT (SGPT)	8000		
64.	Misc:	AMYLASE	200		
65.	Misc:	ASOT QUANTITATIVE	200		
66.	Misc:	BILIRUBIN-TOTAL	8000		
67.	Misc:	BILIRUBIN-DIRECT	8000		
68.	Misc:	CALCIUM	600		
69.	Misc:	CHOLESTEROL	8000		
70.	Misc:	CK	200		
71.	Misc:	CREATININE	8000		



S.#	Category	Name of the Articles	Req Qty	Packing	Manufacturer
72.		ELECTROLYTE (Na,K,Cl)	4500		
73.		GAMAGT	900		
74.		GLUCOSE	7000		
75.		LDL CHOLESTEROL	500		
76.		HDL CHOLESTEROL	500		
77.		HbA1C	100		
78.		LDH	1500		
79.		MAGNESIUM	600		
80.		PHOSPHORUS	600		
81.		RA FACTOR QUANTITATIVE	300		
82.		TOTAL PROTEIN	600		
83.		TRIGLYCERIDES	600		
84.		UREA	7000		
85.		URIC ACID	300		
86.		HCO ₂ (BI-CARBONATE)	500		

Chemistry Auto analyzer (Specifications)

- Fully automated or Semi Automated Chemistry analyzer with random access system
- Through output 100 tests.
- Bench top random access clinical chemistry analyzer
- Several sample positions available.
- Reagents System Open.
- Sample type is serum / plasma/ urine/CSF, whole blood.
- 9 wave length 340,405,505,546,578,600,650 and 700 nm (Mono & Bio Chromatic)
- Refrigerated reagents control with peltier element.
- Outer segment with 50 bar codes readable positions
- All positions can be used for calibration / control / normal / pediatric and stat samples.
- State of the art host query interface is available.
- Host: Rs232 or Ethernet (TCP/IP) through LIS 2A Protocol.
- Temperature 16-30 C°
- Humidity <80 % RH
- Altitude up to 2000 m
- Electrical voltage 220-240 vat, or 110-120 frequency 50/60 Hz, power max 400A
- All Position can be used for calibration / control/ normal / pediatric and state sample.



S.#	Name of the Articles	Test Required	Manufacturer	Packing With MI/Test
1.	<u>CBC 03 Point Differential</u> Sample Volume: 14 µl of Whole Blood in normal 3-Part Mode, Parameters: 21 or more Hematology Parameters, including 3-Part Differential Reagent System: Isotonic Diluents Lyse, Cleaner. Throughput: 60 or more tests / hour. Sampling Method: Open tube system Sample Types: Whole Human Blood. Calibration: 1 or more Measurement Automatic and Manual (Factors) Calibration of WBC, HGB, RBC, PLT, MCV (Or HCT), RDW, and MPV. User Interface: Menu driven user interface with touch-screen Data Capacity: 50000 or More, Including RBC, PLT, and WBC 3-Part Histogram Printer Interface: Via USB port any Windows Compatible Printers Built-in Printer: Thermal Printer. Display: 320x240 -dots, high-contrast, backlit, color graphics LCD (Liquid Crystal Display) External Keyboard: USB keyboard	5000		
2.	<u>CBC 05 Part Differential</u> Throughput: 40 to 60 tests/hour Sample: Whole Human Blood Sample Volume: 80 to 110 ul of whole blood sampling method: closed and open tube sampling , auto sampler with integrated bar code reader and sample mixer Parameters: 26 CBC Parameters including 5 part WBC differential Reagent Sensing: YES Reagents: Dil-Diff, Lyse-5P, Diff-5P Calibration: Manual and Automatic mode Multi User Mode: Multilevel user modes with user ID User Interference: 600x800 color graphics LCD, touchscreen User Language: Multilingual user interface Software Upgrade: Via USB Port (USB Memory Stick) Data Storage Capacity: 70,000 to 100,000 records External Printing: Via USB port, any windows compatible printer External Key Board: Via PS/2 or USB Certificates: FDA Approved and CE Marked.	3000		



3.	COAGULATION:-			
	PT (Prothrombin Time)	8000		
	APTT	8000		
	SPECIFICATIONS OF COAGULATION AUTO ANALYZER:			
	Throughput: 100 or more tests/hour. Sample Capacity: 30 or more. Stat Sample: Available. Repetition: Automatic. Sample Volume Sensing: Yes Reagent Volume Sensing: Yes Clot Detection: Yes. Reloaded ability: Cuvettes, Sample, Reagents reloadable at Any time. Calibration: Fully Automatic. Method: Clotting, Chromogenic Immunologic. LIS Interface: Bar code reader and results interface able. Data Storage: more than 100000			

S.No	N.B			
01	Analyzers / Machines and Kits should be FDA approved or IVD CE marked and made in USA, Europe or Japan.			
02	Analyzers / Machines should be Brand New, refurbished are not acceptable.			
03	Clinical Chemistry Fully Auto analyzer:-			
	Supplier has to place One main and One backup Fully auto analyzer at Peoples Medical College Hospital Laboratory Nawabshah SBA. Main Chemistry Fully auto analyzer should have throughput of 200 tests /hour alone, and with ISE 800 tests per hour.			
04	Hematology Analyzer Fully Automatic or equivalent :-			
	Supplier has to place (02) 03 part differential analyzers at Peoples Medical College Hospital Laboratory Nawabshah SBA.			
5	Coagulation Analyzer Fully Automatic:-			
	Supplier has to place One Coagulation Fully Automatic analyzer with water bath at Peoples Medical College Hospital Laboratory Nawabshah SBA.			
06	All the Analyzers should be fully automated, with Barcode reader and results interface able. Supplier will be responsible for interfacing of their analyzer and bear its costs.			
07	Kit Sizes, expiry and stability should suit our work load requirements.			
08	The supplier will provide and take care of Computers, UPS, Bar code Scanners, Stabilizer and quality water for their Analyzer.			
09	The supplier will be responsible for providing Thermal / laser printer with their consumable for their analyzer; supplier will bear the cost of repair and maintenance. CBC Printing Roll and Printing paper for Reports. Provide Micro Scope for micro Scope			
10	The supplier will also be responsible for calibration of analyzer at their own expense.			
11	24 hours customer back up service is essential and supplier will bear the cost of repair and maintenance on weekly/ Monthly must provide.			
12	Certificate of analysis of kits is required.			
13	The date of manufacturing of the items expiry thereof batches number must be printed in Red on labels of bottles/containers/packages/ampoules / bags etc. The shelf life for Manufacturer's and imported items should not in any case be less than 70% at the time of supply / receipt.			







**PEOPLES MEDICAL COLLEGE HOSPITAL NAWABSHAH
SHAHEED BENAZIR ABAD**

**BIDDING DOCUMENT PURCHASE OF MEDICAL LABORATORY EQUIPMENT
FOR THE YEAR 2026-27**

S.#	Category	Name of the Articles	Req Qty	Amount in word of per No/ bottle/ pkt / box in	Amount in figure of per No/ bottle/ pkt / box in
1.	Misc:	Acid Solution for Selectra 1000 ml	10 bottles		
2.	Misc:	Blood Lancet	300 Nos		
3.	Misc:	Blue Tips	600 pkts		
4.	Misc:	Yalow.Tips	200 pkts		
5.	Misc:	Calibrator Chemistry for Selectra	20 Nos		
6.	Misc:	Capillary Tube No Heparin	50 Box		
7.	Misc:	Cover Slip 18 x 18	10000 Nos		
8.	Misc:	Cover Slip 22 x 22	10000 Nos		
9.	Misc:	Dengue kit	500 Nos		
10.	Misc:	Distil water 30 lit can	100 Nos		
11.	Misc:	E.S.R PIPPET (Disposable)	4000 Nos		
12.	Misc:	E.S.R stand	100 Nos		
13.	Misc:	EDTA Vacuum Tube	5000 Nos		
14.	Misc:	Field stain A+B	20 Nos		
15.	Misc:	Gel Tube Vacuum	5000 Nos		
16.	Misc:	Glass Slide	1000 Nos		
17.	Misc:	HBsAg (Device)	5000 Nos		
18.	Misc:	HBV DNA PCR 96 test	30 Nos		
19.	Misc:	HBV DNA Extraction Kit 250 Test	15 Nos		
20.	Misc:	HCl N/ For Selectra	30 Nos		
21.	Misc:	HCV Device	5000 Nos		
22.	Misc:	HCV DNA PCR 96 test	30 Nos		
23.	Misc:	HCV RNA DNA Extraction Kit 250 Test	15 Nos		
24.	Misc:	HIV Device	3000 Nos		
25.	Misc:	ICT Malaria Device	1000 Nos		
26.	Misc:	Juster 100-1000 Micro	10 Nos		
27.	Misc:	Juster 10-100 Micro	10 Nos		
28.	Misc:	Juster 5-50 Micro LIT	10 Nos		
29.	Misc:	Juster10-100 Micro LIT	10 Nos		
30.	Misc:	Micro Scope(student best scope) with LED Camera (UK, USA, Germany or equivalent)	02 Nos		
31.	Misc:	Micro Pipette EDTA	30 pkts		




S.#	Category	Name of the Articles	Req Qty	Amount in word of per No/ bottle/ pkt / box in	Amount in figure of per No/ bottle/ pkt / box in
32.	Misc:	Multi Sticks 10 parameter	100 bottles		
33.	Misc:	PCR reaction Tube 0.1 ml	200 pkts		
34.	Misc:	Plain Vacuum Tube Red Cape	5000 Nos		
35.	Misc:	Pregnancy Kit 1x 50	20 Boxes		
36.	Misc:	PT Bottles	4000 Nos		
37.	Misc:	Quality Control High	20 Nos		
38.	Misc:	Quality Control Normal	20 Nos		
39.	Misc:	Reaction Rotor for Selectra	30 Nos		
40.	Misc:	RNA DNA Free Gloves	100 Nos		
41.	Misc:	RNA DNA Free sample cup 1.5 ml	200 Nos		
42.	Misc:	RNA DNA Free sample cup 2 ml	200 Nos		
43.	Misc:	RNA DNA Free tips 01 ul to 10 ul	100 Nos		
44.	Misc:	RNA DNA Free tips 10 ul to 100 ul	100 Nos		
45.	Misc:	RNA DNA Free tips 100 ul to 1000 ul	100 Nos		
46.	Misc:	Simple Cup for Selectra 2 ml	1000 Nos		
47.	Misc:	Sodium Citrate Bottle Vacuum	4000 Nos		
48.	Misc:	Sunny Plaster	100 pkts		
49.	Misc:	Surgical Gloves	1000 Nos		
50.	Misc:	System Solution 1000 ml for Selectra	30 bottles		
51.	Misc:	T.B Rapid Device	500 Nos		
52.	Misc:	Test-tube Disposable	8000 Nos		
53.	Misc:	Test-tube Rack	50 Nos		
54.	Misc:	Thermo Shaker Incubate for PCR 24 Whole	01 No		
55.	Misc:	Typhoid Device	4000 Nos		
56.	Misc:	Trope T Device (UK, USA, Germany or equivalent)	2000 Nos		
57.	Misc:	Urine Continuer	10000 Nos		
58.	Misc:	Urine Multi Strip	50 Box		
59.	Misc:	Vacuum Tube for CP	20000 Nos		
60.	Misc:	VDRL	5,000 Nos		
61.	Misc:	ALBUMIN	1000		
62.	Misc:	ALKALINE PHOSPHATASE	8000		
63.	Misc:	ALT (SGPT)	8000		
64.	Misc:	AMYLASE	200		
65.	Misc:	ASOT QUANTITATIVE	200		
66.	Misc:	BILIRUBIN-TOTAL	8000		
67.	Misc:	BILIRUBIN-DIRECT	8000		
68.	Misc:	CALCIUM	600		
69.	Misc:	CHOLESTEROL	8000		
70.	Misc:	CK	200		
71.	Misc:	CREATININE	8000		



S.#	Category	Name of the Articles	Req Qty	Amount in word of per No/ bottle/ pkt / box in	Amount in figure of per No/ bottle/ pkt / box in
72.		ELECTROLYTE (Na,K,Cl)	4500		
73.		GAMAGT	900		
74.		GLUCOSE	7000		
75.		LDL CHOLESTEROL	500		
76.		HDL CHOLESTEROL	500		
77.		HbA1C	100		
78.		LDH	1500		
79.		MAGNESIUM	600		
80.		PHOSPHORUS	600		
81.		RA FACTOR QUANTITATIVE	300		
82.		TOTAL PROTEIN	600		
83.		TRIGLYCERIDES	600		
84.		UREA	7000		
85.		URIC ACID	300		
86.		HCO2 (BI-CARBONATE)	500		

Chemistry Auto analyzer (Specifications)

- Fully automated or Semi Automated Chemistry analyzer with random access system
- Through output 100 tests.
- Bench top random access clinical chemistry analyzer
- Several sample positions available.
- Reagents System Open.
- Sample type is serum / plasma/ urine/CSF, whole blood.
- 9 wave length 340,405,505,546,578,600,650 and 700 nm (Mono & Bio Chromatic)
- Refrigerated reagents control with peltier element.
- Outer segment with 50 bar codes readable positions
- All positions can be used for calibration / control / normal / pediatric and stat samples.
- State of the art host query interface is available.
- Host: Rs232 or Ethernet (TCP/IP) through LIS 2A Protocol.
- Temperature 16-30 C'
- Humidity <80 % RH
- Altitude up to 2000 m
- Electrical voltage 220-240 vat, or 110-120 frequency 50/60 Hz, power max 400A
- All Position can be used for calibration / control/ normal / pediatric and state sample.



S. #	Name of the Articles	Test Required	Manufacturer	Per Test Cost	Per Kit Cost
1	<u>CBC 03 Point Differential</u> Sample Volume: 14 µl of Whole Blood in normal 3-Part Mode, Parameters: 21 or more Hematology Parameters, including 3-Part Differential Reagent System: Isotonic Diluents Lyse, Cleaner. Throughput: 60 or more tests / hour. Sampling Method: Open tube system Sample Types: Whole Human Blood. Calibration: 1 or more Measurement Automatic and Manual (Factors) Calibration of WBC, HGB, RBC, PLT, MCV (Or HCT), RDW, and MPV. User Interface: Menu driven user interface with touch-screen Data Capacity: 50000 or More, Including RBC, PLT, and WBC 3-Part Histogram Printer Interface: Via USB port any Windows Compatible Printers Built-in Printer: Thermal Printer. Display: 320x240 -dots, high-contrast, backlit, color graphics LCD (Liquid Crystal Display) External Keyboard: USB keyboard	5,000			
2	<u>CBC 05 Part Differential</u> Throughput: 40 to 60 tests/hour Sample: Whole Human Blood Sample Volume: 80 to 110 ul of whole blood sampling method: closed and open tube sampling , auto sampler with integrated bar code reader and sample mixer Parameters: 26 CBC Parameters including 5 part WBC differential Reagent Sensing: YES Reagents: Dil-Diff, Lyse-5P, Diff-5P Calibration: Manual and Automatic mode Multi User Mode: Multilevel user modes with user ID User Interference: 600x800 color graphics LCD, touchscreen User Language: Multilingual user interface Software Upgrade: Via USB Port (USB Memory Stick) Data Storage Capacity: 70,000 to 100,000 records External Printing: Via USB port, any windows compatible printer External Key Board: Via PS/2 or USB Certificates: FDA Approved and CE Marked.External Key Board: Via PS/2 or USB Certificates: FDA Approved and CE Marked.	3,000			



S. #	Name of the Articles	Test Required	Manufacturer	Per Test Cost	Per Kit Cost
3	COAGULATION:-				
	PT (Prothrombin Time)	8000			
	APTT	8000			
	SPECIFICATIONS OF COAGULATION AUTO ANALYZER:				
	Throughput: 100 or more tests/hour. Sample Capacity: 30 or more. Stat Sample: Available. Repetition: Automatic. Sample Volume Sensing: Yes Reagent Volume Sensing: Yes Clot Detection: Yes. Reloaded ability: Cuvettes, Sample, Reagents reloadable at Any time. Calibration: Fully Automatic. Method: Clotting, Chromogenic Immunologic. LIS Interface: Bar code reader and results interface able. Data Storage: more than 100000				

OTHER DIRECTIONS (EVALATION CRITERIA):

1. TECHNICAL PROPOSAL should have the following documents.

- Original Tender Purchase Receipt.
- Copy of the Bid offer with quoted items with terms & conditions (without rates) with signature & Stamp.
- Photostat copy of Pay Order / Bank Draft of Bid Security showing without rates.
- Valid NTN / GST Certificate.
- Valid SST Certificate for Janitorial and Deployment of Security Guards
- Valid Professional Tax Certificate (From Excise and Taxation Department).
- An affidavit of Rs. 100/- (Non Judicial Stamp paper) that the firm is not blacklisted in any Government Department.
- Sample of quoted item should be provided at the time of opening of tender otherwise tender will be rejected.
- Three years' Experience and past performance in relevant field.
- Last Three Year Income Tax returns Challan.
- Three years Annual Turnover statement with Bank Certificate as per required in tender documents (must be year-wise).
- Soft Copy to be submitted in USB of complete Technical Proposal in MS Excel, Word or Searchable PDF files format, otherwise the bid will be ignored.

2. FINANCIAL PROPOSAL should have the following documents.

- Original Pay Order / Bank Draft of Bid Security.
- Original Copy of Bid offer with quoted price.
- Soft Copy in USB containing complete Financial Proposal in MS Excel, Word or Searchable PDF files format, otherwise the bid will be ignored.





GOVERNMENT OF SINDH
HEALTH DEPARTMENT
(PROCUREMENT MONITORING & INSPECTION CELL)

NOTIFICATION

No. SO(PM&I)/2026-27/E.28/(PMCHN): A Procurement Committee under Rule-7 of Sindh Public Procurement Rules 2010 (Amended up-to-date) is hereby constituted comprising of the following Officers for procurement of Drugs/Medicines (Local Purchase 15%) (other than items already approved under Frame Work Contract), Medical & Laboratory Equipment, Generic Consumables (Instrument / Consumable items), Other Miscellaneous items, Janitorial services, Diet of Patients, Lifesaving Medical Supplies (Oxygen), Security Guards, Uniform / Protective Clothing, in respect of Peoples Medical College Hospital Nawabshah, Shaheed Benazir Abad for the financial year 2026-27.

01	Medical Superintendent, Peoples Medical College Hospital Nawabshah Shaheed Benazirabad.	Chairman
02	Additional Medical Superintendent, Peoples Medical College Hospital Nawabshah Shaheed Benazirabad.	Member
03	Representative of Deputy Commissioner Nawabshah Shaheed Benazirabad.	Member
04	Representative of District Health Officer Nawabshah Shaheed Benazirabad.	Member
05	Senior Pharmacist, PMC Hospital Nawabshah Shaheed Benazirabad.	Member

TORs:

- The TORs/ Functions / Responsibilities of the Procurement Committee in accordance with Rule-8 of SPP Rules 2010 shall be as under:
- 1) Preparing and /or Reviewing bidding documents.
 - 2) Carrying out technical as well as financial evaluation of the bids;
 - 3) Preparing evaluation report as provided in Rule-45;
 - 4) Making recommendations for the award of contract to the competent authority; and
 - 5) Perform any other function ancillary and incidental to the above.

SECRETARY HEALTH
GOVERNMENT OF SINDH

No. SO(PM&I)/2026-27/E.28/(PMCHN):

Karachi dated, the 07th August 2026

Copy forwarded for information and necessary action to:-

1. The Managing Director, Sindh Public Procurement Regulatory Authority, Karachi.
2. The District Accounts Officer, Nawabshah Distt. Shaheed Benazir Abad.
3. The Chairman & all members of the Committee
4. PS to Minister, Health & Population Welfare Department, Govt. of Sindh, Karachi.
5. PS to Secretary Health, Govt. of Sindh, Karachi.
6. PS to Additional Secretary (PM&I), Health Department, Govt. of Sindh, Karachi.



(Signature)
(DEEN MUHAMMAD ABRO)
SECTION OFFICER (PM&I)





**GOVERNMENT OF SINDH
HEALTH DEPARTMENT**

NOTIFICATION

No. SO(PM&I)/2026-27/E.28/(PMCHN): A Complaint Redressal Committee under Rule-31 of Sindh Public Procurement Rules 2010 (Amended up-to-date) is hereby comprising the following Officers for scrutinizing the complaints of aggrieved bidders against tender invited by the Medical Superintendent, Peoples Medical College Hospital Nawabshah, Shaheed Benazirabad for the financial year 2026-27.

01	Director Health Services Nawabshah , District Shaheed Benazirabad.	Chairman
02	District Accounts Officer, District Shaheed Benazirabad.	Member
03	An independent professional from the relevant field.	Member

ToR's:

- To scrutinize the complaints from the aggrieved bidders and decide the cases strict in accordance with SPP Rules 2010.


**SECRETARY HEALTH
GOVERNMENT OF SINDH**


No. SO(PM&I)/2026-27/E.28/(PMCHN):

Karachi, dated, the 07th August 2026

Copy forwarded for information and necessary action to :

1. The Managing Director, Sindh Public Procurement Regulatory Authority Karachi.
2. The Chairman and all members of Complaint Redressal Committee.
3. The District Accounts Officer, Nawabshah Distt. Shaheed Benazirabad.
4. PS to Minister Health Govt. of Sindh, Karachi.
5. PS to Secretary, Health Department, Govt. of Sindh, Karachi.
6. PS to Additional Secretary (PM&I), Health Department, Govt. of Sindh, Karachi.


(DEEN MUHAMMAD ABRO)
SECTION OFFICER (PM&I)



DAILY AWAMI AWAZ

DATED 14-08-2026.



روزاني
عوامي آواز

سڄي دنيا جي لاءِ
CERTIFIED
ABC

ڪراچي، حيدرآباد ۽ سکر مان هڪ ئي وقت شايع ٿيندڙ

سال 37، شمارو 215، جمعو 14 آگسٽ 2026، 30 صفر المظفر 1448ھ، جلد 12، قیمت 50 روپيا

ٽينڊر نوٽيس

سنگل لٽيج - ٽراپيڪل پروسيسر ڊيٽ لائن تي بي آر اي يون 2010 (ڊسمبر قبل جي لاڳو ٿيڻ جي مطابق) ميڊل ميڊيڪل ڪاليج اسپتال، شهيد بينظير آباد لاءِ مالي سال 27-2026 لاءِ هيٺين خدمتون/شپن لاءِ فعال SST/SRB/GST/NTN رجسٽرڊ فرمن (جن جي لاڳو هجي) ۽ سندن مڃاڻ نڪاتن کان مهيننڊ ٿيل ٽينڊر طلب ڪيا ويا آهن.

واڪ صرف EPADS سسٽم ذريعي جمع ڪرايا ويندا. هتي آڇون قبول نه ڪيون وينديون. دلچسپي رکندڙ واڪ ڏيندڙن کي پنهنجون آڇون جمع ڪرائڻ لاءِ EPADS تي سٽ پروڪيورمينٽ پورٽل/EPADS تي رجسٽر ڪرائڻ جي ضرورت آهي.

ٽينڊر دستاويزن تي نوٽيس جي اشاعت جي تاريخ کان 07 سيپٽمبر 2026 تائين منجهند 03:00 وڳي تائين هيٺ ذڪر ڪندڙ جي آفيس مان فرم جي اصل ليٽر هيد تي هڪ تحريري درخواست جمع ڪرائڻ سان گڏ 2,000 روپيا جي ناقابل واپسي ٽينڊر دستاويز ڦيرس سان گڏ حاصل ڪري سگهجن ٿا.

ٽينڊر EPADS ذريعي 08 سيپٽمبر 2026 تائين منجهند 01:00 وڳي تائين جمع ڪرايا ويندا. جيڪڏهن ٽينڊر ڊويزن/لڳا ساڳي ڏينهن منجهند 02:00 وڳي، ميڊيڪل سپر اسپتال، PMCH، نوابشاهه جي آفيس وٽ، واڪ ڏيندڙن يا ٽين جي مڃاڻ نڪاتن جي موجودگي و گولڊ وينڊا جيڪي شريعت طرفن ڇاهيندا.

خرپاري اجنسي SPPRA رولز 2010 (جسٽن ٽرمس هڪيل) ۽ واڪ دستاويزن ۾ شامل شرط ۽ ضابطن جي مطابق هٿن ۾ يا سڀني واڪن کي قبول يا رد ڪرڻ جو حق محفوظ رکي ٿي.


سلسلو	تفصيل	ٽينڊر خرپاري تاريخ	واڪ دستاويز جمع ڪرائڻ ۽ کولڻ جي تاريخ	ٽينڊر جي لاڳت
1	ميڊيڪل لائبريري ايپروپريٽ جي خرپاري	اشاعت جي تاريخ کان وٺي 07 سيپٽمبر 2026 منجهند 03:00 وڳي تائين	08 سيپٽمبر 2026	Rs. 2,000
2	هندو ٽيڪسٽ بڪس ٽرڪس ۽ ميڊيڪل 15 ٽوڪل پريچر (I.P) جي خرپاري (پارٽ 8 بلڪل) (پارٽ 3 روزا)	ساڳو	ساڳو	Rs. 2,000

نوٽ

« سڀني واڪ جيڪي 2 سيپٽمبر 2026 سڳوڙي، ميڊيڪل سپر اسپتال، PMCH، نوابشاهه جي آفيس جي آڏو ڏنل هجن گهرجن.

« واڪ يا ٽيڪسٽ بڪس ۽ ٽرڪس لاءِ جيڪي ٽيڪسٽ بڪس ڏنل هجن گهرجن.

« جيڪڏهن جيڪو ڪو به ڪوٺيل جواب ڏئي ٿي، انهن لاءِ ڪو به ڪوٺيل جواب ڏيڻ جي سڳي سڳي نه ڪوٺيل ۽ ڪوٺيل.



PEOPLES MEDICAL COLLEGE HOSPITAL
Nawabshah



[illegible][illegible]

نمبر شار	تفصیل	نیزہ خرچے کے تاریخ	بذریعہ کی دستاویزات پیش کرانے اور مکمل کرنے کی تاریخ	نیزہ کی ناقابل واپسی ٹیکس
1	طبی تشخیصی لیپارکری کے آلات کی خریداری	اشاعت کی تاریخ سے 07 ستمبر 2026 تک 3:00 بجے تک	08 ستمبر 2026	2,000
2	بیمہ کے اقسام ہونے والی اشیاء	ایسا	ایسا	2,000
3	اوربٹ اور وائس کی خریداری — 15% مقامی خریداری (LP) (حصہ A: یک روزہ نقدی سرمایہ کاری) (حصہ B: روزمرہ کی بنیاد پر خریداری)	ایسا	ایسا	2,000

[illegible]

میڈیکل سیرنڈنٹ
پیپلز میڈیکل کالج ہسپتال، نواب شاہ
شہید بیظیر آباد

WORK FOR INDIA
www.workforindia.com



آوارہ 16 ربیع الاول 1448ھ 30 اگست 2026ء
 www.jang.com.pk
 JangDotComDotPk @ jangdotcom جang abdulaziz
 پاکستان کے روزنامہ گراہیہ
 ABC CERTIFIED
 باقاعدہ تصدیق شدہ شاعت
 THE JANG KARACHI
 روزنامہ جنگ
 جی جی
 میر ظلیل الرحمن
 بانی
 90 سالہ
 آوارہ 16 ربیع الاول 1448ھ 30 اگست 2026ء
 236 نمبر



**تصحیحی نوٹس
بولی جمع کرانے کی آخری
تاریخ میں توسیع**

۱۔ تین مہینے کی مدت میں ۱۰۰ فیصد کی شرح پر (تعمیری) منسلک
۲۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۳۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۴۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۵۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۶۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۷۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۸۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۹۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر
۱۰۔ (تعمیری) ۱۰۰ فیصد کی شرح پر (تعمیری) ۱۰۰ فیصد کی شرح پر

INF 157 3775 26

I WORK FOR SINDH  THE SINDH SAHITYA AKADEMI



**ANNUAL PROCUREMENT PLAN
FINANCIAL YEAR 2026-27**

PEOPLES MEDICAL COLLEGE HOSPITAL NAWABSHAH SHAHEED BENAZIR ABAD

S.No	Description of Procurement	Required Quantity	Estimated total Cast	Source of Funds ADP/ Non ADP	Proposes procurement Method	Timing of Procurement			
						1 st Quarter	2 nd Quarter	3 rd Quarter	4 th Quarter
						Single Stage two Envelop	34,344,300	34,344,300	34,344,300
						Single Stage two Envelop	2,015,500	2,015,500	2,015,500
						Single Stage two Envelop	1,904,500	1,904,500	1,904,500

1.	Purchase of Drug and Medicine 15% L.P (Bulk & Day to Day)	as Per Required	137,377,200	Non ADP	Single Stage two Envelop	34,344,300	34,344,300	34,344,300	34,344,300
2.	Medical and Laboratory Equipment	as Per Required	8,062,000	Non ADP	Single Stage two Envelop	2,015,500	2,015,500	2,015,500	2,015,500
3.	Generic Consumables (Instrument / Consumable items	as Per Required	7,618,000	Non ADP	Single Stage two Envelop	1,904,500	1,904,500	1,904,500	1,904,500

MEDICAL SUPERINTENDENT
Peoples Medical College Hospital Nawabshah
Shahed Benazir Abad

(Signature)

