NATIONAL

OFFICE OF THE DISTRICT HEALTH OFFICER DISTRICT THATTA

NOTICE INVITING TENDER

District Health Officer District Thatta invites e-bids through E-Pak Acquisition and Disposable System (EPADS) from Suppliers/firms/Manufacturers / Importers / Authorized Distributors on Active Taxpayers List of the FBR/SRB (whichever is applicable) for the supply of Medical Equipment's, Machinery, Instruments, Furniture / Fixture under following ADP Scheme 876, 877

01	876	UP GRADATION OF DISTRICT COUNCIL DISPENSARY MEHAR CITY TO THE LEVEL OF BHU AT UC MEHAR, TALUKA GHORABARI
02	877	ESTABLISHMENT OF (02) DESPENSARIES NAMELY JAM BIJAR AT VILLAGE JAFAFAR JOKHIO, UC BHAMBHORE, TALUKA MIRPUR SAKRO AND VILLAGE BABU MALLAH, UC CHOWBANDI, TALUKA MIRPUR SAKRO

for the financial year 2026-27 & 2027-28 under *Rule-46(2) Single Stage-Two Envelopes Procedure.* Bidding documents containing detailed Terms & Conditions, can be viewed /downloaded from https://portalsindh.eprocure.gov.pk/#/ Electronic Bids should be submitted through EPADS ONLY, Manual bid shall not be received. Interested Bidders are required to register themselves on EPAD System at the link: https://sindh.eprocure.gov.pk/#/supplier/registration for submission of electronic-bids.

The Bids, prepared in accordance with the instructions in the bidding documents, must be submitted on EPADS by 30-11-2026 at 11:00 A.M. The original instrument of tender fee Rs.3,000/- and fixed bid security mentioned against quoted items must reach the procuring agency before the deadline for submission of e-bids, which will be opened on the same day at 11:30 A.M. at the address @ Fatima Engineer Auditorium, Jinnah Sindh Medical University (JSMU), Karachi.

N.B:-

- Any query for e-bidding may contact at District Health Officer District Thatta contact No. 0310 0031823
- In case Govt. announces any Public Holiday or any unfavorable circumstances, the tender / bids will be submitted and opened on next working day, with same Venue and Time.
- The Purchaser reserves the right to reject any / all bids under the relevant provisions of SPP Rules 2010.
- In case of any difficulty prospective bidders may contact EPADS Helpline 051-111-137-237 during working days/hours.

DISTRICT HEALTH OFFICER
DISTRICT THATTA

INF-KRY 4900/26
iWORK FOR SINDH
www.iwork4sindh.com
JOB PORTAL BY INFORMATION DEPARTMENT

Phone: (+92) 51-9205000

از دفتر ڈسٹرکٹ ہیلتھ آفیسر، ضلع ٹھٹھہ

نوٹس طلبی ٹینڈر

ڈسٹرکٹ ہیلتھ آفیسر، ضلع ٹھٹھہ، کو مالی سال 27-2026 اور 28-2027 کے لیے، قاعدہ 46(2) کے تحت سنگل اسٹیج، دولفافہ طریقہ کار کے مطابق اے ڈی پی اسکیم نمبر 876 اور 877 کے تحت طبی آلات، مشینری، آلات، فرنیچر/فکسچر کی فراہمی کے لیے ای پاک ایکوزیشن اینڈ ڈسپوزل سسٹم (EPADS) کے ذریعے سپلائرز، فرموں، مینوفیکچررز/درآمد کنندگان/مجاز ڈسٹری بیوٹرز سے ای بڈز مطلوب ہیں۔ بولی دہندگان کے لیے ضروری ہے کہ وہ ایف بی آر/ایس آر بی (حسب اطلاق) کی فعال ٹیکس دہندگان کی فہرست میں شامل ہوں۔

01	876	یوسی مہر، تعلقہ گھوڑا باڑی میں ڈسٹرکٹ کونسل ڈسپنسری مہر سٹی کو بنیادی مرکز صحت (BHU) کی سطح پر اپ گریڈ کرنا۔
02	877	دو ڈسپنسریوں کا قیام، جن کے نام جام بجار، واقع گاؤں جعفر جوکھیو، یونین کونسل بھنبھور، تعلقہ میر پور سکھرو، اور گاؤں بابو ملاح، یونین کونسل چوہندی، تعلقہ میر پور سکھرو

تفصیلی شرائط وضوابط پر مشتمل بولی کے دستاویزات درج ذیل ویب سائٹ سے دیکھی اور ڈاؤن لوڈ کی جاسکتی ہیں:

https://portalsindh.eprocure.gov.pk/#/

الیکٹرانک بولیاں صرف EPADS کے ذریعے جمع کرائی جائیں گی۔ دستی بولیاں وصول نہیں کی جائیں گی۔ دلچسپی رکھنے والے بولی دہندگان کو الیکٹرانک بولیاں جمع کرانے کے لیے درج ذیل لنک پر EPADS سسٹم میں اپنی رجسٹریشن کروانا ضروری ہے:

https://sindh.eprocure.gov.pk/#/supplier/registration

بولی کے دستاویزات میں درج ہدایات کے مطابق تیار کردہ بولیاں 30 نومبر 2026 کو صبح 11:00 بجے تک EPADS پر جمع کرانا ضروری ہیں۔ ٹینڈر فیس مبلغ -/3,000 روپے کا اصل ادائیگی کا ثبوت اور پیش کردہ اشیا کے مقابل درج مقررہ بڈ سیکورٹی، ای بڈز جمع کرانے کی آخری تاریخ سے پہلے پروکیورنگ ایجنسی تک پہنچ جانا چاہیے، بولیاں اسی روز صبح 11:30 بجے درج ذیل پتہ: فاطمہ انجینئر آڈیٹوریم، جناح سندھ میڈیکل یونیورسٹی (JSMU)، کراچی پر کھولی جائیں گی:

نوٹ:

- الیکٹرانک بولی سے متعلق کسی بھی استفسار کے لیے ڈسٹرکٹ ہیلتھ آفیسر، ضلع ٹھٹھہ رابطہ نمبر: 0310-0031823 سے رابطہ کیا جاسکتا ہے۔
- اگر حکومت کی جانب سے کوئی عام تعطیل کا اعلان کیا جاتا ہے یا کوئی ناموافق صورت حال پیدا ہوتی ہے تو ٹینڈر/بولیاں اگلے کاروباری روز اسی مقام اور مقررہ وقت پر جمع اور کھولی جائیں گی۔
- خریدار کو سندھ پبلک پروکیورمنٹ (SPP) رولز 2010 کی متعلقہ دفعات کے تحت کسی بھی یا تمام بولیوں کو مسترد کرنے کا حق حاصل ہے۔

کسی بھی دشواری کی صورت میں ممکنہ بولی دہندگان کام کے دنوں اور اوقات میں EPADS ہیلپ لائن 051-111-137-237 پر رابطہ کرسکتے ہیں۔

ڈسٹرکٹ ہیلتھ آفیسر
ضلع ٹھٹھہ

INF-KRY 4900/26

iWORK FOR SINDH
www.iwork4sindh.com
JOB PORTAL BY INFORMATION DEPARTMENT

...صوبہ ہے، جس کی فنڈنگ وزارتِ انفارمیشن ٹیکنالوجی اینڈ ٹیلی کمیونیکیشن اور اگنائٹ جانب سے کی جاتی ہے، جبکہ اس کا انتظام پی ٹی سی ایل اور ...یل ایم کے ٹی کے پاس ہے۔ سال 2022 میں ...صوبہ سندھ کے دوسرے بڑے انکیوبیشن سینٹر کے ...طور پر قائم ہونے والا این آئی سی حیدرآباد اسٹارٹ ...پس کو مینٹرشپ، سرمایہ کاروں تک رسائی، صنعتی ...روابط، منظم کاروباری معاونت اور قومی و بین الاقوامی سطح پر پبلسٹی کے مواقع فراہم کرتا ہے۔

...سے بڑے مسائل، سہیل راجپوت

...وی پازیٹو ہوگئے، متاثرین کو انصاف دلایا

...نے کراچی سے آنے والی ایک درخواست پر نوٹس لیا تو معلوم ہوا کہ کراچی کے صنعتی علاقے میں سندھ ...سپلائز سوشل سیکیورٹی انسٹیٹیوشن کا سرکاری اسپتال کلثوم ویلکا میں ایک ہی سرنج سے کئی کئی مریضوں کو انجکشن لگانے کے باعث علاقے کے 170 بچے ایچ آئی وی پازیٹو ہوگئے جس میں درخواست دہندہ کے ...بچے بھی شامل تھے جس پر جب کارروائی کی تو 64 ...افسران و عملے کے لوگ ملوث پائے گئے جو بجٹ بچانے کے لیے کم سے کم سرنج کا استعمال کررہے تھے جس پر متعلقہ افسران کو کارروائی کی ہدایت بھی کی اور ...ان بچوں کا کراچی کے بڑے نجی اسپتالوں سے ...ٹریٹمنٹ شروع ہوا اور ان کے لیے تعلیم کا سلسلہ بھی شروع کرا دیا ہے۔ میں نے سال 2024 میں ...صوبائی محتسب کا چارج لیا تو اس وقت ہر سال 8 ہزار شکایتیں درج ہوتی تھیں۔ سہیل راجپوت نے ...بتایا کہ صوبائی محتسب اعلیٰ کے پاس سب سے زیادہ ...پانی، بجلی اور پولیس کے متعلق شکایات موصول ہوتی ...ہیں تاہم انہیں حیدرآباد کے بارے میں معلومات ...ملتی رہتی ہیں، یہاں بھی بہت مسائل ہیں جنہیں ترجیح ...بنیادوں پر حل کرنے کی ضرورت ہے۔ رانی باغ ...سمیت حیدرآباد کے دیگر تفریحی مقامات پر خصوصی ...توجہ دی جانی چاہیے۔ قبل ازیں صدر حیدرآباد ...پریس کلب جنید خانزادہ نے کہا کہ حیدرآباد کے ...صحافیوں کیلئے خصوصی گرانٹ کی منظوری کے لیے ...سہیل راجپوت نے بہت اہم کردار ادا کیا اور پہلی ...مرتبہ ڈیڑھ کروڑ روپے کی گرانٹ سندھ کابینہ سے ...منظور کروائی تھی جس میں وقت کے ساتھ ساتھ ...اضافہ ہوتا گیا، بعدازاں سینئر صحافی لالہ رحمان سومرو ...نے سہیل راجپوت کو اجرک اور سینئر صحافی علی حسن نے ...سندھی ٹوپی کا تحفہ پیش کیا۔ اس موقع پر صوبائی محتسب ...اعلیٰ کے رجسٹرار کیمشر سجاد حیدر، افتخار احمد، شبیر علی بابر، ...محمود قادری، سیکریٹری حیدرآباد پریس کلب اشوک ...شرما، سینئر صحافی محمد حسین خان، حامد شیخ، حمید الرحمان ...اور دیگر صحافی بھی موجود تھے۔

روزنامہ ایکسپریس

01/11/2026

ABC Certified
روزاني
پنهنجي اخبار
ايڈيٽر
محمد علي قاضي
ڪراچي
epaper.pahenjiakhbar.com
جلد 7 آچر پهرين نومبر2026ع بمطابق 20 جمادي الاول 1448 هـ شمارو 355 قيمت 40 رپيا

آفيس آف دي
ڊسٽرڪٽ هيلٿ آفيسر
ڊسٽرڪٽ ٺٽو

ٽينڊر گھرائڻ جو نوٽيس

ڊسٽرڪٽ هيلٿ آفيسر ڊسٽرڪٽ ٺٽو پاران هيٺ ڄاڻايل اي ڊي پي اسڪيمن 876 ۽ 877 تحت طبي اوزارن، مشينري، طبي آلات ۽ فرنيچر/ فڪسچر جي فراهمي لاءِ، ايف بي آر/ايس آر بي (جيڪو به لاڳو ٿئي) جي فعال ٽيڪس ڊيندڙن جي فهرست ۾ شامل سپلائرز/ فرمن/ مينوفيڪچررز/ امپورٽرز/ ۽ بااختيار ڊسٽريبيوٽرز کان اي-پيڊز (EPADS) ذريعي اليڪٽرانڪ آڇون طلب ڪجن ٿيون.

01	876	ضلعي ڪائونسل ڊسپينسري ميڙ شهر ڪي بي سي ميڙ، تعلقي گھوڙاباري ۾ بنيادي صحت مرڪز (BHU) جي درجي تائين اپ گريڈ ڪرڻ.
02	877	تعلقي ميرپور ساڪرو جي يونين ڪائونسل پنهور جي ڳوٺ جعفر جوڪيو ۾ جام بچار ڊسپينسري ۽ يونين ڪائونسل چوبندي جي ڳوٺ بابو ملاح ۾ ڊسپينسري سميت مجموعي طور پن (02) ڊسپينسرين جو قيام.

مالي سال 27-2026ع ۽ 28-2027ع لاءِ رولز 46(2) تحت هڪ مرحلو به لفافا طريقيڪار مطابق خريداري ڪئي ويندي. تفصيلي شرطن ۽ ضابطن تي مشتمل آڇ دستاويز ويب سائيٽ https://portalsindh.eprocure.gov.pk/#/ تي ڈسي يا ڊائون لوڊ ڪري سگھجن ٿا.

اليڪٽرانڪ آڇون صرف EPADS ذريعي جمع ڪرائڻيون پونديون. هٿرادو آڇون قبول نه ڪيون ويندیون. دلچسپي رکندڙ آڇ ڊيندڙن لاءِ لازمي آهي ته اهي اليڪٽرانڪ آڇون جمع ڪرائڻ لاءِ لنڪ https://sindh.eprocure.gov.pk/#/supplier/registration ذريعي EPADS نظام تي پنهنجي رجسٽريشن ڪرائين.

آڇ دستاويز، جيڪي آڇ دستاويزن ۾ ڈنل هدايتن مطابق تيار ڪيل هجن اهي 30-11-2026 صبح 11:00 وڳي تائين EPADS تي جمع ڪرائڻ لازمي آهن. ٽينڊر فيس 3,000 رپيا ۽ پيش ڪيل شين جي سامهون ڄاڻايل مقرر ڪيل آڇ جي ضمانتي رقم جا اصل دستاويز. اليڪٽرانڪ آڇون جمع ڪرائڻ جي آخري وقت کان اڳ خريداري ڪندڙ اداري وٽ پهچڻ گھرجن. آڇون ساڳئي ڏينهن صبح 11:30 وڳي فاطمه انجنيئر آڊيٽوريم، جناح سنڌ ميڊيڪل يونيورسٽي (JSMU)، ڪراچي ۾ کوليون ويندیون.

نوٽ:

- اليڪٽرانڪ آڇن بابت ڪنهن به قسم جي پڇا ڳاڇا لاءِ ڊسٽرڪٽ هيلٿ آفيسر، ڊسٽرڪٽ ٺٽو سان فون نمبر 0310-0031823 تي رابطو ڪري سگھجي ٿو.
- جيڪڏهن حڪومت طرفان ڪنهن به قسم جي عام موڪل جو اعلان ڪرڻ يا ڪا اٽوٽندڙ صورتحال پيدا ٿئي جي صورت ۾ آڇون وڌندڙ ڪم ڪار واري ڏينهن تي ساڳئي هنڌ ۽ ساڳئي وقت مطابق جمع ڪرايون ۽ کوليون ويندیون.
- پرچيز ڪي سنڌ پبلڪ پروڪيورمينٽ رولز 2010ع جي لاڳاپيل شقن تحت ڪنهن به يا سموريون آڇن ڪي رد ڪرڻ جو اختيار حاصل هوندو.
- ڪنهن به قسم جي ڏکيائي جي صورت ۾ امڪاني آڇ ڏيندڙ ڪم ڪار واري ڏينهن / آفيس وقت دوران EPADS هيلپ لائن نمبر 051-111-137-237 تي رابطو ڪري سگھن ٿا.

صحيح
ڊسٽرڪٽ هيلٿ آفيسر
ڊسٽرڪٽ ٺٽو

INF/KRY.No.4900/26

iWORK FOR SINDH
A JOB PORTAL BY INFORMATION DEPARTMENT
www.iwork4sindh.com

GOVERNMENT OF SINDH
HEALTH DEPARTMENT

NOTIFICATION

No.SO(PM&I)2-1/2022-23(Main)/PC: In supersession of this Department's Notification of even number dated: 20.07.2022, 18.10.2023, 03.11.2023, 04.01.2024 & 29.01.2025 and with the approval of Hon'ble Chief Minister Sindh vide CMIT Summary, the Health Department's Procurement Committee (PC) for procurement of Drug/Medicines, Surgical/Disposable items, X-Ray Films / Chemicals, Machinery Equipments, Furniture / Fixture (etc.) is hereby re-notified as under:

1.	Professor Dr. Amjad Siraj Memon, Vice Chancellor, Jinnah Sindh Medical University (JSMU), Karachi.	**Chairman**
2.	Vice Chancellor, Liaquat University Medical & Health Sciences (LUMHS), Jamshoro.	Member
3.	Vice Chancellor, Peoples University of Medical & Health Sciences for Women, Nawabshah-Shaheed Benazirabad.	Member
4.	Vice Chancellor, Shaheed Mohtarma Benazir Bhutto Medical University Larkana.	Member
5.	Prof. Dr. Sajida Qureshi, Professor of Surgery, DUHS / Dr. Ruth K.M. Pfau CHK.	Member
6.	Chief Executive Officer, Health Care Commission Sindh.	Member
7.	Special Secretary, Chief Minister's Inspection, Enquiries & Implementation Team Department, Govt. of Sindh.	Member
8.	Additional Secretary (GA), SGA&CD, Govt. of Sindh.	Member
9.	Director General Health Services Sindh Hyderabad.	Member
10.	Executive Director, Jinnah Postgraduate Medical Centre Karachi.	Member
11.	Medical Superintendent, Dr.Ruth K.M. Pfau Civil Hospital Karachi.	Member
12.	Medical Superintendent, Liaquat University Hospital Hyderabad.	Member
13.	Medical Superintendent, Ghulam Muhammad Mahar Medical College Sukkur.	Member
14.	Medical Superintendent, Peoples Medical College Hospital Nawabshah, SBA.	Member
15.	Additional Secretary (PM&I), Health Department, Govt. of Sindh.	Member / Secretary

TOR's:

a) Preparing and /or reviewing bidding documents;
b) Carrying out technical as well as financial evaluation of the bids;
c) Preparing evaluation report as provided in Rule 45;
d) Making recommendations for the award of contract to the Competent Authority.
e) Monitoring of the procured medicines with regard to their quality and genuiness.
f) The Committee shall be authorised to have vigilance on distribution of the supply to different Hospitals including all health facilities.
g) The Committee would be empowered to Co-opt. Expert(s) incase the committee desires any assistance in any specific/relevent fields.
h) The committee shall have all the powers to procure the medical/surgical equipments, machines and furniture or any other item to be installed and / or fixed within the Hospital in Sindh province at any level whether Tertiary, District & Taluka Hospitals.
i) The powers of the Procurement Committee would extend to all the projects under the Health Department for procurement of medicines, furniture, instruments & machines etc.
j) Perform any other function ancillary and incidental to the above;

SECRETARY HEALTH

No. SO(PM&I)2-1/2022-23(Main)/PC:


GOVERNMENT OF SINDH PM & I CELL HEALTH DEPARTMENT

Karachi, dated 26th October 2026

26-10-2026

C.C. to:

1. The Chairman and all Members of the Committee.
2. The Accountant General Sindh, Karachi.
3. The Managing Director, Sindh Public Procurement Regulatory Authority, Karachi.
4. The PS to Chairman, CMIE&ITD Karachi.
5. The PS to Chief Executive Officer, Health Care Commission Sindh.
6. The PS to Special Secretary, CMIE&ITD, Karachi.
7. The Deputy Secretary (Staff) to Chief Minister Sindh / Chief Secretary Sindh.
8. The PS to Minister Health, Social Welfare and Public Health Engineering Dept. Sindh, Karachi.
9. The PS to Secretary Health, Govt. of Sindh, Karachi.
10. The PS to Addl. Secretary (PM&I), Health.



26-10-2026

(DEEN MUHAMMAD ABRO)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH
HEALTH DEPARTMENT
(PROCUREMENT MONITORING & INSPECTION CELL)

NOTIFICATION

No. SO(PM&I)2-1/2022-23/PC(TEC-III): In supersession to this department's notification of even number dated: 13.10.2023, on recommendation of Health Procurement Committee (HPC) and with the approval of worthy Secretary Health, the Experts Committee for Machinery / Equipment / Lab. Equipment / Furniture & Fixture etc. is hereby re-constituted with the following composition and ToRs:

i.	Dr. Shahid Pervez, Histopathologist,Aga Khan University Hospital Kyc.	Chairman
ii.	Dr. Benish Sultan, Associate Professor & Head of Pathology Lab, JSMU, Karachi.	Member
iii.	Prof. Dr. Tariq Mehmood, Representative of PAF, JPMC, Karachi.	Member
iv.	Dr. Syed Farjad Sultan, Associate Professor, Anaesthesia Department, DUHS / Civil Hospital Karachi.	Member
v.	Prof. Dr. Sajida Qureshi, Professor of Surgery, DUHS / Dr. Ruth K.M. Pfau Civil Hospital Karachi.	Member
vi.	Prof. Dr. Syed Mahmood Hassan, Professor of Pathology, JSMU, Karachi.	Member
vii.	Dr. Shoaib Malik, Associate Professor, Anesthesia Department, JPMC Karachi.	Member

TORs:

a) To revisit and approve proposed specifications of Machinery / Equipment etc. prepared by the Technical Committee.
b) To approve bidding documents prepared by Technical Committee along-with verification / certification of technical evaluation reports finalized by Technical Committee.
c) The committee will also check the samples as per given specification in the tender document (wherever applicable).
d) To submit final recommendation to the Procurement Committee within seven (07) days for approval.

SECRETARY HEALTH

No. SO(PM&I)2-1/2022-23/PC(TEC-III): Karachi, dated the 18th February 2025

A copy is forwarded for information & necessary action to:-

1. The Chairman, Health Procurement Committee / VC, JSMU, Karachi.
2. The Chairman and all members of the Experts Committee.
3. PS to Minister Health, Social Welfare and Public Health Engineering Dept. Sindh, Karachi.
4. PS to Secretary Health.

18/2/2025

(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH
PM & I CELL
HEALTH DEPARTMENT

GOVERNMENT OF SINDH
HEALTH DEPARTMENT
(PROCUREMENT MONITORING & INSPECTION CELL)

NOTIFICATION

No. SO(PM&I)2-1/2022-23/PC(TEC-II): In supersession to this Department's Notification of even number dated: 15.08.2022, on the recommendation of the Chairman, Health Department's Procurement Committee, Technical Experts Committee for Machinery / Equipment is hereby re-constituted comprising on following, to assist the Health Procurement Committee for preparation of bidding documents and evaluation of tender of companies / bidders for purchase of Machinery / Equipment / Furniture etc. for the year 2022-23:

a.	Mr.S.Zamir Hussain Shah,Bio Medical Engineer,Dr.Ruth K.M.Pfau Civil Hosp. Kar.
b.	Mr. Abdul Majid, Electrical Engineer, NICH, Karachi.
c.	Mr. Qazi Muhammad Amir, Bio-medical Technician, Elector-medical Workshop.
d.	Mr. Sunny Mohiuddin, Dy. Director (Procurement),Directorate of Pharmacy,Sindh, Kar.
e.	Mr. Rajeev, Pharmacist, Dr. Ruth K.M. Pfau Civil Hospital Karachi.

ToRs:

i. To prepare bidding documents, Specifications and Evaluation Criteria.
ii. To review existing formulary list and segregate into seven (07) Therapeutic Groups and submit to technical experts / leading professors of respective therapeutic groups for approval.
iii. To carry out scrutiny of submitted bidding documents.
iv. To prepare Bid Qualification Report (BQR), Bid Evaluation Report (BER) and submit to respective therapeutic groups experts for review and approval.
v. To prepare financial comparative statement and final Bid Evaluation Report (BER) and get endorsement from respective experts of therapeutic groups before final approval by Health Department Procurement Committee.
vi. To submit scrutinized working to Experts Committee constituted for Machinery / Equipment / Furniture / Fixture & Lab Equipment etc.

SECRETARY HEALTH

No. SO(PM&I)2-1/2022-23/PC(TEC-II): Karachi, dated the 01st February 2023

A copy is forwarded for information & necessary action to:-

1. The Chairman, Health Procurement Committee / VC, JSMU, Karachi.
2. The Executive Director, NICH, Karachi.
3. The Medical Superintendent, Dr. Ruth K.M.Pfau Civil Hospital Karachi.
4. All members of the Technical Experts Committee (Machinery / Equipment).
5. The P.S. to Secretary Health.

01/02/2023
(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH PM & I CELL HEALTH DEPARTMENT

GOVERNMENT OF SINDH
HEALTH DEPARTMENT

NOTIFICATION

No.SO(PM&I)2-1/2013/CRC): In supersession of this department's notification of even number dated: 09.01.2023 and in pursuance of Rule-31 of Sindh Public Procurement Rules 2010, Health Department, Govt. of Sindh is pleased to re-constitute Complaint Redressal Committee (CRC) comprising of following officers for scruitinizing the complaints of aggrieved bidders against tender invited by the Health Procurement Committee, Health Department under Frame Work Contract System and Health Institutions/Hospitals of Sindh Province :-

1)	Secretary Health, Govt. of Sindh	Chairman
2)	Representative of Accountant General Sindh, Karachi.	Member
3)	Independent professional from the relevant field.	Member

ToR's:

- To scrutinize the complaints received from the aggrieved bidders and decide the same strictly in accordance with SPP Rules 2010.

SECRETARY HEALTH

No.SO(PM&I)2-1/2013/CRC): Karachi, dated, the 10th July, 2024

Copy forwarded for information and necessary action to :

- The Chairman and all members of Health Procurement Committee.
- The Accountant General Sindh, Karachi.
- The Managing Director, Sindh Public Procurement Regulatory Authority Karachi.
- The Medical Superintendents (All).
- The District Health Officers (All).
- The Directors Health Institutions / Projects (All).
- The Chairman and all members of Complaint Redressal Committee.
- PS to Minister Health & Population Welfare Deptt., Govt. of Sindh, Karachi.
- PS to Secretary Health Govt. of Sindh, Karachi.
- PS to Additional Secretary (PM&I), Health Department, Govt. of Sindh, Karachi.

10/7/2024

(ZULFIQAR ALI DARS)
SECTION OFFICER (PM&I)

GOVERNMENT OF SINDH
PM & I CELL
HEALTH DEPARTMENT

ANNUAL PROCUREMENT PLAN FOR THE FINANCIAL YEAR 2026-27-2027-28

Name of the Procuring Agency: District Health Officer Thatta

Sr.No	Name of Procurement (Description) Estimated	Estimated Cost (Rs)	Procurement Method	Tentative / Time of Procurement				Remarks (if any)
				1st Qtr	2nd Qtr	3rd Qtr	4th Qtr	
01	PROCUREMENT OF MACHINERY/EQUIPMENT, FURNITURE/FIXTURE AND OTHER ITEMS FOR ESTABLISHMENT OF 02 DISPENSARIES NAMELY JAM BIJAR KHAN AT VILLAGE JAFFAR JOKHIO UC BHAMBORE TALUKA MIRPUR SAKRO AND VILLAGE KHUDA DINO SHAIKH UC BHAMBORE TALUKA MIRPUR SAKRO.	PKR. **2.291 M ADP-877**	Open Competitive Bidding	-	✓	✓	✓	E-bid through E-Pak Acquisition and Disposable System (EPADS)
02	PROCUREMENT OF MACHINERY, EQUIPMENT, FURNITURE/FIXTURE AND OTHER ITEMS FOR UPGRADATION OF DISTRICT COUNCIL DISPENSARY MEHAR CITY TO THE LEVEL OF BHU AT UC MEHAR TALUKA GHORABARI DISTRICT THATTA.	**PKR.9.851 M ADP-876**		-	✓	✓	✓	

DISTRICT HEALTH OFFICER
DISTRICT THATTA

District Health Officer
District Thatta

CamScanner

OFFICE OF THE DISTRICT HEALTH OFFICER
DISTRICT THATTA

HEALTH DEPARTMENT, GOVERNMENT OF SINDH

STANDARD BIDDING DOCUMENT

PROCUREMENT OF MEDICAL EQUIPMENT, MACHINERY, MEDICAL DEVICES, INSTRUMENTS, FURNITURE/FIXTURE AND OTHER ITEMS FOR FOLLOWING ADP SCHEMES

S#	ADP No#	NAME OF SCHEME
1	876	UP GRADATION OF DISTRICT COUNCIL DISPENSARY MEHAR CITY TO THE LEVEL OF BHU AT UC MEHAR, TALUKA GHORABARI
2	877	ESTABLISHMENT OF (02) DESPENSARIES NAMELY JAM BIJAR AT VILLAGE JAFAFAR JOKHIO, UC BHAMBHORE, TALUKA MIRPUR SAKRO AND VILLAGE BABU MALLAH, UC CHOWBANDI, TALUKA MIRPUR SAKRO

FINANCIAL YEAR 2025-26 & 2026-27

DUE ON _____ at 11:00 AM

@Fatima Engineer Auditorium, Jinnah Sindh Medical University, Karachi

(Single Stage – Two Envelope Procedure)

District Health Officer
Thatta

SECTIONS OF THE TENDER

TENDER TITLE
SECTION A SUPPLY, INSTALLATION OF MEDICAL MACHINERY/EQUIPMENTS **SECTION B** SUPPLY, OF HOSPITAL& GENERAL FURNITURE,FIXTURE ITEMS **SECTION C** MEDICAL INSTRUMENTS & DEVICES, IT EQUIPMENTS AND MISCELLANEOUS ITEMS

District Health Officer
Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

TABLEOFCONTENT

S.No.	DESCRIPTION	PAGE
1.	SalientFeature	
2.	Terms&Conditions	
3	DocumentsChecklist	
4.	EvaluationCriteria&BidderDetails	
5.	TechnicalSpecificationsofMedicalEquipment	
6	HumanResourceIncludingDetailofTechnicalTeam(FormA)	
7	MajorSpareParts(FormC)	
8	ComplianceSheet/TechnicalEvaluationSheet(FromD)	
9	FinancialEvaluationSheet(FormE)	
10	ManufacturerAuthorizationForm(FormF)	
11	ListofInstallation/ReferenceClient(FormG)	
12	PerformanceCertificate(FormH)	
13	BidLetterForm	
14	ContractAgreementForm	
15	BankGuaranteeForm	
16	IntegrityPact	
17	BidForm& PriceSchedule	
18	ScheduleofRequirements	

District Health Officer
Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

SALIENTFEATURESOFTHETENDER

1.	NameofWork&Address	PURCHASEOFMEDICALEQUIPMEN/MACHINERYFOR **Office of the District Health Officer District Thatta**
2.	Time&DateofissueofTenders	From the date of publishing to one day beforeopeningofTechnicalProposal,duringofficehours. Itshouldbe (AsmentionedinNotice invitingtender).
3.	PlaceofIssuanceofTender	**Office of thethe District Health Officer District Thatta**
4.	Dateofpre-bidmeeting	IncaserequestedbybiddersorinformedbyProcuringAgency. Mentionedpre-bidmeetingdateifany applicationreceived considered for discuss same mentioneddate
5.	MethodofopeningofTender	SingleStage–TwoEnvelopes;asperSPPRA-Rules2010 (Amendedup-to-date) on online EPAD system.
6.	Date&TimeofSubmissionofTenders	Aspublishedinnewspapers Itshouldbe (AsmentionedinNotice invitingtender).
7.	Date&TimeofOpeningofTenders/TechnicalProposals	Aspublishedinnewspapers/ on EPAD system Itshouldbe (AsmentionedinNotice invitingtender).
8.	DateofOpeningofFinancialProposals	AsinformedbytheChairmanofProcurementCommitteetoallparticipants
9.	Venueofsubmission&openingofTenders	ViceChancellor,JinnahSindhMedicalUniversity(JSMU),Karachi.
10.	ValidityofTenders	90daysasperSindhPublicProcurementRules,2010(Amendedtill date).
11.	AmountofEarnestMoney/BidSecurity	Fixed bid security on each quoted item amount inshape of Pay Order / Call Deposit from schedule Bankin favor of**District Health Officer District Thatta.**to be submitted along with Financial Proposal andalsoattachedphotocopywithoutshowingamountof thatpayorderelsetheofferwillberejected.
12.	SupplyofEquipment	Withinupto16weeksafterthePurchaseorderin caseofDDP.However,thisperiodisextendableaspermutualagreement/demand ofthesituation.
13.	InstallationPeriod	Itwillstartafterreceivingofequipmentatsite. It will start after satisfactory installation of equipmentatsite.
14.	MaintenancePeriod	It will start after the satisfactory installation of theequipment and issuance of Installation Certificatefrom the concerned official. This period will remainfunctionaltillthreeyearswithreplacementofparts.
15.	WarrantyPeriodwithreplacementofPartsandservices	As mentioned in technical specification of theequipment,otherwisethewarrantyperiodofparts withFreeofcostreplacementwillbeforthreeyears.
		Alongwiththreeyearsfreeservice&maintenancefromthedateof Installation.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Office Thatta

16.	(a)ContractAgreement StampPaperrequirementforAgr eement	TheContractorshallenter &executeaformal Agreementasperthe"Form"annexedwithsuchmodificationas may be necessary. - Rs.0.35%oftheContractValueorasprescribedbyGovernmentL aws.
	OpeningofL.C(LetterofCredit	Letter of Credit (L.C.) will be opened by ProcuringAgency as consignee and as per the prevailing rule ofsame. Performa invoice of the quoted amount on C&Fbasis of the country of origin along with insurancecover note if applicable will also be submitted by thecontractor(s). Bank charges in establishment of Letter of Credit(L.C.) fluctuation in currency if any and othermiscellaneous charges on logistics, transportationclearingfromseaport/airportwillalsobepaid/ bear bythecontractor(s).
17.	ReleaseofEarnestMoney/BidSecurity	Toun-successfulbidders,afterfinancialbidisopenedortechnical bidisrejected. Earnest Money will be released to successful Bidderafterpurchaseorderisreleasedbutafter5%Security Depositsubmission.
18.	SecurityDeposit	5% Security Deposit in the shape Pay Order/BankGuaranty will be submitted by Contractor withProcuringAgencywhichwillbereleasedasperitem20.Bank Guarantee of equal amount will also beacceptable. SecurityDeposit isdepositedas peritem–19.
19.	ReleaseofSecurityDeposit	Aftercompletionofinstallationandmaintenanceof DefectLiabil ity period.
20.	VariationinContractPrice	Novariation in priceshallbeallowedonanygroundincluding currency fluctuation/variation orwhatsoever.
21.	Discrepancy	IfthereisanydiscrepancybetweenSalientFeaturesof Tender /Bidding Documents and respectivecontents mentioned elsewhere, Salient Features /Terms&Conditions andAdditionalTerms& ConditionsofTenderwillgovern.
22.	Taxes	Alltaxes willbedeductedasperprevalentlawsofCountry.
23.	Approvedmakes	Asmentionedinitemsspecifications.
24.	SpecialNoteRegardingEquipment	Technical offers / Commercial offers failing todemonstratebelowdetailswouldberejected: a.Quotedsystemmustbeofadvanced&latestversion.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Office Thatta

		b. Tendermustcovercompleteequipment. c. Tendermustcovercompleterangeofdispo sables/kits d. Tendermustprovidealltechnicaldetailsuptothesatisfacti onoftheend user. Itemsshouldbequalityapprovedfromtheconcernedinternatio nalbodyoftherespectiveindustry.
25.	(a)InspectionofImportedequipmentman ufacturing sitebytheclient.	Aftertheaward ofPurchaseOrder,Suppliershalltaketwo persons to the manufacturing site and bear allexpenses of visit in case the cost any equipment ofSupplyOrder is 50million or above.
	(b)Training	Supplierwillprovide on-sitesuccessfultrainingtoallthe personnel working on/operating the saidEquipment/Machineaslongastheneedprevail.
26.	Maintenance.	MaintenancecostforallitemsforTwoyearsfromthedate of successful Installation shall be undertaken bythe Contractor (Maintenance includes all Parts &Labor,etc.withsufficientstaff,duringmaintenance period).
27.	DefaultinPreventiveMaintenance,Brea kdownandEmergencyCalls.	• In case of default by the Contractor with respect tomaintenance,breakdownandemergencycalls,thesame will be carried out within 24 hours by theProcuring Agency and the cost so incurred will bepaid fromtheRetention Money. • Moreoveranadditional10%oftheamountspentwould be charged from the concerned contractorbeingdefaulter.
28.	CosttobequotedinBOQ.	• The Contractors shall quote price in DDP only, ofthe Equipment's including custom clearance etc.and delivering of Equipment at site withInstallation&Commissioning cost. • It should also include cost during maintenanceperiodincludingthepartsandmaintenanc estaff.
29.	TaxExemption.	In case of purchase on C&F basis, the requiredcertificatewillbe issuedbythisofficethattheimporthas been made for this hospital, so as to avail thefacility of exemption of duties/taxes, as perGovernment Rules/Policy according to the SindhPublic Procurement Rules, 2010 (Amended till date).Clearancefromthe customandotherchargeswillbe theresponsibilityofbidder.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar. _____

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI. _____

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop. _____

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar. _____

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi. _____

District Health Officer Thatta

30.	Criteriaforbidevaluation.	MeritPointEvaluation(**Most Advantageous Bid**)Theitemsrankedhighestinmeritpoints(obtainedthrough andbasedontechnicalandfinancialevaluation)shallbe consideredasmostadvantageousbid.
31.	AwardofContract	TheProcuringagencywillawardthecontracttothesuccessfulB idderwhosebidhasbeendeterminedtobesubstantially responsive and has been determined to be the**Most Advantageous Bid,** provided further that the Bidder isdeterminedtobequalifiedtoperformthecontract satisfactorily.

District Health Officer
Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

TERMS&CONDITIONSOFTENDER

1. TheContractorsresponsiblefor**Supply,Installation,TestingandCommissioningof MedicalEquipment, Machinery items, Instruments &Furniture & Fixture** while Installing Equipment at no extra-cost. So, the Contactors should take into account allsuchexpenditureswhilequoting thebid.
2. **In Technical Bid**, the bidder must provide original data sheet, technical brochure, and all other relevantdocuments, along with copy of Pay Order / Bank Draft / Bank Guarantee without showing the amountfigure otherwise, the bid will be ignored/rejected. The technical evaluation will be done on the basis ofcriteriagivenintenderdocument.
3. The vendor has to fill up **compliance sheet** with YES/NO or put value as desired in the column againstprovided specification. These specifications must be highlighted in the brochure attached; else the offerwillberejected.
4. **In Financial Bid**, the vendor should mention financial offer along with scanned copy of Pay Order / Bank Draft / BankGuaranteeFixed bid security on each quoted item amount infavorof,District Health Officer District Thatta**Bid Security (in Original) must be submitted in sealed envelope to the procuring Agency before opening of tender.**
5. The Supply order shall be placed on DDP basis the bidder should supply the items within time as specifiedin Tender Document i.e. 16 weeks. This period is extendable as per mutual agreement/demand ofsituation.
6. The disclosure of firm's price at the time of opening of Technical bid will result in the rejection of thebid.
7. Notenderwillbeentertainedwithoutearnestmoney.Theearnestmoneywillbeforfeitedto Government Treasury, in case of non-submission of security money within **seven (7) days** of receipt oftheofferletter.
8. OnlyManufacturer,itsSoleAgentorAuthorizedDistributorscanquoteDevice,Equipment/Machinery.Nosub-distributoris allowedtoquote.
9. **For Single Stage Two Envelope Sections of the tender**, first scrutiny of Technical Bids will be performedby the Procurement Committee. Financial Bids of only those firms will be opened who are qualifiedtechnically.
10. CertificatefromtheSoleAgent/ManufacturerhasfullyequippedWorkshopTrainedPersonnel/Engineers and Equipment should beprovided in the technical bid, and it can bephysically verified bythe administration of**District Health Officer District Thatta.**The warranty period will commencefromthedateofinstallationoftheEquipmentat**Office of the District Health Officer District Thatta.**
11. Firm must provide complete detailsoftheir financial standing,listingofsimilaritemssupplied inPakistan,availablelocalstockofitemsand detailsofofficewithnumberof personals.
12. The vendor has to quote only one rate for each item as per tender specifications. Hand written tendersoranyoverwriting,cutting,shouldbesigned.
13. Vendorsshouldquotetheitemstheyintendtosupply.
14. The bidder shall submit with the bid, an **Undertaking** on Non-Judicial stamp paper of Rs.50/- dulyattested by Oath Commissioner, stating that the medical equipment will be supplied brand new, freefromdefectsaspertheapprovedspecification,alsoprovidetwoyearswarrantyforserviceandmaintenanc

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

e and will bind himself for availability of spares for a period of **10 years** from the date ofsupply.Incaseofwindingupofhisbusiness,thesupplierwillbeboundtomakeanalternatearrangement of spares and maintenance of medical equipment. In case of default, depreciated cost ofmedicalequipmentwillberecoverablefromvendor.

15. If it has been found that the information submitted by the bidder regarding his qualificationandprofessional, technical, financial, legal or managerial competence as supplier / contractor is / was falseandmateriallyinaccurateor incompleteatanystage,thenthebidder willbedisqualified.
16. Suppliers / Contractors will be responsible for supply at consignee end, free of cost and also responsiblefor freeinstallationwork, if required.
17. Equipment will be handed over to the Store Keeper will be handed over to Store Keeper **Office of the District Health Officer District Thatta** (Priceescalationwill notbeallowed.)
18. If vendor doesn't supply medical equipment in accordance with the supply order or fails to supplyaccording to termsand conditionsorincase of any other default,the PerformanceSecurity andSecurity Deposit would be forfeited and further action as per SPPRA rule would be taken against suchvendor.
19. If the supplier fails to give supply and install within the stipulated period, Liquidated charges will beimposed.
20. Thedecisiononcetakenwillbefinaland willnotbechallengedin anyCourtofLaw.
21. AcopyofIncomeTaxandGSTRegistrationCertificateshouldbeattachedalongwithTechnicalBid.
22. The bidding firm must be registered with PNRA department for minimum past three years (for Radiologyitems).
23. ThebiddershallfurnishcopyofATL (ActiveTaxPayerListcertificate).
24. CopyofvalidProfessionalTax(Excise&Taxation)Certificate,IncomeTaxCertificateandGSTRegistrationCertificateshouldbeattachedwiththetechnicalbid.
25. Thebiddershallfurnishthreeyearsincome TaxReturnCertificate.
26. Conditional Tender(s) against the Government rules and non-compliance of instructions mentioned inTenderForm&Handwrittentender(s)will notbeentertained.
27. The vendor should submit a **Data Sheet** according to the Tender Specifications Form as mentioned inthe tender and highlight the tender specifications in their Technical Data Sheet. Additional Features canbementionedseparately.
28. In case the informationprovidedbybidderinData Sheet asper**Clause No.30**is incomplete orincorrect,thequoteditemwill beliablefor rejection.
29. SuppliermusthavetoprovideManualbookwithcircuitdiagramwhereapplicable.
30. Suppliermusthavetoarrangetrainingonsiteforconcernedtechnicalstafffortheusupplyofequipment
31. ProcurementCommittee,,**Office of theDistrict Health Officer District Thatta**
32. therighttorejectoracceptany /alltender(s)asperSPPR-2010(Amended2019).

District Health Officer Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

33. Thesystemwillhavetheminimumuptime95%incaseofdowntimebeyondthisperiodthefollowingpenaltywill beapplied:

a) 95%to100%Nopenalty.

b) 90%to95%thewarrantyperiodwillbeextendedbythenumberofdayssystemindown.

c) 85%to90%warrantyperiodwillbeextended 1.5timethedowntimeperiod

d) 80%to85%warrantyperiodwillbeextendedtwotimeforthedowntimeperiodandsupplierissupposedtogi vejustificationofdelay.

e) Uptimelessthan80%hospitalwillcomplaintothePrincipaltotakeactionagainstthesupplierandwarrant\' will bethethreetimeforthedowntimeperiod.

f) However,PPM(Periodic PreventiveMaintenance)willnotbeconsideredinthedowntimeandPPMofthesystemwillbedoneas permanufacturerrecommendation.

34. PaymentofDDPbasis,thevendorshallsubmitBill/Invoiceforthepurchaseordervalueafterdelivery,installatio nandissuanceofsatisfactoryreportbytheenduser.

35. Bids shall remain valid for **90 days** after the date of bid opening and same may be extended in terms ofRule38(2)(3)(4)ofSPPRARules.

36. Bidshouldbeinclusiveof allGovernmenttaxes(ifapplicable)andthesame willbepaidbytheContractor except withholding tax and 0.35% Stamp Duty which will be deducted at source in office oftheAccountantGeneralSindhKarachi.

37. The bid security will be forfeited to the Government, if the bidder withdraws his bid after opening andbefore the expiry of the bid validity period or fails to sign the contract in stipulated time if the bid isaccepted.

38. The Procuring Agency may reject all or any bid at any time prior to the acceptance of a bid or proposals,subject to the relevant provision of SPP Rules, 2010 (Amended 2019 up-to-date). Payment will be madewithin fourweeksafterreceiptofbiii / invoicedulyfills inall respects.

39. Pricesquotedshallremainvalidupto 90Days.

40. RegistrationfromSindhRevenueBoardisnotrequiredinprocurementofGoods.

I/Weagreeto abovementionedterms&conditions:

NameofContractor_____Signature_____

(CNICNO_____(Copy mustbeattached).)

FullAddress _____

_____

RubberStamp_____

District Health Officer Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

SECTIONA

Supply,Installation,TestingandCommissioningofMedicalEquipmentandMachineryItems

DOCUMENTCHECKLIST

Please review the following list of all documents to be enclosed with the Technical Proposal. These arethe "MANDATORY DOCUMENTS" non-submission of any one of the mandatory documents will lead todisqualificationandnofurtherassessmentoftenderwillnotbedonehence"TECHNICALLYREJECTED"

S.NO	DOCUMENTDESCRIPTION	YES/NO	PAGE#
1	TenderPurchaseReceipt(Original)Copy of Pay Order		
2	BidSecurity(Copywithvalue hiddeninTechnicalProposal;Originalin Financial)		
3	BiddingDocuments(Dulyfilled,Signed&StampedbyBidder)		
4	TechnicalProposalonBidder'sLetterhead		
5	Compliancesheetonletterhead(Mandatory)		
6	Undertaking(aspersampleform)		
7	Certificate(aspersampleform)		
8	IncomeTax&GSTRegistrationCertificate		
9	ProfessionalTaxCertificate(Sindh)		
10	SECPIncorporationCertificate(whereapplicable)		
11	PNRARegistrationCertificate(whereapplicable)		
12	PECRegistrationofFirm&QualifiedEngineer'sCertificate(whereapplicable)(MandatoryforServiceworks)		
13	DRAP Establishment licenseshowingImporter/Manufacturerofmedicaldevices.		
14	DRAP Product Establishment /Registration		
15	CompanyProfile		
16	BankAccountMaintenanceCertificate		
17	IncomeTaxReturns(last3years)		
18	AuditedFinancialStatement(last3year)		
19	WorkshopforAfterSalesServices(Mandatory)		
20	TechnicalTeamDetails(Workshopdetails&TechnicalStaffBiodata withTrainingCertificates)		
21	Manufacturer'sAuthorizationforthesoledistributioninPakistan (AttestedbytheEmbassy)		
22	Catalogue/Brochures(original)		
23	Product TechnicalDataSheet		
24	ISO9001ofbidder(whereapplicable)**(website linkmustbeprovidedforverification**		
25	ISO13485ofManufacturer(whereapplicable)**(websitelinkmustbeprovidedfor verification)**		
	CE/FDA510(K)/ JIS/MHLW(whereapplicable)**(websitelinkmustbeprovidedfor verification)**		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Thatta

Note:

- Mandatorydocumentsarementioned.
- Allpagesofthebid,exceptforun-amendedprintedliterature,shallbeinitialedbythe Bidder.

BIDDER'SDETAILSFORNOTICEPURPOSES

BidderName:	
Address:	
TelNo.	
FaxNo.	
Contact PersonName:	
MobileNo.	
EmailAddress:	

District Health Officer, Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

SECTIONA

Supply,Installation,TestingandCommissioningofMedicalEquipmentandMachineryItems

Note:

EVALUATIONCRITERIA

Financialproposalmustbesubmittedoncompanyletterheaddulysignedandstamped.Bidderisrequiredtotypetheirofferinfigureandaswell asinwordsofthetotalamount;elsetheofferwouldberejected.

EVALUATION/QUALIFICATION

1. Theprocuringagencyreservestheright toevaluateandcomparethebidsonitemizedbasisORonbasisofagroupofsimilarnaturegoodsORgoodscompatible witheachother.
2. Thefollowingmeritpointsystemforweighingevaluationfactors/criteriawillbeappliedfortechnicalproposals.
3. Biddersachievingminimum70%markswillbeconsideredasqualifiedbesidecompliantofallmandatoryclauses.Documentaryevidencemustbeattachedinsupportofeachparameter.
4. Any Bid not meeting the mandatory requirements of evaluation criteria will be disqualified /rejected straightawayandwill notbeconsideredforfurtherevaluation.
5. Providewebsitelinkofproductcertificationforonlineverification.

A. PRODUCTEVALUATION

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
1	**ConformitytothePurchaser's Specifications(Mandatory)**	**30**
1.1	Fullycompliantwiththerequiredspecifications	30
1.2	Compliantwithminordeviation(upto5%subjecttomainfunctionisnotaffected)	25
1.3	Compliantwithminordeviation(upto10%subjecttomainfunctionisnotaffected)	20
1.4	Non-complianttorequiredspecifications	Disqualify
2	**TrainedProductSpecialist**	**10**
2.1	TwomarksforeachForeignTechnicalGraduateEngineerwithPECRegistrationinSindhforthequotedproduct(Factory/OEMlevelservicetraining).	06
2.2	TwomarksforeachForeignTrained ScienceGraduateinSindhforthe quotedproduct (Factory/OEMlevelapplicationtraining).	04
3	**ProductCertification**	**10**
	• USAFood&DrugAdministration(USA-FDA) • EuropeanCommunity(CE)MDD • JapanIndustrialStandard(JIS) /Japan QualityAssuranceOrganization(JQAO)/CertificateissuedbyMinistryofHealth,Laborand WelfareGovt.ofJapan	
3.1	AllThreeoftheabove	10
3.2	Any two	7
3.3	OnlyOne	3
	TOTALMARKSPRODUCTEVALUATION(A)	**50**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer

B. BIDDEREVALUATION

S#	PARAMETERS/SUB-PARAMETERS	TotalMarks
4.	**Manufacturer'sAuthorization(Mandatory)**	**10**
4.1	BidderisOEMdirectrepresentativeortheSoleDistributorforlast10yearsormore. (Authorizationcertificateattestedbyembassymustbeprovided)	10
4.2	BidderisOEMdirectrepresentativeortheSoleDistributorformorethan 05yearsbut lessthan10years.(Authorizationcertificateattestedbyembassymustbeprovided)	8
4.3	BidderisOEMdirectrepresentativeortheSoleDistributor for05yearsorless (Authorizationcertificateattestedbyembassymustbeprovided)	5
4.4	Nonsubmissionofvalidmanufacturer'sauthorization	Disqualify
5.	**HumanResource**	**06**
5.1	Diploma of Associate Engineer (DAE) in electrical / electronic / biomedical /mechatronics/mechanical/industrial.DAEcertificatemustbesubmitted.(***0.5mark foreachcertificate***)	2
5.2	GraduateEngineerwithPECRegistrationinbiomedical,electronics,andmechatronics, mechanical,industrial.PECregistrationcardofthe engineermustbesubmitted.(***1markfor eachEngineer***)	4
6.	**Networking**	**04**
6.1	NetworkingsetupacrossPakistan(1markforeachsetup)(verifiable)(proofof registeredofficeandworkshopmustbeprovided)	04
7.	**PastExperience/Performance**	**10**
7.1	Bidder's prior experience for supplying the **same nature of product.** Performancecertificate of at least 05 public hospitals having minimum 5 years' Experience alongwithInstallationReportandSupplyorder/PurchaseOrderofthefirm inlast5years,onletterhead,signedand stampedby theHead of institutionofpublicsector. ***Onemarkforeachaftersalesatisfactoryperformancecertificate.***	05
7.2	Bidder's prior experience for supplying the **same nature of product.** Performancecertificate of at least 05 reputableprivate hospitals having minimum 5 years'Experience alongwithInstallationReportandSupply order/ Purchase Order of thefirminlast5years,onletterhead,signedandstampedbytheHeadofinstitutionof privatesector.***Onemarkforeachaftersalesatisfactoryperformancecertificate.***	05
8.	**Turnovertotalduringlastthree(03)financialyears(AuditedStatementsofAccounts andIncomeTaxReturnFormsmustbeattachedassupportingdocuments)**	**10**
8.5	Turnoverlast3 yearsequalorabovePKR1200 million	10
8.4	Turnoverlast3 yearsequalorabovePKR1000million	08
8.3	Turnoverlast3 yearsequalorabovePKR 800million	06
8.2	Turnoverlast3yearsequalorabovePKR 600 million	04
8.1	Turnoverlast3yearsbelowPKR600million	00
9.	**Bonuspoints**	**10**
9.1	FreeofCostExtendedWarranty(inaccordancewiththestandardwarrantyperiod) atleastfortwoyearincludingparts &services &labor. (Bidderofferinggreaterextendedwarrantyperiodwillgetthefullmarks)	10
9.2	FreeofCostExtended Warranty(inaccordancewiththestandardwarrantyperiod) at leastforoneyearincludingparts&services&labor.	05
	TOTALMARKSBIDDEREVALUATION(B)	50
	GRANDTOTAL(A+B)	100

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Thatta

BIDDEREVALUATIONCRITERIA

Bid found to be responsive with maximum accumulative points (Technical + Financial) shall be accepted as the **mostadvantageousbid**asperSPPR,2010(Amended2021).

Bidswillbeevaluated onthebasisoffollowingevaluationcriteria.

A. TECHNICALEVALUATION

i) Technicalevaluationhas70%weightage

ii) BidswillbeevaluatedonaBidder'scapacityandcapabilitytoundertakesuppliesandonquoteditemquality,performanceandconformity totherequiredspecifications.

iii) Technical capacity of a bidder includes availability of workshop, tools, technical staff, number ofinstallation and experiencein the province of Sindh forprompt servicesto KarachiDivision.Past performance of the bidder at**District Health Officer District Thatta**will also bereviewed.

iv) A bidder must achieve a minimum of 70 points out of total 100 points in technical evaluation, tobeconsideredforfinancialevaluation.

B. FINANCIALEVALUATION

i) Financialevaluationhas30%weightage.

ii) Only financial proposal of bidders with responsive technical proposal will be opened. Financialproposal of bidders who do not meet required mandatory documents and minimum 70 pointsoutoftotal100pointsintechnicalevaluationwillbereturnedunopened.

iii) Bidderwhohasofferedthelowestvaluefortheitemwillget30points.

iv) All remaining bidders after the lowest shall get points on their financial proposal, calculated asperthefollowing formula.

$$\frac{\text{Valueoflowestbidder} \times 30}{\text{Valueofhigher bidder [2nd, 3rd, 4th,etc.]}}$$

C. FINALBIDEVALUATION

i) Pointsobtainedbybiddersintechnicalandfinancialevaluationshallbeproratedandfinalizedas follows.

Evaluation	Weightage
Technical	70%
Financial	30%
Total	**100%**

ii) Bidofabidderwhoisfoundtechnicallyresponsivewithhighestaccumulatedpointswillbeconsideredas themostadvantageousbid.

District Health Officer
Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

SECTIONA

Supply,Installation,TestingandCommissioningofMedicalEquipmentandMachineryItems

S. No	Name of items	QTY	FIXED BID SECURITY
1.	**Autoclave** **Specification:** The outer surfaces and the chamber of the device must becompletelymade ofstainlesssteel Sterilizationbasketmadeofstainlesssteelmaterialinsidethedevice The volume used in the device should be at least 50-60liters The device should be able to perform fully automatic, pre-vacuum,steam filling,sterilization,steam dischargeandfinalvacuumoperationsautomatically. Coversealonthedevicemustbeprovidedwithsiliconebased gasket Shouldhavealockablecoversystemthatdoesnotopenwhenthere ispressure inside Double-walled device, steam generator should be placedbetween walls The water required for reinforcing the steam of the steamgeneratorislocated inatankandthewaterlevelinthistankshould be observable Thevacuumpumpused inthedevicemustbeofliquidringtype Freshairintakeintothecellduringthedryingprocess mustbeprovidedviatheHEPAfilter The pressure gauges of the sterilization chamber must beseparatelyandonthe panel Device should have PLC control system, sterilizationtemperature between 110 ° C and 140 ° C at 1 ° C intervals,sterilization time at intervals of 1 minute up to 99 minutesand final vacuumrepeatnumberbetween2-9 On the control panel of the device control 'must bewarningsystemforopen,overpressure,heaterfailureandlowwaterlevel acid' In case the electricity is cut off again during thesterilizationphase:Ifthetemperatureofthecabinis within±1 ° C limits, it should continue its operation and if the alarmexceeds the limits, it should warn the power failure visuallyand audibly. Thedevice musthavebothelectrical(heater-stopping),mechanical (overpressure) safety system and safety valveagainstpressure **Country of Origin**: UK/USA/EUROPE/OR EQUAVLENT. **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years startingfromdateofinstallationofequipmentto theconsigneeend.	02	1,500
2.	**Binocular Microscope 50000-200000** **Specification:** Standardmicroscopewithinfinitycolorcorrectedopticsbinocular head,adjustableinterpapillarydistance Tubeheadwithan inclinationof30DegreeormoreRotating quadruple/quintuplenose piece 2-layerwithmechanicalslidingstage,XYmovingrangeCondensercarrier,vertically/horizontaladjustableBrightfieldcondenser, N.A.of 1.25 Aperture iris diaphragmLEDillumination 220V50 Hz, AC	01	15,000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

MEDICAL SUPERINTENDENT PEOPLES MEDICAL COLLAGE HOSPITAL NAWABSHAH SBA

MEDICAL SUPERINTENDENT GMMMC HOSPITAL SUKKUR

District Health Officer Thatta

	UserAdjustableSettings: Coaxialcoarseandfinefocusingsystem,withfocusingstopmechanismVariable intensity controlsystem Eyepiecelenses:- 10x wide field, focusable,Objective lenses (Achromat):-4xScanninglens 10X 20X 40x 100xoilimmersion **CountryofOrigin**:UK,USA,Europe, JapanorEquivalent. **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty**:Warrantyperiodshallbeofthreeyearsstartingfromdateofsatisfactory installationof equipment		
3.	**Centrifuge (Bench Top)** **Specification:** Maximum capacity: 12 x 15 mL Maximum speed : 300-5000 rpm, adjustable in increments of 100 rpm Time setting: 1-99 minutes, in increments of 1 minute: HOLD Temperature range: 2°C to 40°C Control system: Microprocessor User interface: Large, brightly-lit LED display Noise level: <55 dBA Power Supply (VAC/ Hz): 220-240 VAC/ 50 Hz **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.	01	750
4.	**Centrifuge Machine** **Specification:** Maximum capacity: 12 x 15 mL Maximum speed : 300-5000 rpm, adjustable in increments of 100 rpm Time setting: 1-99 minutes, in increments of 1 minute; HOLD Temperature range: 2°C to 40°C Control system: Microprocessor User interface: Large, brightly-lit LED display Noise level: <55 dBA Power Supply (VAC/ Hz): 220-240 VAC/ 50 Hz **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of Satisfactory installation of equipment to the consignee end.**	01	7,500

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Karachi.

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer

5.	**Incubator with pathology** **Specification:** Temperature range 10 to 100° C Control Accuracy ± 0.7° C Volume 60 -100 liters No of Racks Two Temperature Control PID Display LED Operating Voltage 220V **Ventilation & Control** Forced air circulation Parameters adjustable temperature (Celsius or Fahrenheit), Programmed stored in case of power failure **Over Temperature Protection** with audible and visual alarm Over temperature monitor **Texture** Fully insulated stainless steel door with 2-point locking (compression door lock) Inner glass door Easy-to-clean interior, **Temperature** Min. 10 °C above ambient up to +80 °C **Accessories**: Complete with standard accessories, software, Operating manual, Services Manual, error code book, part list and software if any **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warrantyperiod shallbeofthreeyearsstartingfromdateofinstallation of equipmentto the consignee end.**	01	9,000
6.	**Infant Warmer** **Specification:** • **Type:** Infant radiant warmer with modular, user-friendly design for optimal thermal support of newborns. • Designed for **neonatal intensive care, delivery rooms, emergency departments and operating theatres.** • **Application:** Thermoregulation, pre-warming, manual heating and servo-controlled baby mode. *Heating & Temperature Control* • **Heating Source:** Radiant heater with ceramic resistances in warmer head for **even heat distribution.** • **Operating Modes:** Baby (servo) mode, Pre-warm mode, Manual mode. • **Skin Temperature Set Range:34 – 38 °C** in **0.1 °C increments.** • **Temperature Measurement Accuracy:** ± 0.1 °C. • **Heater Output:** 0 – 100 % in 10 % increments. • **Skin Temperature Display Range:18 °C – 42 °C.** • Intelligent alarm system for prolonged warm-up, sensor failures and safety limits. *Mechanical & Physical* • **Dimensions (approx):** o Width: **90 cm** o Depth: **110 cm** o Height (standard): **177 cm** o Height (adjustable, optional): **177 – 197 cm** o Mattress Size: **70 × 55 cm** o Maximum Baby Load: **10 kg** o Standard Mattress Height: **93 cm**	02	3,300

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Thatta

	○ Adjustable Mattress Height (optional): **93 – 113 cm.** • **Bed Tilting:** ± 12°. • **Wheels:** Castor wheels with brakes for mobility. (manual) *Illumination & Display* • **Examination Light:LED lamps** (typically 2 × 5 W). • Clear digital display for temperature, settings and alarm status. *Control Panel & Features* • **Control Panel:** Digital with mode and setting buttons. • **SmartSilence:** Hands-free alarm silence function to maintain sterile hands. • **SmartTilt:** Electronic Trendelenburg and reverse Trendelenburg positioning. • **SmartMemory System:** Stores two memory positions for bed tilting angles. **Country of Origin:** UK, USA, Europe, Japan, (Any One Certificates of CE, FDA, MHLW) **Should be:** ISO-13485 (Mandatory) **Warranty:** Warranty period shall be of three years starting from date of satisfactory installation of equipment		
7.	**Chemistry Analyzer** **Specification:** • Membrane keyboard/Touchscreen • Optional external keyboard • High contrast graphical LCD display OPTICAL • Photometric range: -0.1 to 2.3 Absorbance or better • Wavelengths: automatic by 12 positions filter wheel • 7 Standard interference filters: 340, 405, 505, 546, 578, 620 and 660 nm or better • 5 positions for optional filters • Detector: Photo diode (320-1000 nm) or better Light source: quartz-halogen lamp or better. FLOWCELL • Metal, with quartz windows • Measuring volume 30 µL or better PRINTER • Internal matrix printer • Normal paper • External printer port available MULTIPLETESTING • Up to 9replicates • Means, SD and CV% POWERREQUIREMENTS • 100-240 VAC nominal, 50/60 Hz Battery back-up to retain data/ Compatible External imported online UPS • The bidder will provide kits for 100 patients along with the machine free of cost at the time of installation **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warrantyperiod shallbeofthreeyearsstartingfromdateofinstallation of equipmentto the consignee end**	01	6,000
8.	**Microscope** **Specification:** Standard microscope with infinity color corrected optics Binocular head, adjustable inter papillary distance Tube head with an inclination of 30 Degree or more Rotating quadruple/quintuple nose piece 2-layer with mechanical sliding stage, XY moving range Condenser carrier,	01	7,500

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

	vertically/horizontal adjustable Bright field condenser, N.A. of 1.25 Aperture iris diaphragm LED illumination 220V 50 Hz, AC User Adjustable Settings: Coaxial coarse and fine focusing system, with focusing stop mechanism Variable intensity control system Eyepiece lenses: - 10x wide field, focusable, Objective lenses (Achromat): - 4xScanning lens 10X 20X 40x 100x oil immersion **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years startingfromdateofinstallationofequipmentto theconsigneeend..		
9.	**Suction Machine** **Specification:** Should have ability to be used in the emergency dept., intensive care, operating theatre. The pump should have accurate vacuum setting. Should have piston/cylinder system for a vibration free and noiseless operation less than 50 db. The vacuum range should start from -95 kPa / 713 mmHg or better. The flow rate should be adjustable from 40 to 60 liter/min. Power Requirement: Line Voltages 100-240 VAC Line Frequency 50/60Hz **Should be Supplied with:** Two reusable suction jars of 4 liters Mobile Trolley Original with two lockable castors Antibacterial Filter Clamp Holder Disposable Patient Tubing Power Cord Operating Manual **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years startingfromdateofinstallationofequipmentto theconsigneeend.	04	3,600
10.	**Ultrasound Machine with printer &Trolley** **Specification:** Digital Ultrasound Scanner with digital beam former. The System should be capable to handle multi frequency probes. **Probes:** **Multi frequency Convex Probe with center frequency between 3 to 5 MHz** Measurement Modes: B, M and combination thereof and pw Doppler. M. Mode sweep: 4 speed or more. Have must **PW Doppler** in standard Gray scale: 256	01	110,760

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer, Malir Katta

	Sensitivity time gain: 6-8 steps Depth: 24 cm or more Two Active Probe Connector or more Focusing system: 3 steps and dynamic System dynamic range: upto150dB Speckle Reduction Imaging Frame rate: 250 frames/ sec or more Keyboard: Alpha numeric with track ball / Touch pad Tissue Harmonics: Tissue Harmonic imaging Cine memory of 256 frames minimum Post processing: Image inversion, edge/echo enhancement correlation / persistence/Dynamic range/Gamma Curve. Image magnification 4x or more in real time. Display Monitor: 12" LCD / TFT screen Operating Voltage: 220V, 50 Hz **Built-in battery backup** of minimum 2 hours during scanning. Gel: 05 liters in tubes Trolley compatible with the machine for smooth operations and transportation (Local-subject to approval of sample) 20 thermal paper Rolls Thermal Printer 256-Gray scale (Sony, Mitsubishi or equivalent) **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years startingfromdateofinstallationofequipmentto theconsigneeend.		
11.	**Urine Analyzer** **Specification:** A urine analyzer is a device used in the clinical setting to perform automatic urine testing. High tech Urine Analyzer also performs automated microscopy and automatic image evaluation. Urine analyzer based on photometer principal with 4 differential wavelengths Unit analyze check Specific Gravity, Leucocytes, Nitrite, pH, Protein, Glucose, Ketones, Urobilinogen, Bilirubin, Blood, Creatinine, Albumin and Micro albumin to Creatinine ratio in urine Microscopy Analysis: RBC, WBC, Crystals, Yeast Cells, Bacteria, Hyaline Casts, Sperm, Mucus, Pathological casts, non-squamous epithelial cells, Squamous epithelial cells Automatic calibration supported for strip and reference cuvette Automated Microscopy and automatic image evaluation Possibility to connect external Key board Connectivity: Connection to external keyboard and bar code reader 220V 50 Hz, AC User Adjustable Settings: Sample Type: Urine and Dialysate Wave length from 490-650nm or better Throughput not less than 200 with strip reading and 100 for microscopy Displayed Parameters: Digital touch screen display Data Storage Capacity: 200 measurements Accessories:	**01**	**24,000**

MEDICAL SUPERINTENDENT
PEOPLES MEDICAL COLLEGE HOSPITAL
NAWABSHAH SBA...

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer

	Complete with standard and operation accessories; With Built-in Thermal Printer or External Laser Printer Compatible Imported Online Sine wave UPS with Battery backup for 30 minutes **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty:** Warranty period shall be of three years startingfromdateofinstallationofequipmentto theconsigneeend..		
12.	**Wash Bath** **Specification:** KeyFeatures: - Electronictimeandtemperaturecontrol,upto80 °C - Auto start temperature controlled / temperature pre-selection Specification: - TankVolumeCapacity= 18 Literormore - Heater output Performance= 1000Wattorbetter - ConstructiveMaterial ofTank: Stainless Steel AISI304 - PowerRequirement: Line voltages:220-240VAC - Line Frequency: 50/60 Hz Countryoforigin:USA/EUROPE/JAPAN OR EQUIVALENT **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. Warranty:03 yearsfromthedateofsatisfactoryinstallation.	01	4,500

MEDICAL SUPERINTENDENT
GMMMC HOSPITAL SUKKUR

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer
Thatta

SECTIONB
SupplyofHospital&GeneralFurnitureItems

Technical Evaluation Criteria for Section "B"MaximumAllocableMarksScore 70%toQualify

CheckList(MandatoryDocuments)

Ifanybidderfailedtosubmitanyofthefollowingdocuments,thesameshallbedisqualified.

Sr.No.	DOCUMENTDESCRIPTION	YES/NO	PAGE#
1	TenderPurchaseReceipt(Original) Copy of Pay Order		
2	BidSecurity(2%)CopywithvaluehiddeninTechnicalProposal;OriginalinFinancial)		
3	TechnicalProposalonBidder'sLetterhead		
4	Compliancesheetonletterhead		
5	Undertaking(aspersampleform)		
6	IncomeTax&GSTRegistrationCertificate		
7	ValidProfessionalTaxCertificate		
8	FBRIncomeTaxReturns(last3years)		
9	AuditedFinancialStatement(last3year)		
10	Factory/WorkshopDetails(dulyequipped)		
11	ValidAuthorizationLetterfromManufacturerinreferencetothisparticulartender.		
12	Catalogue/Brochures /Drawing		
13	ISO9001ofBidders/Manufacturer(whereapplicable)**(websitelinkmustbeprovidedfor verification)**		

Note:

- Allpagesofthebid,exceptforun-amendedprintedliterature,shallbeinitialedbythe Bidder.
- Documentsmustbesubmittedinbinding,withfilledTableofContent,PageNumberingandtaggingwherev errequired.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer
Thatta

EVALUATION/QUALIFICATION

1.1 Theprocuringagencyreservestheright toevaluateandcomparethebidsonitemizedbasisORonthebasisofagroupofsimilarnaturegoodsORgoodscompatiblewitheachother.

Thefollowingmeritpointsystemforweighingevaluationfactors/criteriawillbeappliedfortechnicalproposals.

1.2 Biddersachievingminimum**70%marks**willbeconsideredonlyforfurtherprocess.Documentaryevidencemustbeattachedinsupportofeachparameter.

1.3 AnyBidnotmeetingthemandatoryrequirementsofevaluationcriteriawillbedisqualified/rejectedstraightawayandwillnotbeconsideredforfurtherevaluation.

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
1	**ConformitytothePurchaser'sSpecifications/SampleExamination**	**50**
1.1	Fullycompliancewiththerequiredspecifications/Sample	50
1.2	Compliantwithminordeviation(upto10%subjecttomainfunctionisnotaffected)	40
1.3	Non-compliancefromthestatementofrequiredspecifications	**Disqualify**
2	**LastthreeYears'ExperienceoffivePublicSectorTertiaryCareHospitals(2Marksofeach)**	**10**
3	**LastthreeYears'ExperienceoffivePrivateSectorTertiaryCareHospitals(2Marksofeach)**	**10**
4	**Average Annual Turnover during last three (03) financial yearsDulyverifiedfrom FBRTaxReturns**	**10**
4.1	Morethan500(M)	10
4.2	Morethan 300(M)to400(M)	08
4.3	Morethan 200(M)to300(M)	06
4.4	Morethan 100(M)to200(M)	04
5	**Manufacturer'sAuthorization**	**10**
5.1	ItselfManufacturer (AttachedvalidISO9001/13485) (AttachedFBRstatusshowingManufacturerInNTNverification)	10
5.2	ValidAuthorizationofManufacturerofgoods (Attachedvalidauthorizationfrommanufacturer).	05
6	**DeliveryScheduleatconsigneesend**	**10**
6.1	Within20days	10
6.2	Within30days	08
6.3	Within45days	06
	TOTALMARKS	**100**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

BIDDEREVALUATIONCRITERIA

Bid found to be responsive with maximum accumulative points (Technical + Financial) shall be accepted as the **mostadvantageousbid**asperSPPR,2010(Amended2021).

Bidswillbeevaluated onthebasisoffollowingevaluationcriteria.

A. TECHNICALEVALUATION

i) Technicalevaluationhas70%weightage

ii) BidswillbeevaluatedonaBidder'scapacityandcapabilitytoundertakesuppliesandonquoteditemquality,performanceandconformity totherequiredspecifications.

iii) Technical capacity of a bidder includes availability of workshop, tools, technical staff, number ofinstallation and experiencein the province of Sindh forprompt servicesto Past performance of the bidder at **Office of theDistrict Health Officer District Thatta**will also bereviewed.

iv) A bidder must achieve a minimum of 70 points out of total 100 points in technical evaluation, tobeconsiderèdforfinancialevaluation.

B. FINANCIALEVALUATION

i) Financialevaluationhas30%weightage.

ii) Only financial proposal of bidders with responsive technical proposal will be opened. Financialproposal of bidders who do not meet required mandatory documents and minimum 70 pointsoutoftotal100pointsintechnicalevaluationwillbereturnedunopened.

iii) Bidderwhohasofferedthelowestvaluefortheitemwill get 30 points.

iv) All remaining bidders after the lowest shall get points on their financial proposal, calculated asperthefollowing formula.

$$\frac{\text{Valueoflowestbidder}}{\text{Valueofhigher bidder [2nd, 3rd, 4th,etc.]}} \times 30$$

C. FINALBIDEVALUATION

i) Pointsobtainedbybiddersintechnicalandfinancialevaluationshallbeproratedandfinalizedas follows.

Evaluation	Weightage
Technical	70%
Financial	30%
Total	**100%**

Bidofabidderwhoisfoundtechnicallyresponsivewithhighestaccumulatedpointswillbeconsideredas themostadvantageousbid.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

SECTIONB

Supply of Hospital & General Furniture

S. No	Name of items	QTY	FIXED BID SECURITY
1.	**Almirah** **Specification:** Size: 3"x6"x18" made of MS Sheet 20 gauge solid 4 shelves Inter lock system with fine quality brass handle Spray paint • **Country of Origin:** IMPORTED/ EXPORT QUALITY.Should be:ISO-13485,CE/ISO-14001(Anyone). **Warranty:** Warrantyperiodshallbeofthreeyearsstartingfromdateofinstallationofequipmenttotheconsignee end.	04	1,800
2.	**Baby Cort** **Specification:** Mobile baby cot with removable bassinet made of acrylic Mounted on 4 swivel castors, heavy duty frame MS steel round pipe Transfer bars connect all lower distal portions of the 4 legs, providing maximal structural strength Lower shelf for equipment placing Clutch mechanism for tilting of bassinet With padded mattress, detachable for easy cleaning Mattress covers removable via side zipper Basinet sets and removes smoothly from cart frame **Material:** High resistance to corrosion (tropical environment) Frame: epoxy powdered coated Mattress: high-density with Rexene cover: flexible highly tear resistant, anti-static, disinfectant, liquid proof and washable Caster frame/bracket: steel or nylon Caster brake: total-lock type (wheel and rotational lock) Caster wheel: single wheel, mold-on type, non-hooded (for easy maintenance) Wheel bearing: sealed bearing in the swivel and the wheel **Dimensions :** Frame: 30 x 18 x 36 (inches) approx. Swivel castor wheels: 3" (imported) Mattes : 2" thickness Carrying capacity: 33-50 kg Knockdown construction: yes Accessories included: 1 x removal acrylic bassinet, 1 x fitting removable mattress with cover **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone). **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.	02	450

MEDICAL SUPERINTENDENT
GMMMC HOSPITAL SUKKUR

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer
Thatta

3.	**Delivery Table with Foot Step** **Specifications:** - Gynecological examination and delivery table, 3 sections - Mounted on 4 sturdy supports, all Anished with fixed height - All sections fitted with non-removable padded upholstery 4" Mattress (Master) - Robust mechanics allow for manual repositioning of all sections between - &ynecological and obstetric use - Transfer bars connect all lower distal portions of the 4 supports, providing maximal - structural strength - Back section: - Adjustable via secured clutch lever, safe for patient and operator - Pelvic section: - Fixed base position for sitting - Sides of this section are fit with side rails collapsable - Padded lithotomy poles are height adjustable, set with robust clamps with heavy knob - This section integrates a support for a slide-out basin-tray - Leg section: - Recesses entirely downwards, 90 degrees - When elevated and fully extended, all sections align to a perfectly flat surface. - High resistance to corrosion (tropical environment) - Frame: epoxy coated steel; all metal parts are epoxy coated steel - Feet covered with PVC - Padded upholstery: high-density polyurethane foam, density 27-33 kg/m3 - Cover: plastic, flexible highly tear resistant, anti-static, flame retardant, disinfectant- - and liquid proof, washable - Basin-tray: Austenitic stainless steel, 18/10 - All 3 sections extended, including padding: 6.5 x 2.5 x 3 ft - Carrying capacity: 135-165 kg - Slide-out basin-tray, capacity: 4L - Knockdown construction **Country of Origin: IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).** **Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end.**	02	1,500
4.	**Emergency Resuscitation Trolley** **Specification:** • Mobile resuscitation trolley with overhead warmer with integrated baby bassinet along with Cushioned mattresses. • Microprocessor control. • Audible and Visual high and low temperature alarms. • Display of skin Temperature, Air Temperature • Heater output (%), • APGAR Timer, • Resuscitation system **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW**	01	6000

MEDICAL SUPERINTENDENT
PEOPLES MEDICAL COLLAGE HOSPITAL
NAWABSHAH SBA...

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer
Thatta

	(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate. **Warranty: Warranty period shall be of three years starting from date of satisfactory installation of equipment.**		
5.	**Examination Couch** **Specifications:** - Dimensions: 72" x 24" x 28" - Body frame made up of 16SWG 1¼" MS round pipe - Welded base for stable operation - Backrest adjustable, with range up 40° to 50° and down 10° to 15° with friction clutch lever - Lying area 3" thick in two parts upholstered with foam &Rexene - PVC boots on feet for grip and noiseless operation - Electrostatic epoxy powder coated paint finish - OR Equivalent **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone). **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.	04	1,800
6.	**I/V Stand** **Specification:** Outer Stainless steel 16SWG pipe, Stainless steel telescopic inner rod Four stainless steel hooks, Spring loaded for drop safety Heavy iron base, epoxy powdered coating finished Mobile on castors **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone). **Warranty:** Warranty period shall be of three years starting from date of installation of equipment to the consignee end.	04	180
7.	**Instrument Trolley** **Specification:** Movableinstrumenttable,withtwoflatnon-removableshelves/traysHeavycarriagemountedon4swivelcastors,ofwhichtwowithbrake Transferbarsconnectalllowerdistalportionsofthe4base,providingmaximalstructuralstrengthShelves havesmoothfinishingallowingforeasycleaning. Top shelfof1mm304SSsheetprovided withguardallaround Lowershelf1mm304SS sheetisfullyflatsurfaceandfixedtothetableframerightabovetheswivel castors Trolley frame made of 18SWG 304 SS round pipe Dimension: 18 x 27 x 31(inches) Castors:2"(imported) **CountryofOrigin:**IMPORTED/EXPORT QUALITY.Shouldbe:ISO-13485,CE/ISO-14001(Anyone). **Warranty:** Warrantyperiodshallbeofthreeyears startingfromdateofinstallationofequipmenttotheconsigneeend.	02	2,400
8.	**Iron Rack** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	04	1,600
9.	**New Born Resuscitation Trolley** **Specification:** • Mobile resuscitation trolley with overhead warmer with integrated baby bassinet along with Cushioned mattresses.	02	12,000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

	• Microprocessor control. • Audible and Visual high and low temperature alarms. • Display of skin Temperature, Air Temperature • Heater output (%), • APGAR Timer, • Resuscitation system **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of satisfactory installation of equipment.**		
10	**Office Chair Wooden** **Specifications:** Round Shape made of sheesham wood Polish in good material. - Seat back cushioned with cloth and foam(Master Foam **-OREquivalent** **CountryofOrigin:**EXPORTQUALITY.Shouldbe:ISO-13485,CE/ISO-14001(Anyone). **Warranty:** Warrantyperiodshallbeofthreeyearsstartingfromdateofinstallationofequipmenttotheconsignee end.	18	2,700
11	**Office Table** **Specification:** - woodframewithlaminationtop inheavydutyfinish - Threedrawersonone sidewithhandles,locksandslidingonimportedrailsinfullextension - Highqualitymatt finishwoodlacquerpolish - Dimensions: L42"xW 24"xH26" - OREquivalent **CountryofOrigin:**EXPORTQUALITY.Shouldbe:ISO-13485,CE/ISO-14001(Anyone). **Warranty:** Warranty period shall be of three years starting fromdateofinstallationofequipmenttotheconsignee end.	08	3,600
12	**Patient Bed with Mattress** **Specifications:** Dimensions:78"x36"x22" Patientlyingareaintwo sections Reinforced10 stripsof1mmthickMSsheetplatform Bedframemadeof16SWGMStubular pipe Back-restadjustablebyindividualscrewcrank Crankhandletobelatchedunderfootend HeadboardandfootboardofABS famewithbakelite/HDFboard Headboardand footboard,lockable&easilyremovable Holdersformattressatfourpositionsonthebed Hooksforurinebagand NGbagateachsideofthebed HoldersocketsforIVpoleatdiagonalendsofthebed Mobileonfourimported5"castors(imported),two withswivel-typelockat diagonal end Workingloadcapacity:150Kg Electrostaticepoxypowdercoatedfinish FoamMattress'A'GradewithRexeneCover	04	6,000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Karachi.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Malir

	Dimensions: 78"x 36"x4" 10yearswarrantyormore OREquivalent **CountryofOrigin:**IMPORTED/EXPORTQUALITY.Should be: ISO-13485 (Mandatory), CE/JIS/ ISO-14001 (AnyTwo). **Warranty:** Warrantyperiodshallbeofthreeyearsstartingfromdateofinstallationofequipmenttotheconsignee end.		
13	**Revolving Chair** **Specifications:** - Mediumbackwitharmrestsrevolvingchair - MouldedP UFoam -Seat:20"x18" - Fabricupholsteredseat&backcushions - Highqualitygas springliftforheightadjustment **-OREquivalent** **CountryofOrigin:**EXPORTQUALITY.Shouldbe:ISO-13485,CE/ISO-14001(Anyone). **Warranty:** Warrantyperiodshallbeofthreeyearsstartingfromdateofinstallationofequipmenttotheconsignee end.	04	2,400
14	**Revolving stool stainless steel** **Specification:** Revolving patient examination stool. Top made of stainless-steel molded sheet. Central 1 3/4" round pipe with screw mechanism Legs made of round pipe fitted together with 5/8" round pipe for stability strength Complete frame is chrome plated (high quality) Rubber boot on feet for best floor grip **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	03	675
15	**S.S Bench** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	05	4,500
16	**Stretcher** **Specifications:** - Dimensions: 75" x 26" x 32" (L x W x H) - Frame made of 16SWG MS pipe - Fixed stretcher top with 2" foam mattress Rexene covered - Rubber bumper guards on all side for wall safety - Utility basket in base trolley - IV pole holder at diagonal end - Patient safety side guards of stainless-steel pipe which is slidable up and down - Mobile on four 8" rubber castors (diagonal breakable) - Electrostatic epoxy powder coated steel finish - OR Equivalent Accessories: 1) IV TRANSFUSION POLE 2 hooks with height adjustment & tight screw Outer MS pipe epoxy powder coated, Inner Chromed steel pipe **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone). **Warranty:** Warranty period shall be of three years starting from date of installation of	04	3,000

MEDICAL SUPERINTENDENT PEOPLES MEDICAL COLLAGE HOSPITAL NAWABSHAH SBA

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

	equipment to the consignee end.		
17	**Wheel Chair** **Specifications:** - Dimensions:32"x28"x20" - Mainframeconstructedfrom 16SWGMSroundpipe - Seatandbackupholsteredwithfoam&Rexene - Armrestofplastic - Rear wheelof8"non-swivel type - Front wheelof5"swiveltype - Rubberbumpersonchairsides - HolderforIVpoleononeside - Utilitybasketunderneath **Country ofOrigin:** IMPORTED/EXPORTQUALITY.Should be: ISO-13485 (Mandatory), CE/JIS/ ISO-14001 (AnyTwo). **Warranty:** Warrantyperiodshallbeofthreeyearsstartingfromdateofinstallationofequipmenttotheconsigneeend..	04	2,100

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer, Thatta

SECTIONC

Supply&InstallationofMedicalInstruments& Devices,ITEquipmentandMiscellaneousItems

Technical Evaluation Criteria for Section "C"MaximumAllocableMarksScore 70%toQualify

DocumentCheckList(MandatoryDocuments)

Ifanybidderfailedtosubmitanyofthefollowingdocuments,thesameshallbedisqualified.

Sr.No.	DOCUMENTDESCRIPTION	YES/NO	PAGE#
1	TenderPurchaseReceipt(Original) Copy of Pay Order		
2	BidSecurity(2%)Copywithvaluehidden inTechnicalProposal;OriginalinFinancial)		
3	TechnicalProposalonBidder'sLetterhead		
4	Compliancesheetonletterhead		
5	Undertaking(aspersampleform)		
6	IncomeTax&GSTRegistrationCertificate		
7	ValidProfessionalTaxCertificate		
8	FBRIncomeTaxReturns(last3years)		
9	AuditedFinancialStatement(last3year)		
10	Factory/WorkshopDetails(dulyequipped)		
11	ValidAuthorizationLetterfromManufacturerinreferencetothisparticulartender.		
12	Catalogue/Brochures /Drawing		
13	ISO9001ofBidders/Manufacturer(whereapplicable)**(websitelinkmustbeprovidedfor verification)**		

Note:

- Allpagesofthebid,exceptforun-amendedprintedliterature,shallbeinitialedbythe Bidder.
- Documentsmustbesubmittedinbinding,withfilledTableofContent,PageNumberingandtaggingwherev errequired.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Malir

EVALUATION/QUALIFICATION

1.1 Theprocuringagencyreservestherightto evaluateandcomparethebidsonitemizedbasisORonthebasisofagroupofsimilarnaturegoodsORgoodscompatibl ewitheachother.

1.2 Thefollowingmeritpointsystemforweighingevaluationfactors/criteriawillbeappliedfortechnicalproposals.

1.3 Biddersachievingminimumtotal**70%marks**willbeconsideredtoqualify.Documentaryevidencemustbeattachedi nsupportofeachparameter.

1.4 AnyBidnotmeetingthemandatoryrequirementsofevaluationcriteriawillbedisqualified/rejectedstraightawayan dwill notbeconsideredforfurtherevaluation.

A. PRODUCTEVALUATION

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
1	**ConformitytothePurchaser'sSpecifications(Mandatory)**	**40**
1.1	Fullycompliantwiththerequiredspecifications	40
1.2	Compliantwithminordeviation(upto10%subjectto mainfunctionisnotaffected)	30
1.3	Non-complianttorequiredspecifications	Disqualify
2	**ProductCertification**	**15**
2.1	USAFood&DrugAdministration(USA-FDA)andEuropeanCommunity(CE)MDDandJapan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) /CertificateissuedbyMinistryofHealth,Laborand WelfareGovt.ofJapan	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan QualityAssuranceOrganization(JQAO)/CertificateissuedbyMinistryofHealth,Laboran dWelfareGovt.ofJapan**OR** USA-FoodandDrugAdministration(USA-FDA)andEuropeanCommunity(CE)	12
2.3	USA-FoodandDrugAdministration(USA-FDA)	10
2.4	EuropeanCommunity(CE)MDD	08
2.5	JapanIndustrialStandard(JIS)/JapanQualityAssuranceOrganization(JQAO)/Certificateissued byMinistryofHealth,LaborandWelfareGovt.ofJapan	08
	TOTALMARKSPRODUCTEVALUATION(A)	**55**

B. BIDDEREVALUATION

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
1	**Manufacturer'sAuthorization**	**10**
2	**DeliveryScheduleatconsignee'send**	**10**
2.1	Upto40daysDDP	10
2.2	Morethan40 days	00
3	**PastExperience/ PerformanceofLast3years**	**10**
3.1	LastthreeYears'ExperienceoffivePublicSectorTertiaryCareHospitals(1Mar ks ofeach)	05
3.2	LastthreeYears'ExperienceoffivePrivateSectorTertiaryCareHospitals(1Ma rks ofeach)	05

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer

S#	PARAMETERS/SUB-PARAMETERS	Total Marks
4	AverageAnnualTurnoverduringlastthree(03)financialyears (AuditedStatementsof AccountsandIncomeTaxReturnFormsmustbeattachedassupportingdocuments)	10
4.1	TurnoverbelowPKR100million	00
4.2	TurnoverabovePKR100 million	04
4.3	TurnoverabovePKR200 million	06
4.4	TurnoverabovePKR300 million	08
4.5	TurnoverabovePKR500 million	10
5	ReplacementWarranty	05
5.1	ReplacementWarrantyfor1year	05
	TOTALMARKSBIDDEREVALUATION(B)	45
	GRANDTOTAL(A+B)	100

MEDICAL SUPERINTENDENT
PEOPLES MEDICAL COLLAGE HOSPITAL
NAWABSHAH SBA

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer

BIDDEREVALUATIONCRITERIA

Bid found to be responsive with maximum accumulative points (Technical + Financial) shall be accepted as the **mostadvantageousbid**asperSPPR,2010(Amended2021).

Bidswillbeevaluated onthebasisoffollowingevaluationcriteria.

A. TECHNICALEVALUATION

i) Technicalevaluationhas70%weightage

ii) BidswillbeevaluatedonaBidder'scapacityandcapabilitytoundertakesuppliesandonquoteditemquality,performanceandconformity totherequiredspecifications.

iii) Technical capacity of a bidder includes availability of workshop, tools, technical staff, number ofinstallation and experiencein the province of Sindh forprompt servicesto KarachiDivision.Past performance of the bidder at**District Health Officer District Thatta**will also bereviewed.

iv) A bidder must achieve a minimum of 70 points out of total 100 points in technical evaluation, tobeconsideredforfinancialevaluation.

B. FINANCIALEVALUATION

i) Financialevaluationhas30%weightage.

ii) Only financial proposal of bidders with responsive technical proposal will be opened. Financialproposal of bidders who do not meet required mandatory documents and minimum 70 pointsoutoftotal100pointsintechnicalevaluationwillbereturnedunopened.

iii) Bidderwhohasofferedthelowestvaluefortheitemwillget30points.

iv) All remaining bidders after the lowest shall get points on their financial proposal, calculated asperthefollowing formula.

$$\frac{\text{Valueoflowestbidder}}{\text{Valueofhigher bidder [2nd, 3rd, 4th,etc.]}} \times 30$$

C. FINALBIDEVALUATION

i) Pointsobtainedbybiddersintechnicalandfinancialevaluationshallbeproratedandfinalizedas follows.

Evaluation	Weightage
Technical	70%
Financial	30%
Total	**100%**

ii) Bidofabidderwhoisfoundtechnicallyresponsivewithhighestaccumulatedpointswillbeconsidereda s themostadvantageousbid.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

SECTIONC

Supply&InstallationofMedicalInstruments& Devices,ITEquipmentandMiscellaneousItems

S. No	Name of items	QTY	FIXED BID SECURITY
1.	**Air conditioner with Stabilizers** **Specification:** BTU Capacity 18000 BTU Compressor W 5300 Low Voltage Startup 150V Eco Friendly Refrigerant R410A Energy Efficient Class A+ 60% Electricity s Auto Voltage Adaptation 150V-260V **COOLING** BTU/H Capacity (H/S/L*) 19100/17750/5118 Output Capacity (W) (H/S/L*) 5600/5200/1500 Power Input (W)(H/S/L*) 2100/1529/390 Energy Efficiency Ratio (EER) 3.6 HEATING BTU/H Capacity (H/S/L*) 22550/19600/4260 Output Capacity (W) (H/S/L*) 6600/5750/1450 Power Input (W)(H/S/L*) 2200/1580/390 Energy Efficiency Ratio (EER) 3.6 With compatible imported/export quality stabilizer. Complete installation and commissioning are the responsibility of the vendor Warranty three years including all parts Compressor warranty ten years Import/Export Quality ISO/CE certified Warranty 03 years from the date of installation	01	6,000
2.	**Ambo Bag (Adult)** **Specification:** • Made of high-quality silicone and special plastic • Autoclavable • Silicon size: 550ml (±10%) • O2 tube and reservoir valve attachable • Ventilation capacity: 600ml per one time • OR Equivalent Configuration • Resuscitator Body (1 ea) • Resuscitator Mask (1 ea) • Oxygen Reservoir (1 ea)	04	780

MEDICAL SUPERINTENDENT
GMMMC HOSPITAL SUKKUR

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer

	Country of Origin: UK, USA, Europe, Japan or equivalent. **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of satisfactory installation of equipment to the consignee end.**		
3.	**Ambo Bag (Infant)** **Specification:** • Made of high-quality silicone and special plastic • Autoclavable • Silicon size: 550ml (±10%) • O2 tube and reservoir valve attachable • Ventilation capacity: 600ml per one time • OR Equivalent Configuration • Resuscitator Body (1 ea) • Resuscitator Mask (1 ea) • Oxygen Reservoir (1 ea) **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of satisfactory installation of equipment to the consignee end.**	**04**	**780**
4.	**B.P Apparatus** **Specifications:** • The maximum number of cuff inflations for each SP in the mercury • Measurement is five, counting all MIL attempts and blood pressure attempts. • The rationale for this is twofold: to minimize the discomfort to the SP of frequent cuff inflations and to accomplish data collection for this measurement within the time allowed • Mobiel Stand. ISO 13485 **FDA/CE/JIS APPROVEDUSA/EUROPE/JAPAN** **Manufacturers from other countries (other than USA, Europe or Japan) will be accepted if equivalent has FDA 510(k), CE, JIS or JQAO certifications (Any Two Certificates**	**04**	**960**
5.	**Baskets** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	**06**	**90**
6.	**Blood Grouping Viewing boxes** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	**01**	**60**
7.	**Centrifuge Tubes (Glass)** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	**06**	**360**
8.	**Centrifuge Tubes (Plastic)** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	**06**	**360**

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Hyatta

9.	**DNC Kit** • Metallic Catheter 1 • Uterine Sound 1 • Sims Speculum Medium 1 • Set D&E Sponge Holder 1 • Hegar's Dilator 0.8 Cm 1 • Kidney Tray 1 • Voisellum 8" 1 • Set Uterine Chortle 1 • Plain forceps 8" Bowl 4-Inch 1 **Standards: ISO 13485** **CE/FDA approved** **Local export quality**	02	3,000
10	**Deep Freezer** **Specification:** Twin Door Convertible Chest Freezer with 4 in 1 Convertible modes Normal, Fresh, Chill & Freeze Mode Up to 35% Energy Saving (with R600A) Low Voltage Startup 5- Way Evaporator for Faster Cooling Up to10 years compressor warranty Import/Export Quality ISO/CE certified **Warranty 03 years from the date of installation**	02	5,100
11	**Delivery Light for Labor Room** **Specification:** Input voltage: 220-240VAC, 50/60Hz. International ratings LED Driver with High Power Factor are installed as standard. 50,000 hrs. life span, low luminous decay **Country of Origin: IMPORTED/ EXPORT QUALITY.** **Should be: ISO-13485,** **CE/ ISO-14001 (Anyone).** **Warranty 03 year from the date of installationCertification: ISO9001 Certified Manufacturer**	02	1,800
12	**Dressing Drums** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	03	675
13	**ESR Pipettes** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	03	540
14	**ESR Racks** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	02	360
15	**Electric Stabilizer** IMPORTED/ EXPORT QUALITY. **Should be:** ISO/CE certified **Warranty:** Warranty period shall be of three years starting from date of satisfactory installation of equipment to the consignee end.	01	900
16	**Gas Burner** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	02	120

MEDICAL SUPERINTENDENT
PEOPLES MEDICAL COLLAGE HOSPITAL
NAWABSHAH SBA

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer
Matta

17	**Glass Pipettes Various Sizes Corrected** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	06	2,160
18	**Glass Slides** **Country of Origin:** IMPORTED/ EXPORT QUALITY. Should be: ISO-13485, CE/ ISO-14001 (Anyone).	02	12
19	**Glass Ware** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	01	6000
20	**Hemoglobin Meter** **Specification:** • LED Display. • Measuring Rang 0-40 gm/di of HB. • Zero Adjustment Automatic. • Dilution ratio 1.251. • Key Board Soft touch 4 days. • Calibration Automatic. • Wavelength 530 nm. • Detector High sensitive silicon photo diode. **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end.**	01	1500
21	**Hepatitis (B,C, & HIV AIDs Kits Syphils** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	04	660
22	**Ice Cold Box** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	05	5,250
23	**Instruments Tray** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	02	510
24	**Jester Pipettes Fixed various sizes** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	04	1,800
25	**Jeter Pipettes adjustable various sizes** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	06	2,700
26	**Lab Chemicals** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	01	240
27	**Lancets (Pack)** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	04	780
28	**NVD Kit** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	02	3,000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc.):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer
Thatta

29	**Nebulizer (Baby & Adult)** **Specifications:** • Continuous high-pressure performance. • Ultra-powerful compressor for both humidification and high flow • High-pressure nebulization of viscous medication. • Optimize medication delivery by aerosolizing to a particle size of < 2 microns. • Lightweight, durable and easy-to-clean design. • Detachable door that protect componentry. • Maximum Compressor Pressure: up to 80 psi • Quiet Operation: with a 51 dBA noise level, the heavy-duty air compressor is one of the quietest continuous aerosol compressors on the market. • Continuous high-pressure performance. **ISO 13485** **FDA/CE/JIS APPROVEDUSA/EUROPE/JAPAN** **Manufacturers from other countries (other than USA, Europe or Japan) will be accepted if equivalent has FDA 510(k), CE, JIS or JQAO certifications (Any Two Certificates**	04	4,800
30	**Oxygen Cylinders with flow meter** **Specification:** • Cylinder240 CFT. • BOCTested. • BrandNew. • With Flow meter. CountryofOrigin:IMPORTED. Should Be: ISO Approved Or CE Marked. Warranty:ShouldbeStandard02YearsWarranty	03	3,600
31	**Pipettes Stands** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	01	270
32	**Reagent** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	01	3000
33	**Refrigerator (Chiller) for sample storage** **Specification:** For storage of specimens, diagnostic kits, reagents and Anti serum Controlled fan cooling system Glass door with safety insulation Safety thermostat Fan based refrigeration system 220V 50 Hz, AC User Adjustable Settings: Temperature control should be available on panel with 2 levels of security Temperature Set Ranges: +2 to +12 Degree Centigrade Alarm Settings: +15 Degree Centigrade Gross Volume: 260 Liters or more Hold over Time: 30 Minutes Displayed Parameters: Color display for temperature and settings Digit audible and visual alarm for door ajar, power failure, automatic defrosting system, temperature indicator, digital temperature display alarm, silence button, lights switch & temp Alarms must be battery driven Door opening alarm Complete with standard and operation accessories Three shelves Internal light Servo Controlled Voltage Stabilizer with surge protection facility	01	6000

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

	Operating Manual with a Soft Copy **Country of Origin: UK, USA, Europe, Japan or equivalent** **Should be: ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW** **(Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two Certificate.** **Warranty: Warranty period shall be of three years starting from date of installation of equipment to the consignee end.**		
34	**Refrigerator (ILR)** **Refrigerator 16 cu ft.** Handle less design, low voltage technology-can run at as low as 135 volts, no use of stabilizer. IMPORTED/ EXPORT QUALITY. **Should be:** ISO/CE certified **Warranty:** Warranty period shall be of three years starting from date of satisfactory installation of equipment to the consignee end.	01	6000
35	**Set of Dressing Instruments** • **Mosquito forceps 5" STR+CVD** 6 • **Artery Forceps 6" STR+CVD** 6 • **Kocher forceps 6" STR+CVD** 6 • **Needle Holder 6"** 2 • **Needle Holder 7"** 2 • **Scalpel #3+4** 4 • **Stitch Scissor** 2 • **Dressing Scissor 6"** 2 • **Dressing forceps Plain 6"** 2 • **Dressing forceps tooth 6"** 2 • **Towel Forceps** 4 • **Instrument box 12"** 1 • **IRIS Scissor** 2 • **Metzenbun Scissor 7"** 2 • **Cruet** 2 **Standards: ISO 13485** **Local Export QualityCE/FDA approved**Warranty period shall b of three years starting from date of installation of equipment to the consignee's end.	02	2,400
36	**Stainless steel test tube racks** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	02	150
37	**Stethoscope** **Specifications:** • Dual Head Stethoscope for Adults • Lightweight aluminum chest piece construction • Non-chill bell and diaphragm rim for patient comfort • Diaphragm diameter 1 3/4", bell diameter 3/4" • Proprietary diaphragm enhances acoustic response by 50% over traditional designs • Adjustable aluminum binaurals with PVC ear tips • Flexible 21" PVC tubing • Included with spare pair of black mushroom ear tips **ISO 13485** **FDA/CE/JIS APPROVED** **USA/EUROPE/JAPAN**	04	360

MEDICAL SUPERINTENDENT
PLES MEDICAL COLLAGE HOSPITAL
NAWABSHAH SBA

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer
Matta

	Manufacturers from other countries (other than USA, Europe or Japan) will be accepted if equivalent has FDA 510(k), CE, JIS or JQAO certifications (Any Two Certificates		
38	Stop Watch Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	01	105
39	Surgical Blades Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	06	216
40	Table Lamp Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	01	165
41	Test Tube Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	04	360
42	Test Tube Holder Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	06	540
43	Test Tube Racks Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	01	90
44	Tray Kidney shaped 10 inches Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	02	210
45	Tray Kidney shaped 8 inches Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	02	180
46	Tube Seal Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	02	60
47	Urine Meter Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	01	90
48	Vaccine Carrier Certification: ISO9001 Certified Manufacturer Make: Local Export Quality or Equivalent	05	1,500
49	Vacuum Extractor apparatus Specification: GENERAL DESCRIPTION Vacuum extractor, Bird (anterior & posterior), manual, complete set. INTENDED USE An assembly of hand-operated devices designed to facilitate the delivery of a fetus during vaginal childbirth or Caesarean section by enabling traction to be applied to the fetal head by means of a suction cup that is attached to the scalp. It typically consists of a cup with a soft inner liner to prevent tissue damage, a hand-operated pump, and a handle used to apply traction. It is typically used for dystocia, uterine inertia, maternal exhaustion, maternal/fetal distress. This is a reusable device. TECHNICAL SPECIFICATIONS Vacuum extractor system to assist vaginal deliveries.	01	1,500

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer

	Hand-operated suction pump supports the vacuum extraction. Vacuum range, adjustable: from 0 to -85kPa/-638mmHg or better Gauge dial, minimal graduation: 5kPa/40mmHg or better Accuracy vacuum gauge: ± 2.5kPa or better All parts of the vacuum extractor can be disassembled. All parts can be disassembled and are autoclavable at 121°C requires no other than routine maintenance, manageable at the level of delivery rooms. **SYSTEM IS SUPPLIED AS A COMPLETE SET CONSITSTING OF** 1 x Vacuum pump 2 x Bird type anterior suction cup, stainless steel: diameter rounded edges 1 x Bird type posterior suction/extraction cup rounded edges 2 x Soft vacuum extractor cups "silc-cup" 2 x Extraction handle, stainless steel 1 x Cover plug for vacuum collection bottle, with 3 holes (2 connectors +1 vacuum gauge) 1 x Connector, angled, chrome plated 1 x Connector angled with screw valve, chrome plated screw, cap and gasket 1 x Vacuum gauge (0 to -100 kPa/- 800mmHg) with straight connector, chrome plated 1 x Vacuum collection bottle, plastic 500ml 1 x Basket, epoxy coated, supporting vacuum collection bottle and the vacuum pump, with large hook allowing to fit it to the delivery table. 2 x Tube, silicone transparent, inner/outer diameter approx.: 6/11mm, respective lengths: 50 and 150cm IMPORTED/ EXPORT QUALITY. **Should be:** ISO/CE certified **Warranty:** Warranty period shall be of three years starting from date of satisfactory installation of equipment to the consignee end		
50	**Vacuum Suction Apparatus** **Specification:** • Shouldhaveabilitytobeusedintheemergencydept.,intensivecare,operatingtheatre. • Thepump shouldhaveaccuratevacuum settingwiththemembrane vacuumregulator • Shouldhavepiston/cylindersystemforavibrationfreeandnoise lessoperation. • The vacuum range should start from -95 kPa / 713mmHg orbetter. • Theflowrateshouldbeadjustablefrom60 or more liter/min. • PowerRequirement: Line Voltages 100-240 VACLineFrequency 50/60Hz **Shouldbe Suppliedwith:** Tworeusablesuction jarsof 4liters MobileTrolleyOriginalwithtwolockablecastorAntibacterialF ilter ClampHolder Disposable Patient TubingPower Cord OperatingManual **Country of Origin: UK, USA, Europe, Japan or equivalent.** **Should be:** ISO-13485 (Mandatory) & CE, JIS, FDA, MHLW (Other than UK, USA, Europe & Japan) will be accepted if equivalent have any Two	01	1,500

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Malir

	Certificate. **Warranty**: Warrantyperiodshallbeofthreeyearsstartingfromdateofsatisfactory installationof equipment.		
51	**Wooden Boxes** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	02	60
52	**Wooden Slide Box** **Certification: ISO9001 Certified Manufacturer** **Make: Local Export Quality or Equivalent**	01	300

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer

HUMANRESOURCEINCLUDINGDETAILOFTECHNICALTEAM(FORMA)

S.#	Name	Designation	PostedLocation	DateofJoining	Education	Trainings

Note:SupportingDocumentsmustbeprovidedforeachemployee.

CEO/Proprietor/ManagingDirector

Name:_____

EmailAddress:_____

CellPhone:_____

Director/GeneralManagerServices

Name:_____

EmailAddress:_____

CellPhone:_____

Director/GeneralManagerSales

Name:_____

EmailAddress:_____

CellPhone:_____

TechnicalFocalPersonforthisProject

Name:_____

EmailAddress:_____

CellPhone:_____

Address:_____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar._____
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI._____
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop._____
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar._____
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi._____

District Health Officer

SPECIALIZEDTESTINGANDCALIBRATIONTOOLS(FORMB)

SupportingDocumentsmustbeprovidedforeachspecializedtool

S.#	ToolDescription	Make	Model	DateofPurchase	Date of LastCalibration

MAJORSPAREPARTS(FORMC)

- Purchasedocumentsmustbeprovided,
- it may include Sensor, PCB, LCD display, Pumps, motors, detectors, Laser and other itemsrecommendedbymanufactureretc.,listofrecommendedsparepartsmustalsobeprovided.
- Partssuchascables,connectorswillnot beconsideredmajorspareparts.
- Bidderusinganinventorymanagementsoftwaremayprovideasoftwaregeneratedlistinstead ofthebelowformat.

S.#	PartNumber	PartDescription	Make	Model	DateofPurchase	Qty

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer, Thatta

COMPLIANCESHEET/TECHNICALEVALUATIONSHEET(FORM D)

Namethe Vendor:

Make/Manufacturer:

Model: Yearofintroduction:

CountryofOrigin:

Country of Manufacturing: Country ofMajorpartssourcing:

Usefulllifeaspermanufacturer(inYears):CE(MDD)link:

FDA(US)Link:

JIS/MHLW Link:

ISO13485Link:

Pre-installationrequirement.

ElectricalConnection:

Grounding:

BackupSupply:

Water(RO/DIwithflow):

Drain:

AnyotherRequirement:

S.#	TenderSpecification	OfferedSpecification	Page#ofthesupportingDocument

Note:Supportingdocumentsmustbeprovided

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Thatta

FINANCIALEVALUATIONSHEET(FORME)

	AnnualTurnoverforthelast3yearsYear(PKR)
Year2025	
Year2024	
Year2023	

FinancialInformation inPKR	Year2025	Year2024	Year2023
TotalAssets (TA)			
TotalLiabilities(TL)			
CurrentAssets(CA)			
CurrentLiabilities(CL)			
Total/ GrossRevenue(TR)			
ProfitsBeforeTaxes (PBT)			
NetProfit			
CurrentRatio			

SupportingDocument		Yes/No
BankStatement	Year2025	
	Year2024	
	Year2023	
AuditedAccounts	Year2025	
	Year2024	
	Year2023	
BankAccountMaintenanceCertificate		

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Thatta

MANUFACTURER'SAUTHORIZATION(

FORMF)

[SEECLAUSE 11.1(A)OFTHEINSTRUCTIONTO BIDDERS]

To:

District Health Officer District ThattaWHEREAS_____**[nameoftheManufacturer]**__who are established and reputableManufacturersof _____**[nameand/ordescriptionofthegoods]**_____havingfactoriesat____ _____**[addressoffactory]**_____doherebyauthorize**[nameandaddressofBidder/Agent]** _____to submita bid, and subsequentlyfollow-up/ negotiateand signtheContractwithyouagainstInvitationforBids(IFB)/TenderNoticeforthegoodsmanufactured,byus, underthepatentnameof

_____forperformanceofthecontract.

We hereby commit and assure our full guarantee and warranty / guarantee as per Clause 12 of the GeneralConditions of Contract for the goods offered for supply by the above mentioned firm against this Invitationfor Bids.

[SignatureforandonbehalfofManufacturer]N
ote:
This letter of authority should be on the letterhead of the Manufacturer and should be signed by a personcompetentandhavingthepowerofattorneytobindtheManufacturer.ItshouldbeincludedbytheBidderin its bid.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI.
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.

District Health Officer Thatta

LISTOFINSTALLATION/REFERENCECLIENT(FORMG)

Equipment: _____

Make/Manufacturer: _____

Model: _____

S.#	Department	NameofInstitute	City	Dateof Installation	DateofSatisfactory Certificate

Note: Attachthefollowingsupportingdocuments.

- Performancecertificate(issuedinlast oneyear)onletterheadofbidder/Hospital
- InstallationReportsignedbyHeadofdepartmentandBiomedicalEngineer
- Supplyorder/purchaseorder

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar. _____

b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI. _____

c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop. _____

d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar _____

e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi. _____

District Health Officer, Thatta

PERFORMANCECERTIFICATE(FORMH)

[onbidder'sletterhead]

Thisistocertifythattheinstituteissatisfiedbytheperformanceofthefirmandequipment asdetailedbelow.

PurchaseOrderNo: ,dated
Make&Modelno.:
SerialNumber:
Quantity:
Nameoftheconsignee:

TheFirmhasfulfilledhiscontractualobligationregardingthefollowingservices:
SatisfactoryInstallation,performance,andcommissioningofequipment.
Furnishingdetailedoperationandmaintenancemanualforeachequipment.
Trainingoftheoperators/usersinoperatingtheequipmenttothesatisfactionoftheInstitute.
Maintainuptimeof__%
PerformPreventativemaintenanceat___Intervalduringwarranty/contractperiod
ProvideSparepartstimely

BiomedicalEngineeringDepartment.

Name _____

Designation _____

Date _____

Signaturewith Stamp _____

HeadoftheInstitute/MedicalDirector/MedicalSuperintendent

Name _____

Designation _____

Date _____

Signaturewith Stamp _____

Representativeofthefirm

Name _____

Designation _____

Date _____

Signaturewith Stamp _____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

District Health Officer

BIDLETTERFORM

From:
(Registerednameandaddressofthe
bidder)To:

District Health Officer District Thatta DearSir /Madam,

Having examined the bidding documentand amendment thereon we the undersigned,offer toprovideservicestotheworksincludinginconformitywiththetermsandconditionsofthebiddingdocumentandamendmentsthereon,forthefollowingprojectinresponsetoyourtendercalldated_____

TenderTitle:... ....

We undertake to provide services/execute the above project or it part assigned to us in conformity withthe said bidding documents for an estimated sum of Rs._____(Rupees-_____) (total bid amount in words and figures) whichmay vary in accordance with the schedule of prices attached herewith and coverage options made by DrugTestingLaboratory Sindhoritsuserorganization.

Ifourbidisaccepted,we undertaketo;

1) Provideservices/executetheworkaccordingtothetimeschedulespecifiedinthe biddocument,
2) Obtaintheperformanceguaranteeofbankinaccordancewithbidrequirementsfortheduеperformanceofthe contract, and
3) Agree to abide by the bid conditions, including pre-bid meeting minutes if any, which remain bindingupon us during the entire bid validity period and bid may be accepted any time before the expiration ofthatperiod.
4) Weunderstandthatyouarenotboundtoacceptthelowestoranybidyoumayreceive, nor togiveanyreasonfortherejection ofanybidandthatyouwillnotdefrayanyexpenses incurred byusinbidding.

Place:
Date:

Bidder'ssignature
andseal.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer
Thatta

CONTRACTFORM

THISAGREEMENTmadethe……Dayof ..... (year)BetweentheProcuringAgency(hereinafter "the
_____")ofonepartand……(NameofVendor)of……..(Cityan dcountryofVendor)(Hereinafter"theSupplier")oftheotherpart:
WHEREASthe
_____isdesirousthatcertainSupplies,asdescribedinthebiddocumenta ndbriefly outlinedbelow,shouldbeprovidedbytheVendor.

Dateoftendercall:
Titleoftheproject:
Briefoutlineofthework:

NOWTHISAGREEMENTWITNESSETHASFOLLOWS;

Inthisagreementwordsandexpressionshallhavethesamemeaningsasarerespectivelyassignedtothem inthebiddocumentreferredto.

Thefollowingdocumentshallbe deemedtoformandbereadandconstruedaspartofthisContract,viz..

1) Biddocument(s)
2) Pre-bidconferenceminutesifany,
3) Clarificationonbiddocumentissuedifany,
4) District Health Officer District Thatta award notification.

In case of conflict among documents mentioned above, the documents mentioned above in reverse orderwill prevail over other documents.In consideration of the payments to be made by the **District Health Officer District Thatta**to the Supplier as hereinafter mentioned, the Supplier herebycovenantswith theDistrict Health Officer District Thatta(**TenderTitle**.. .. ..)andtoremedydefects thereinconformity,inallrespects,withtheprovisions ofthecontract.

TheDistrict Health Officer District Thattahereby covenants topay theSupplierinconsiderationofthe provisionof thegoods and servicesandthe remedying ofdefects therein, thecontract price or such other sum as may become payable under the provisions of the contract at the timesandinthemannerprescribedby thecontract.

Briefparticularsoftheserviceswhichshallbesupplied/providedbytheSupplierareasunder:

Solutions,serviceormaterial	Quantity	Unitprice	Amount	Remarks

INWITNESSwhereofthepartiesheretohavecausedthisAgreementexecutedthedayandyearabovewritten.
Signed,sealed,delivered by_____the (for the Procuring agency)Signed,sealed,deliveredby_____the(fortheSupplier)

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

FORMOFPERFORMANCESECURITY
(BankGuarantee)

GuaranteeNo.:_____

_____

Executedon:_____

[LetterbytheGuarantortotheEmployer] Expirydate:_____

NameofGuarantor(Bank)withcompleteaddress(ScheduledBankinPakistan):

_____

NameofPrincipal(Contractor,Manufacturer,Supplieroranybidder)withcompleteaddress:

_____

PenalSumofSecurity(expressinwordsandfigures):

_____

_____

LetterofAcceptanceNo._____Dated:_____

KNOW ALL MEN BY THESE PRESENTS, that in pursuance of the terms of the Bidding Documents and abovesaid Letter of Acceptance (hereinafter called the Documents) and at the are e of the said Principal we, theGuarantor above named, are held and firmly bound unto the**District Health Officer District Thatta**(hereinafter called the Employer) in the penal sum of the amount stated above for thepayment of which sum welland truly to be made to the said Employer, we bind ourselves, our heirs,executors,administratorsandsuccessors,jointly andseverally,firmlyby thesepresents.

THECONDITIONOFTHISOBLIGATIONISSUCHthatwhereasthePrincipalhasacceptedtheEmployer's abovesaidLetterofAcceptancefor _____ (NameofContract)forthe _____(NameofProject).

NOW THEREFORE, if the Principal (Contractor) shall well and truly perform and fulfill all the undertakings,covenants, terms and conditions of the said Documents during the original terms of the said Documentsand any extensions thereof that may be granted by the Employer, with or without notice to the Guarantor,which notice is, hereby, waived and shall also well and truly perform and fulfill all the undertakings,covenants terms and conditions of the Contract and of any and all modifications of said Documents thatmay hereafter be made, notice of which modifications to the Guarantor being hereby waived, then, thisobligation to be void; otherwise to remain in full force and virtue till all requirements of Condition ofContractarefulfilled.Our total liability under this Guarantee is limited to the sum stated above and it is a condition of anyliabilityattachingtousunderthisGuaranteethattheclaimforpaymentinwritingshallbereceivedbyus

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____

c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.** _____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

District Health Officer Thatta

within the validity period of this Guarantee, failing which we shall be discharged of our liability, if any,underthisGuarantee.

We,
_____(theGuarantor),waivingallobjectionsanddefensesundertheContract,doherebyirrevocably and independently guarantee to pay to the Employer without delay upon the Employer's firstwritten demand without cavil or arguments and without requiring the Employer to prove or to showgroundsor reasonsforsuch demandanysumor sumsup to the amountstated above, against theEmployer's written declarationthat the Principal has refused or failed to perform the obligations under the Contract which payment willbeeffectedbytheGuarantor toEmployer'sdesignatedBank&AccountNumber.

PROVIDED ALSO THAT the Employer shall be the sole and final judge for deciding whether the Principal(Contractor)hasdulyperformedhisobligationsundertheContractorhasdefaultedinfulfillingsaid obligations and the Guarantor shall pay without objection any sum or sums up to the amount stated aboveupon first written demand from the Employer forthwith and without any reference to the Principal or anyotherperson.

IN WITNESS WHEREOF, the above-bounden Guarantor has executed this Instrument under its seal on thedate indicated above, the name and corporate seal of the Guarantor being heretoaffixed and thesepresentsdulysignedbyitsundersigned representative,pursuanttoauthorityofitsgoverningbody.

Witness:

_____
Guarantor(Bank)

1._____
(Name,Title,Signature&Seal)

2. _____
(Name,Title,Signature&Seal)

Signature:_____
Name:_____
Title:_____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**

District Health Officer Thatta

INTEGRITYPACT

DECLARATIONOFFEES,COMMISSIONANDBROKERAGEETCPAYABLEBYTHESUPPLIERS/CONTRACTORS/CONSULTANTS

ContractNumber: NO._____ Dated:_____
ContractValue: Rs._____
ContractTitle: _____

M/s._____hereby declares that it has not obtained or induced the procurement of anycontract, right, interest, privilege or other obligation or benefit from Government of Sindh (GoS) or anyadministrative subdivision or agency thereof or any other entity owned or controlled by it (GoS) throughanycorruptbusiness practice.

Withoutlimitingthegeneralityofthe foregoing,**M/s.**_____represents and warrants that ithas fully declared the brokerage, commission, fees etc. paid or payable to anyone and not given or agreedto give and shall not give or agree to give to anyone within or outside Pakistan either directly or indirectlythrough any natural or juridical person, including its affiliate, agent, associate, broker, consultant, director,promoter, shareholder, sponsor or subsidiary, any commission, gratification, bribe, finder's fee or kickback,whetherdescribedasconsultationfeeorotherwise,withtheobjectofobtainingorinducingtheprocurement of a contract, right, interest, privilege or other obligation or benefit, in whatsoever form,**District Health Officer District Thatta,** except that which has been expressly declaredpursuanthereto.

M/s._____certifies that it has made and will make full disclosure of all agreements andarrangements with all persons in respect of or related to the transaction with PA and has not taken anyactionorwillnottakeanyactiontocircumventtheabovedeclaration,representation or warranty.

M/s._____acceptsfullresponsibilityandstrictliabilityformakinganyfalse declaration,notmakingfulldisclosure,misrepresentingfactsortakinganyactionlikelytodefeatthepurposeofthisdeclaration,representationandwarranty.Itagreesthatanycontract,right,interest,privilegeorotherobligationorbenefitobtainedorprocuredasaforesaidshall,withoutprejudicetoanyotherrightandremedies available to PA under any law, contract or other instrument, be voidable at the option of PA.Notwithstanding any rights and remedies exercised by PA in this regard, **M/s.**_____agrees to indemnify PA for any loss or damage incurred by it on account of itscorrupt business practices and further pay compensation to PA in an amount equivalent to ten time thesum of any commission, gratification, bribe, finder's fee or kickback given by **M/s.**_____as aforesaid for the purpose of obtaining or inducing the procurement of anycontract,right,interest,privilegeorotherobligationorbenefit,inwhatsoeverform,fromPA.

M/s._____ District Health Officer District Thatta

District Health Officer, Thatta

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar._____
b. Majid Ali, Senior Equipment Engineer, NICH, KARACHI._____
c. Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop._____
d. Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar._____
e. Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi._____

BidForm&PriceSchedule

BIDFORM

To:

District Health Officer District Thatta DearSir,

HavingexaminedtheBiddingDocuments,thereceiptofwhichisherebydulyacknowledged,we,theundersigned,offertosupplyanddeliverthegoodsspecifiedinthesaidBiddingDocumentsforthesumof _____ **[TotalBidAmountRs.** _____ **],[BidAmountinwords** _____ **only]**orsuchothersumsasmaybeascertainedinaccordancewiththeScheduleofPricesattachedherewithandmadepartofthis bid.

1. Thefreeof cost/donation/ discountsofferedandthemethodologyfortheirapplicationare:

_____

_____

2. Weundertake,ifourbidisaccepted,todeliverthegoodsinaccordancewiththedeliveryschedulespecifiedintheScheduleofRequirements.
3. Ifourbidisaccepted,weshallobtainanunconditionalguaranteeofabankinthesumof5%oftheContractPriceforthedueperformanceoftheContract,intheformprescribedbytheProcuringAgency.
4. Weagreetothevalidityofthisbidfor90daysfromthedatefixedforfinancialbidopeninganditshallremain bindinguponusandmaybeacceptedatanytimebeforetheexpiration ofthatperiod.
5. UntilaformalContractispreparedandexecuted,thisbid,togetherwiththewrittenacceptancethereofandnotificationofaward,bytheProcuringAgency,shallconstituteabindingContractbetweenus.
6. Weunderstandthatyouarenotboundtoacceptthelowestoranybidyoumayreceive.

Name_____

Inthecapacityof_____

Signed_____

Dulyauthorizedto signtheBidforand onbehalf of_____

Date_____

(A) PRICESCHEDULE INPAKRUPEESdelivereddutypaid(DDPBASIS) FORGOODSOFFEREDWITHINTHEPROCURINGAGENCY'SCOUNTRY

S#	DetailedSpecification ofGoods	Model/ CatNo.	Nameof Manufacturer	Country ofOrigin	Portof Shipment	Quantity ofStores	Unit	RateP erUnit	Total Price
1	2	2	4	5	6	7	8	9	10
TotalAmountinPakRs.									

Name_____

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**
e. **Rajeev Brahmani, Pharmacist, Dr. Ruth K.M.Pfau Civil Hosp. Karachi.**

District Health Officer Thatta

Port of Shipment and Country of origin of "MAJOR PART(S) OF THE EQUIPMENT" mustbe clearly reflected separately in the Technical and Financial bids. The "Origin" means the place where the "goods" are mined, grown, produced,

Inthecapacityof_____

Signed_____

Dulyauthorizedto signtheBidforand onbehalf of_____

NOTE:

(B) PRICESCHEDULEINFOREIGNCURRENCY
(CFR/CNF/C&F/CPTBASIS)
FORGOODSOFFEREDFROMOUTSIDETHEPROCURINGAGENCY'SCOUNTRY

S #	DetailedSpecificationofGoods	Model /CatNo	Name of Manufacturer	Country ofOrigin	Port ofShipment	Quantity ofStores	Unit	Curr-Ency	RateP erUnit	Total Price
1	2	2	4	5	6	7	8	9	10	11
TotalAmountinForeignCurrency										

Name_____

Inthecapacityof_____

Signed_____

Dulyauthorizedto signtheBidforand onbehalf of_____

Date_____

NOTE:

Port of Shipment and Country of origin of "MAJOR PART(S) OF THE EQUIPMENT" mustbe clearly reflected separately in the Technical and Financial bids. The "Origin" meanstheplacewherethe"goods"aremined,grown,orproduced.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.**_____
b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.**_____
c. **Qazi Muhammad Amir, Bio-Medical Technician, Electro-Medical Workshop.**_____
d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.**_____
e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.**_____

District Health Officer Thatta

Schedule of Requirements

The entire quantity of the ordered supplies shall be delivered in 16 weeks or earlier from the dateof issuance of supply order / contract award without any penalty.

Delay in the delivery shall result in penalties to be paid by the Supplier without any argument orquestion according to the prevailing SPPRA Rules.

Prepared/Reviewed by following members of Technical Committee (Machinery Equipment / Furniture etc):

a. **S. Zamir Hussain Shah,Bio Medical Engineer, Dr. Ruth K.M.Pfau Civil Hosp. Kar.** _____

b. **Majid Ali, Senior Equipment Engineer, NICH, KARACHI.** _____

c. **Qazi Muhammad Amir, Bio-medical Technician, Elector-medical Workshop.** _____

d. **Sunny Mohiuddin, Dy. Director (Procurement), Directorate of Pharmacy, Sindh, Kar.** _____

e. **Rajeevv Brahman, Pharmacist, Dr. Ruth K.M Pfau Civil Hospital Karachi.** _____

i.	Dr. Shahid Pervez, Histopathologist, Aga Khan University Hospital Karachi.	
ii.	Dr. Binish Arif Sultan, ~~Associate~~ Professor & ~~Head~~ of Pathology Lab, JSMU, Karachi.	
iii.	Prof. Dr. Tariq Mehmood, Professor Emeritus, Radiation Oncology, Member of PAF, JPMC, Karachi.	HEAD OF THE DEPARTMENT ASSISTANT PROFESSOR National Institute of Child Health Karachi
iv.	Dr. Syed Farjad Sultan, Associate Professor, Anaesthesia Department, DUHS / Civil Hospital Karachi.	
v.	Prof. Dr. Sajida Qureshi, Professor of Surgery, DUHS / Dr. Ruth K.M. Pfau Civil Hospital Karachi.	
vi.	Prof. Dr. Syed Mahmood Hassan, Professor of Pathology, JSMU, Karachi.	
vii.	Dr. Shoaib Malik, Associate Professor, Anesthesia Department, JPMC Karachi.	

CONCURRED BY:

Professor Dr. Ali Akbar Siyal
Vice Chancellor, Peoples University of Medical & Health Sciences for Women, Nawabshah-Shaheed Benazirabad/
Member

Professor Dr. Ikram Din Ujjan
Vice Chancellor, Liaquat University Medical & Health Sciences, Jamshoro./
Member

Prof. Dr. Sajida Qureshi
Professor of Surgery,
DUHS / Dr. Ruth K.M. Pfau CHK./ **Member**

Special Secretary,
Chief Minister's Inspection, Enquiries & Implementation Team Department,
Govt. of Sindh/ **Member**

Professor Dr. Nusrat Shah
Vice Chancellor, Shaheed Mohtarma
Benazir Bhutto Medical University
Larkana / **Member**

Dr. Ahson Qavi Siddiqi
CEO, Health Care Commission Sindh/
Member

Mr. Mohammed Ibrahim Memon
Additional Secretary (PM&I), Health
Department, Govt. of Sindh,/
Member / Secretary

Additional Secretary (GA),
SGA&CD, Govt. of Sindh/
Member

Director General Health Services Sindh,
Hyderabad./ **Member**

DIRECTOR GENERAL
HEALTH SERVICES SINDH
HYDERABAD.

Medical Superintendent,
Peoples Medical College Hospital
Nawabshah, SBA./ **Member**

PEOPLES MEDICAL COLLAGE HOSPITAL
NAWABSHAH SBA

Executive Director,
Jinnah Postgraduate Medical Centre,
Karachi./ **Member**

Prof. Khalid Sher
MBBS. MD.FCPS
EXECUTIVE DIRECTOR
Jinnah Postgraduate Medical Centre
Karachi

Medical Superintendent,
Dr. Ruth K.M.Pfau Civil Hospital
Karachi./ **Member**

Dr. S.M. Khalid Bukhari
Medical Superintendent
Dr. Ruth K.M.Pfau
Civil Hospital Karachi

Medical Superintendent,
Liaquat University Hospital,
Hyderabad./ **Member**

Medical Superintendent,
Ghulam Muhammad Mahar
Medical College Sukkur./**Member**

Professor Dr. Amjad Siraj Memon
Vice Chancellor,
Jinnah Sindh Medical University/
Chairman