

TENDER NOTICE NO: 1/ 2026

Purchase of Laparoscope 2DK

A Public Sector Hospital invites sealed tenders on single stage two envelope basis from eligible / authorized distributors/ reputed firms / suppliers registered with income tax / sale tax department on FOR Jauharabad basis. The bids must be reached to the undersigned on or before mentioning date / time, which will be opened in the presence of tenderers and their authorized representatives.

Sr. #	Tender. #	Description	Qty	Tender Fee (Rs.)	Last Date/Time of Bid Submission	Bids Opening Date/Time	Date of Technical Bids Opening
1	1/2026	Purchase of Laparoscope 2DK	1	2,000/-	03.10.2026 at 10:00 hours	03.10.2026 at 10:30 hours	03.10.2026 at 11:00 hours

1. A set of Tender documents along with terms and conditions can be downloaded from PPRA website (www.ppra.org.pk). An amount of Rs. 2,000/- (Non-Refundable) payable in shape of pay order / demand draft in favor of Head LAO, Public Sector General Hospital, Girote Chowk Jauharabad must be submitted along with technical proposal.
2. Earnest money of 2% of the bidding amount in the form of Pay Order/ Bank Draft, in favour of Head LAO, PAEC General Hospital, Jauharabad, must be attached along with the Technical Offer in separate sealed envelope.
3. The Competent Authority reserves the right to reject the tender.

Admin Officer (P)
Public Sector General Hospital
Girote chowk, Jauharabad, Ph: 0454-90239430



Government of Pakistan
PAKISTAN ORDNANCE FACTORIES
Wah CANTT



INVITATION TO BID

1. POFs Supply Chain Management invites bids from the original Manufactures/Authorized Dealers/Supplier registered with Income Tax, Sales Tax Department for stores against following tenders.

Sr. #	TE. No. & Date	Description of Store	Qty	Opening Date
1	0016-HVN-LP-57 Dt. 11-09-2026	Bolt with nut	12 items	06-10-2026
2	0024-HVN-LP-57 Dt. 11-09-2026	Electrical store	16 items	06-10-2026
3	0022-HVN-LP-57 Dt. 11-09-2026	Hydraulic pumps	02 items	06-10-2026
4	0013-HVN-LP-57 Dt. 11-09-2026	SS Seamless pipes	06 items	06-10-2026

2. Bidding documents containing item wise detail, specifications and all other terms & conditions are



TENDER NOTICE NO: 1/ 2026
Purchase of Laparoscope 2DK

A Public Sector Hospital invites sealed tenders on single stage two envelope basis from eligible / authorized distributors/ reputed firms / suppliers registered with income tax / sale tax department on FOR Jauharabad basis. The bids must be reached to the undersigned on or before mentioning date / time, which will be opened in the presence of tenderers and their authorized representatives.

Sr. #	Tender. #	Description	Qty	Tender Fee (Rs.)	Last Date/Time of Bid Submission	Bids Opening Date/Time	Date of Technical Bids Opening
1	1/2026	Purchase of Laparoscope 2DK	1	2,000/-	03.10.2026 at 10:00 hours	03.10.2026 at 10:30 hours	03.10.2026 at 11:00 hours

1. A set of Tender documents along with terms and conditions can be downloaded from PPRA website (www.ppra.org.pk). An amount of Rs.2,000/- (Non-Refundable) payable in shape of pay order / demand draft in favor of Head LAO, Public Sector General Hospital, Giron Chowk Jauharabad must be submitted alongwith technical proposal.
2. Earnest money of 2% of the bidding amount in the form of Pay Order/ Bank Draft, in favour of Head LAO, PAEC General Hospital, Jauharabad, must be attached along with the Technical Offer in separate sealed envelope.
3. The Competent Authority reserves the right to reject the tender.

Admin Officer (P)
Public Sector General Hospital
Giron chowk, Jauharabad
Ph: 0454-90239430

GSR

Latest Technology HD 1920x1080 Laparoscopic advance Surgery System with instruments set

FULL HD VIDEO PROCESSOR

1	Camera Processor	Full HD latest Model
2	Image Quality	Full HD or better
3	Camera Input	Full HD camera Head/Video Endoscopes
4	Video output	X2 DVI out put
5	Image capturing & Recording	Should have Inbuilt recording facility as well as multiple usb ports for external device based/usb storage or better.
6	Special feature	Should be able to use with Rigid & flexible Video Endoscopes
7	Power requirements	100-120VAC/200-240VAC, 50/60Hz

Camera Head Full HD

LED LIGHT SOURCE UNIT

1	Lamp type	High Performance LED , lamp life 20,000hrs or better
2	Color Temperature	6400K
3	Lamp Out let	1
4	Control	Knob or Buttons,
5	Lamp life	20,000 hrs or more
7	Fiber optic light cable	Same manufacturer light cable 4.6mm 240cm qty 01

5- 26" Full HD MONITOR from Same manufacturer

1	Chip technology	Latest chip CMOS or better full HD (1920x1080P)
2	Resolutions	1920*1080 pixels
3	Scan mode	Progressive, 50 /60Hz,
4	camera head buttons	2 to 3 programmable buttons with remote functions
5	Focal length	focal length f=16mm
6	Zooming	Should have Digital OR Optic Zoom
9	Integrated Control	Should be able to control complete cameras system i.e. video processing module, Camera modules, Video Capturing, Still images. Through camera head.
10	Teaching Tools	System should have a teaching tools to point out the surgical area to students
11	Cleaning/Sterilization	Soakable, gas and plasma sterillzable.
1	Max. screen resolution	1920 x 1080
2	Image Format	16:9
3	Video Signal Input	DVI, VGA, S-Video, Composite /FBAS
4	Video Signal Output	DVI, S-Video, Composite/FBAS
5	Power Supply	24V External Power Supply
6-	Monitor quality	From same manufacture

5- RECORDING SYSTEM FROM SAME MANUFACTURER

1	System type	In built system in processor or External Device
2	Still Images	UHD Quality (JPEG)
3	Video Capturing	Full HD quality (MPEG4)
4	Dual function	should be able to capture still images and video recording at same time
5	Data Storage	External USB type or internal hard drive
6	Printer compatibility	Device or system should have compatibility for medical grade printer

50 Liter Endoflator from same manufacturer

a.	Display	LCD or better
b	Controls	Touch screen
c.	Max Gas Pressure	0-30mmHg or equivalent
d.	Max Gas Flow	0-50 l/min or equivalent
e.	Functions	Adult and Pediatrics on One Touch Function, Gas flow must be heated/ thermal insufflation as per body temperature
f.	Integrated Control	Should be control able thorough camera Head & keyboard
g.	Parameter display	Insufflations Pressure, Intraabdominal Pressure; 0-30mmHg, Gas flow: 0-50 l/min, Gas Consumption: 0-999L
h.	Power	100-240 VAC, 50/60Hz

Local- Best Quality Equipment Cart Moveable with multiple shelves and 02 wheels must have locks

Reusable Laparoscopy Advance Surgery Instruments Set from Same Manufacturer

	Latest 4k Telescope 0Deg, 10mm, 30cm or better ,	1
	Latest 4k Telescope 30Deg,10mm,30cm or better,	1
	Trocar 6mm	2
	Trocar 11mm	2
	Marry land dissecting forcep 16mmjaw, 5mm, 36cm	1

Babcock forcep ,5mm, 36cm	1
Scissor curved 5mm, 36cm	1
Johan forcep 5mm, 36cm	1
Gallbladder holding forcep 5mm, 36cm with locking handle	1
Bowel grasper 5mm, 36cm	1
Tooth grasper 5mm, 36cm with locking handle	1
Clip applicator 10mm	1
Needle holder 5mm cvd to left	1
Claw forcep 10mm with locking handle	1
Stone collection forcep 10mm	1
Suction irrigation tube 5mm 36cm	1
Hook electrode 5mm, length 36cm	1
Diathermy lead monopolar	2
Injection cannula	1
Veress needle 15cm	1
Reducer external 10/5mm	2
Knot tier	1
Facial closer forceps	1
Insulated outer tubes for 5mm instruments	2
Trocar seal CAPS for 6mm trocar	20
Trocar seal CAPS for 11mm trocar	20
Camera head disposable covers from same OEM recommen	50p
Nathenson liver retractor medium and large	2
Fan liver retractor 10mm	1
Instruments container for 10 to 12 instruments	1
Articulated Holding Arm with Multiple Movements as well As compatible with OEM exo lens or better	1
Instruments Holding	1
Extension Rod for Articulated holding arm	1
Table clamp fir holding system	1
Telescope sterilization box for telescopes of 10mm	01

Terms & Conditions

All items from same manufacturer, CE Mark FDA Approved

At least 5 latest installations in last 2 years

Demonstration of Equipment / latest technology is mandatory

Instructions for Tender:

1. The quotations should be addressed to Admin Officer PAEC General Hospital Girote chowk, Jauhaurabad. The bidders shall submit their offers with reference to clause No.36 (b) of PPRA rules i.e. Single Stage - Two envelopes bidding procedure as per following details:

- i. The bid shall comprise a single package containing two separate envelopes. Each envelop shall contain separately the financial proposal and the technical proposal.
- ii. Envelops shall be marked as "TECHNICAL PROPOSAL" and "FINANCIAL PROPOSAL" in bold and Legible letters to avoid confusion.
- iii. Initially only envelop marked "TECHNICAL PROPOSAL" shall be opened.
- iv. Envelop marked as "FINANCIAL PROPOSAL" shall be retained in safe custody without being opened.
- v. The technical proposal shall be evaluated with out reference to price and any proposal which do not conform to the specified requirements shall be rejected.
- vi. No amendment in the technical proposal shall be permitted during evaluation.
- vii. The Financial proposal of bids shall be opened publicly at a time, date and venue announced and communicated to the bidders in advance of technically accepted bids only.
- viii. Financial proposal of the bids found non-responsive shall be returned un-opened to the respective bidders.

General terms & conditions:

2. All the Technical and Financial Proposals in Separate Sealed Envelopes clearly marked as "Tender for Laproscope 2DK Tender Notice No.1/2026" should reach to undersigned on or before the tender opening date 03.10.2026 at 10:00 hrs. Technical bids will be opened on the same day at 10:30 hrs at (Conference Room) PAEC General Hospital Girote Chowk Jauhaurabad in the presence of bidders who submitted their offers as per instruction/ requirement and wish to be present. In case of falling holiday on the opening date, the tenders will be opened on the next working day at the same time.

3. No tender will be considered if:-

- a. It is received after the date and time fixed.

- b. The tender and related documents are not signed & Stamp.
 - c. The offer is ambiguous.
 - d. Items offered differ in description etc from those required.
 - e. It is without earnest money or with less earnest money as a general policy.
 - f. Over writing and cutting will be not accepted
 - g. If rates received are higher than reasonable/retail rates.
4. The quotation validity period will be started from the date of **issuance of award letter** and rates will be valid till completion of one year running contract.
5. **Earnest money of 2% of the bidding amount** in the form of **Pay Order/ Bank Draft**, in favour of Head LAO, PAEC General Hospital, Jauhaurabad, must be attached **along with the Technical Offer in separate sealed envelope**. Earnest money will be returned to the unsuccessful bidders soon after finalization of the purchase proposal, and retained in case of successful bidders till satisfactory completion of regular contract agreement.
7. The following are required to be mentioned in the quotation and copy also attached:-
- Letter of registration under sales Tax Act 1990. Copy
 - Name of Supplier
 - Income / Sales Tax Registration No.
 - Active Taxpayer list (ATL)
8. **Special Instructions:**
- i. No changes will be allowed during the currency of contract in supplier's distribution and the supply orders will be generated in the name of same supplier quoted in the tender.
9. **Risk Liquidated Damages:**
- The quantities to be purchased from the suppliers can be increased or decreased as per requirement of PGHJ. The suppliers are to ensure to timely delivery of items and if supplier fails to supply the stores up to the requisite quantity, quality and specifications within the delivery date / specified time. Purchaser will have right to confiscate the submitted earnest / security money, cancel the order and purchase the stores elsewhere at your risk & cost for which no notice shall be issued and the purchaser will not be responsible for any loss that you may suffer due to the cancellation of the order, or in such a case the purchaser may either impose a penalty on you to be paid by you by way of liquidated damages @ 2% of the value of un-supplied stores per day for the number of days by which the delivery is delayed and beyond the said date / specified time up-to a maximum of 10% of the total contract value.
10. **Warranty:**

The supplier will provide a warranty certificate and software certificate that quality of stores is strictly in accordance to the specification given in the attached schedule to this contract and any defective part will be replaced free of all charges (including carriage/ loading/ unloading charges from site to supplier's premises and back).

11. Any correspondence regarding this tender must be addressed to Admin Officer, PAEC General Hospital Girote chowk Jauhaurabad only. For any additional information's please contact on telephone numbers as given below.
12. The Competent Authority reserves the right to accept or reject any or all the offers.
13. If the bidder/supplier shall have to accept unconditionally all "Terms & Conditions" of PGHJ to participate in open tender.

Admin Officer (Proc)
Public Sector General Hospital
Girote chowk Jauhaurabad
Ph: 0454-90239430



FUNDS TRANSFERRED

Transaction reference ID # 254624348677

11-Sep-2026 09:18:22 AM

PAID FROM

ARHAM ZEESHAN
*****1027

TRANSFER TYPE

Interbank

BENEFICIARY

PUBLIC PROCUREMENT
*****0701
HBL/Konnect

PURPOSE

Other

REFERENCE tender jauharabad laproscope

AMOUNT

15,000