

PAC HOSPITAL KAMRA (DISTT ATTOCK)
TENDER NOTICE FOR ELECTRO MEDICAL AND GENERAL
EQUIPMENTS
FINANCIAL YEAR: 2026-27

1. Quotations are invited from the sales tax registered firms for purchase of following item according to the specifications mentioned against each for used in PAC Hospital Kamra.

S No	Description	Qty
(a)	Physio Dispenser / Implant Motor (Dental)	01
(b)	Dermatoscope with Accessories	01
(c)	Fumigation Machine with Accessories	01
(d)	Computer UPS (650 VA-Online)	20
(e)	Vein Finder	As required
(f)	Quilted Fleece Blanket for Hospital	As required
(g)	Bed sheet Cotton White	As required

(a) SPECIFICATIONS OF PHYSIO DISPENSER / IMPLANT MOTOR:-

- (a) Power Supply : AC 100 to 240 V & 50 to 60 Hz
- (b) Max Pump Output : 75 mL/min
- (c) Programs : 8 Programs / Implant Systems
- (d) Dimensions : W 245 x D 235 x H 90 mm to
W 295 x D 265 x H 100 mm
- (e) Torque : 5-80 Ncm
- (f) Motor Speed : 200-40,000 / min
- (g) Display : Coloured & Touch LCD / LED
- (h) Motor with Cord : Autoclave able
- (j) Foot Control : Wireless (With Bluetooth connectivity,
which allows selections of optimal
position)
- (k) Hand Piece : Contra Angle 20:1, Head size smaller
than conventional SG20 (mini head),
double sealing system & stress-less
body shape
- (l) Non-Optic Motor & Cord
- (m) Tube Holder (at least pack of 05)
- (n) Irrigation Tube (at least pack of 05)
- (p) Carrying case
- (q) Coolant Solution Hanger Post

STANDARD / SPECIAL CONDITIONS

- (a) Payment as per PAC Board purchase procedure.
- (b) Item should be as per standard of ISO 13485.
- (c) Two years Warranty/Guarantee is required.
- (d) Store on receipt will be accepted after inspection by EME section of PAC Hospital
- (e) Store must be latest production series (factory new).
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (h) Free of cost training on operation, maintenance and calibration of machine at PAC premises must be provided during warranty period.
- (j) OEM Conformance and Quality Assurance Certificates are required at the time of delivery.
- (k) Authorized distributor / dealer certificate of the equipment / store is required.
- (l) Country of origin and manufacturer certification JIS, FDA, CE or MHLW is required. (Any One)
- (m) Product will be acceptable only with OEM standardized packing and labeled / endorsed Part Number, Batch / LOT No, MFG / EXP Dates. No local / supplier sticker is acceptable.
- (n) If calibration is required, newly calibrated item is to be provided by the vendor.
- (p) The repair / sales Centre of the equipment / machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.
- (q) At least ten years after sale spare supportability is to be provided.
- (r) Delivery period 06 to 08 weeks after signing of contract.
- (s) All transport, shipping, freight charges and unloading of shipment is the responsibility of the Seller.

(b) SPECIFICATIONS OF DERMATOSCOPE WITH ACCESSORIES:-

- (a) Multi-LED illumination.
- (b) High Resolution precision optic system.
- (c) Extended field of view
- (d) Portable
- (e) Weight : 200 to 400 gm
- (f) Dimension : (L x W x D) 195 x 55 x 70 mm to 200 x 60 x 80 mm
- (g) Rechargeable
- (h) Polarized and Non-Polarized light
- (j) Ability to capture Images
- (k) 10x optical zoom

- (l) Tablet, Computer and Smartphone compatibility.
- (m) Compatible Android Tablet or equivalent is required with the equipment.
- (n) **Accessories:-**
 - (i) Adapter / Connector for Tablet or equivalent.
 - (ii) Adapter / Connector for Computer.
 - (iii) Storage Bag

"Or Equivalent"

STANDARD / SPECIAL CONDITIONS

- (a) Payment as per PAC Board purchase procedure.
- (b) Item should be as per standard of ISO 13485.
- (c) Two years Warranty/Guarantee is required.
- (d) Store on receipt will be accepted after inspection by EME section of PAC Hospital
- (e) Store must be latest production series (factory new).
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (h) Free of cost training on operation, maintenance and calibration of machine at PAC premises must be provided during warranty period.
- (j) OEM Conformance and Quality Assurance Certificates are required at the time of delivery.
- (k) Authorized distributor / dealer certificate of the equipment / store is required.
- (l) Country of origin and manufacturer certification JIS, FDA, CE or MHLW is required. (Any One)
- (m) Product will be acceptable only with OEM standardized packing and labeled / endorsed Part Number, Batch / LOT No, MFG / EXP Dates. No local / supplier sticker is acceptable.
- (n) If calibration is required, newly calibrated item is to be provided by the vendor.
- (p) The repair / sales Centre of the equipment / machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.
- (q) At least ten years after sale spare supportability is to be provided.
- (r) Delivery period 06 to 08 weeks after signing of contract.
- (s) All transport, shipping, freight charges and unloading of shipment is responsibility of the Seller.

(c) SPECIFICATIONS OF FUMIGATION MACHINE WITH ACCESSORIES :-

- (a) Power Supply : AC 200 to 240 V & 50 / 60 Hz
- (b) Weight : 6.5 Kg to 7 Kg

- (c) Engine Power : 1100 W to 1200 W
- (d) Body : Stainless Steel / Rugged material
- (e) Disinfectant: : Solution containing hydrogen peroxide and colloidal silver.
- (f) Disinfection Method : Sprays disinfectant as particles smaller than 5 microns.
- (g) User Interface: : Numerical keypad for programming, with optional printer for automatic output.
- (h) Safety Features : Light and sound alerts at the beginning and end of the disinfection cycle.
- (j) Other Features : Delayed start, pause function, password protection for settings, and a special heading system.
- (k) Language : English (UK / USA)
- (l) Compatibility : Compatibility: Compatible with various materials and does not leave residue or wetness.
- (m) Automatic printout feature for retrospective tracking of disinfection processes.
- (n) Accessories:-
 - (i) Remote
 - (ii) Power Cable

"Or Equivalent / Higher"

STANDARD / SPECIAL CONDITIONS

- (a) Payment as per PAC Board purchase procedure.
- (b) Item should be as per standard of ISO 13485.
- (c) Two years Warranty/Guarantee is required.
- (d) Store on receipt will be accepted after inspection by EME section of PAC Hospital
- (e) Store must be latest production series (factory new).
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (h) Free of cost training on operation, maintenance and calibration of machine at PAC premises must be provided during warranty period.
- (j) OEM Conformance and Quality Assurance Certificates are required at the time of delivery.
- (k) Authorized distributor / dealer certificate of the equipment / store is required.
- (l) Country of origin and manufacturer certification JIS, FDA, CE or MHLW is required. (Any One)
- (m) Product will be acceptable only with OEM standardized packing and labeled / endorsed Part Number, Batch / LOT No, MFG / EXP Dates. No local / supplier sticker is acceptable.
- (n) If calibration is required, newly calibrated item is to be provided by the vendor.

- (p) The repair / sales Centre of the equipment / machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.
- (q) At least ten years after sale spare supportability is to be provided.
- (r) Delivery period 06 to 08 weeks after signing of contract.
- (s) All transport, shipping, freight charges and unloading of shipment is responsibility of the Seller.

(d) SPECIFICATIONS OF COMPUTER UPS (650 VA-ONLINE)

- | | | | |
|----|----------------------|---|---|
| 1. | Type | : | Desktop & (Smart Online, Double Conversion) |
| 2. | Waveform Type | : | Line Interactive, Pure Sinewave |
| 3. | Output | : | 400 Watts / 650VA |
| 4. | Main Output Voltage | : | 230 V, 50/60 Hz $\pm$ 3 Hz |
| 5. | Main Input Voltage | : | 230 V, 47-63 $\pm$ 3 Hz Auto Sensing |
| 6. | Battery Type | : | Lead Acid Battery |
| 7. | Warranty / Guarantee | : | One Year OEM Warranty |

STANDARD / SPECIAL CONDITIONS FOR ABOVE MENTIONED ITEM

- (a) Payment as per PAC Board purchase procedure.
- (c) Store on receipt will be accepted after inspection by EME section of PAC Hospital
- (d) Store must be latest production series (factory new).
- (e) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (f) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (g) Free of cost training on operation, maintenance and calibration of UPS at PAC premises must be provided during warranty period.
- (h) Product will be acceptable only with OEM standardized packing and labeled / endorsed Part Number, Batch / LOT No, MFG / EXP Dates. No local / supplier sticker is acceptable.
- (j) Delivery period 06 to 08 weeks after signing of contract.
- (k) All transport, shipping, freight charges and unloading of shipment is responsibility of the Seller.

(e) SPECIFICATIONS OF VEIN FINDER:-

- | | | | |
|----|-----------------------|---|-----------------------------------|
| 1. | Light Source | : | Near Infrared |
| 2. | Projection Resolution | : | Minimum 800 X 400 |
| 3. | Projection Size | : | Very Small, Small, Medium & Large |

4. Projection Brightness : 4 levels adjustable
5. Projection Colour : Green, Green Background, Black White, Red, Purple & Purple Background
6. Depth prompts : Deep, Medium & Shallow
7. Resolution speed should be minimum 50 frames per second.
8. Automatically goes into sleeping mode while not in use.
9. Battery working time minimum 2 hours.
10. Vein Finder could be use with mobile stand, desktop / benchtop and handheld.
11. Mobile stand and desktop / benchtop both are required with the equipment.

STANDARD / SPECIAL CONDITIONS FOR ABOVE MENTIONED ITEM

- (a) Payment as per PAC Board purchase procedure.
- (b) Item should be as per standard of ISO 13485.
- (c) Two years Warranty/Guarantee is required.
- (d) Store on receipt will be accepted after inspection by EME section of PAC Hospital
- (e) Store must be latest production series (factory new).
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (h) Free of cost training on operation, maintenance and calibration of machine at PAC premises must be provided during warranty period.
- (j) OEM Conformance and Quality Assurance Certificates are required at the time of delivery.
- (k) Authorized distributor / dealer certificate of the equipment / store is required.
- (l) Country of origin and manufacturer certification JIS, FDA, CE or MHLW is required. (Any One)
- (m) Product will be acceptable only with OEM standardized packing and labeled / endorsed Part Number, Batch / LOT No, MFG / EXP Dates. No local / supplier sticker is acceptable.
- (n) If calibration is required, newly calibrated item is to be provided by the vendor.
- (p) The repair / sales Centre of the equipment / machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.
- (q) At least ten years after sale spare supportability is to be provided.
- (r) Delivery period 06 to 08 weeks after signing of contract.
- (s) All transport, shipping, freight charges and unloading of shipment is responsibility of the Seller.

(f) SPECIFICATIONS OF BLANKET FOR HOSPITAL:-

**TECHNICAL SPECIFICATIONS
QUILTED FLEECE BLANKET HOSPITAL**

Color	:	Grey Fleece Double Sided (Color should be match with sample of Log Sqn)
Size	:	60"X80"
Filling Material	:	Polyster Filling
Filling GSM	:	300 GSM
Outer Fabric	:	Premium Quality Fleece
Quality	:	A1 Quality
Construction	:	Quilted Feel & Comfort Soft, Warm and Comfortable
Warmth	:	Suitable for provided effective warmth during colder weather
Durability	:	High-quality fabric and filling designed for long -- lasting use
Finish	:	Neat and premium – quality quilting finish

STANDARD / SPECIAL CONDITIONS FOR ABOVE MENTIONED ITEM

- (a) Payment as per PAC Board purchase procedure.
- (b) Blanket color fading warranty / Guarantee is required for One year.
- (c) Store on receipt will be accepted after inspection by Log Sqn of PAC Hospital
- (d) Store must be latest production series (factory new).
- (e) Complete consignment will be replaced by supplier free of cost in case of any defect found during warranty / guarantee period.
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Delivery period 06 to 08 weeks after signing of contract.
- (h) All transport, shipping, freight charges and unloading of shipment is responsibility of the Seller.
- (j) (PAC Hospital Kamra) should be embroided on the Blanket.

(g) SPECIFICATIONS OF SHEET COTTON WHITE:-

**TECHNICAL SPECIFICATIONS
SHEET COTTON WHITE**

Material	:	100 % Cotton
Wt / Sq Mtr	:	175 Gms $\pm$ 5 gms
Weight / Sheet (Minimum)	:	950 Gms
~~Filling GSM~~	~~:~~	~~500 GSM~~
Outer Fabric	:	Premium Quality Fleece
Length	:	240 + 2.5 cm – 1.5 cm
Width	:	135 + 2.5 cm – 1.5 cm
Hem	:	2 cm
Turn in	:	2 cm
Count of Yarn		
Warp	:	36.9 Tex
Weft		36.9 Tex
No of Thread per 25 mm		
Warp	:	60 – 62
Weft		52 - 54
Breaking Strength (Minimum)		
Warp	:	79 Kg
Weft		56 Kg
Shrinkage (Maximum)		
Warp	:	2 %
Weft		2 %
Scouring Loss	:	2 %
Shade	:	White Bleached (The Sheets with poor whiteness Will not be acceptable
Softness	:	The sheet should be soft in handle
Weave	:	Plain
Envelope Folding Flap (Under Mattress Tuck)	:	An Extra 6" – 8" extension of fabric along the head/ foot of the sheet that tucks completely under the mattress base to keep the sheet taut.

STANDARD / SPECIAL CONDITIONS FOR ABOVE MENTIONED ITEM

- (a) Payment as per PAC Board purchase procedure.
- (b) Blanket color fading warranty / Guarantee is required for One year.
- (c) Store on receipt will be accepted after inspection by Log Sqn of PAC Hospital
- (d) Store must be latest production series (factory new).
- (e) Complete consignment will be replaced by supplier free of cost in case of any defect found during warranty / guarantee period.
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (g) Delivery period 06 to 08 weeks after signing of contract.
- (h) All transport, shipping, freight charges and unloading of shipment is responsibility of the Seller.
- (j) (PAC Hospital Kamra) should be embroidered on the Blanket.

2. Sealed Technical & Commercial quotations with the title **"Tender for Electro Medical Equipment with Description"** must be submitted to the undersigned on **29 October, 2026** up to **1030 Hrs** positively. **3% Bid Security** must be present in the sealed Commercial Quotation in the name of **Officer In-Charge Non-Public Fund PAC Hospital Kamra**. The photocopy of **3% Bid Security** with black out value column should be present in the sealed Technical Quotation to ensure that original **3% Bid Security** is present in the sealed Commercial Quotation. Tenders would be opened at **1100 Hrs** on the same date.

3. **Only one offer** can be made in relevant currency for same item and quoted price must be in **PKR**. Multiple rates, if quoted, will be rejected. Multiple quotations against the tender will also be rejected. **Validity of quotation must be upto 30 June, 2027.**

4. **Treasury Challan** Offer must be accompanied with a challan form of **Rs 3,000/-** (obtained from State Bank of Pakistan / Government Treasury) and debit able to **Major Head C02501-20, Main Head 12, Sub Head "A' Miscellaneous (Code Head 1/845/30)**. Only one offer can be made for same item on one challan.

5. **Comparison Sheet (Vis a Vis)** of advertised and offered specifications and also acceptance of standard / special conditions alongwith with documentary evidence (if required) must be accompanied with technical quotation on firm's letter head.

6. **Demonstration** of quoted unit may be taken before technical scrutiny finalization if deemed necessary by the concerned specialist.
7. Previous performance of the suppliers will be considered while accepting / rejecting the tender.

Officer Commanding
PAC Hospital Kamra (Attock)
051-909902352